


Data Path : Z:\HPCHEM1\BNA F\DATA\BF042116\
 Data File : BF086345.D
 Acq On : 21 Apr 2016 14:26
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF042016.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 20 15:45:26 2016



AutoFind: Scans 835, 836, 837; Background Corrected with Scan 831

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.4	29733	PASS
68	69	0.00	2	1.9	560	PASS
69	198	0.00	100	38.1	28726	PASS
70	69	0.00	2	0.9	247	PASS
127	198	10	80	50.9	38344	PASS
197	198	0.00	2	1.1	834	PASS
198	198	100	100	100.0	75398	PASS
199	198	5	9	7.1	5351	PASS
275	198	10	60	24.8	18708	PASS
365	198	1	100	3.0	2245	PASS
441	442	0.01	24	15.5	11601	PASS
442	198	50	100	99.3	74842	PASS
443	442	15	24	19.7	14730	PASS