

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042216\
 Data File : BF086411.D
 Acq On : 24 Apr 2016 1:11
 Operator : UM/SJ
 Sample : H2652-15 4X
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WPG-B5(0-5)-C

Manual Integrations
 APPROVED

sohil
 4/25/2016 12:36:56 PM

Quant Time: Apr 25 07:23:55 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Apr 23 00:24:02 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	171305	20.00	ng	0.00
21) Naphthalene-d8	8.13	136	680858	20.00	ng	0.00
38) Acenaphthene-d10	9.88	164	293456	20.00	ng	0.00
63) Phenanthrene-d10	11.37	188	498022	20.00	ng	0.00
75) Chrysene-d12	14.00	240	358197	20.00	ng	0.00
86) Perylene-d12	15.43	264	252786	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	402993	40.63	ng	0.01
7) Phenol-d6	6.48	99	538909	39.96	ng	0.00
23) Nitrobenzene-d5	7.40	82	282141	24.55	ng	-0.01
41) 2,4,6-Tribromophenol	10.67	330	89603	33.73	ng	0.00
44) 2-Fluorobiphenyl	9.21	172	547654	31.85	ng	0.00
78) Terphenyl-d14	12.94	244	417301	28.95	ng	0.00
Target Compounds						Qvalue
49) Dimethylphthalate	9.60	163	53495	2.52	ng	99
74) Fluoranthene	12.57	202	86802	3.22	ng	97
77) Pyrene	12.80	202	88178	3.53	ng	99
80) Benzo(a)anthracene	13.99	228	46769	2.52	ng	# 87
87) Benzo(b)fluoranthene	15.01	252	37744m	2.72	ng	
89) Benzo(a)pyrene	15.36	252	30945	2.26	ng	# 50

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042216\
Data File : BF086411.D
Acq On : 24 Apr 2016 1:11
Operator : UM/SJ
Sample : H2652-15 4X
Misc :
ALS Vial : 55 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WPG-B5(0-5)-C

Manual Integrations
APPROVED

sohil
4/25/2016 12:36:56 PM

Quant Time: Apr 25 07:23:55 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Apr 23 00:24:02 2016
Response via : Initial Calibration

