

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042315\
 Data File : BF078773.D
 Acq On : 24 Apr 2015 6:22
 Operator : TP/IZ
 Sample : G1987-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-12

Quant Time: Apr 24 07:19:29 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 24 04:11:59 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.76	152	25934	20.00	ng	0.00
21) Naphthalene-d8	7.92	136	110057	20.00	ng	0.00
38) Acenaphthene-d10	11.10	164	56418	20.00	ng	0.00
63) Phenanthrene-d10	13.84	188	112250	20.00	ng	0.01
75) Chrysene-d12	17.71	240	127826	20.00	ng	0.00
86) Perylene-d12	19.42	264	127905	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	3.79	112	9100	5.79	ng	0.02
7) Phenol-d6	5.30	99	9347m	4.74	ng	0.03
23) Nitrobenzene-d5	6.70	82	14096	7.76	ng	0.00
41) 2,4,6-Tribromophenol	12.58	330	5463	11.68	ng	0.01
44) 2-Fluorobiphenyl	9.93	172	27507	6.97	ng	0.00
78) Terphenyl-d14	16.38	244	30163	5.33	ng	0.00
Target Compounds						
49) Dimethylphthalate	10.74	163	17476	4.12	ng	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042315\
Data File : BF078773.D
Acq On : 24 Apr 2015 6:22
Operator : TP/IZ
Sample : G1987-03
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-12

Quant Time: Apr 24 07:19:29 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 24 04:11:59 2015
Response via : Initial Calibration

