

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042318\
 Data File : BF104740.D
 Acq On : 24 Apr 2018 6:58
 Operator : JU/SJ
 Sample : J2503-08
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CHRT25773

Quant Time: Apr 24 08:45:17 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 23 15:38:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	111087	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	439735	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	179152	20.00	ng	0.00
63) Phenanthrene-d10	11.33	188	287647	20.00	ng	0.00
75) Chrysene-d12	13.98	240	204224	20.00	ng	0.00
86) Perylene-d12	15.45	264	150282	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.44	112	762939	105.01	ng	0.02
7) Phenol-d6	6.45	99	805125	90.20	ng	0.00
23) Nitrobenzene-d5	7.37	82	611198	90.76	ng	0.00
41) 2,4,6-Tribromophenol	10.64	330	204136	109.85	ng	0.00
44) 2-Fluorobiphenyl	9.17	172	1117941	98.58	ng	0.00
78) Terphenyl-d14	12.92	244	933101	104.17	ng	0.00
Target Compounds						
31) Naphthalene	8.11	128	47369	2.163	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042318\
Data File : BF104740.D
Acq On : 24 Apr 2018 6:58
Operator : JU/SJ
Sample : J2503-08
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CHRT25773

Quant Time: Apr 24 08:45:17 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 23 15:38:25 2018
Response via : Initial Calibration

