

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042417\
 Data File : BF094579.D
 Acq On : 24 Apr 2017 12:26
 Operator : SJ/MA
 Sample : I2790-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ROLLOFFS-156-VNJ213-155-160-34

Quant Time: Apr 25 02:37:09 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 17 14:59:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.75	152	170154	20.00	ng	0.00
21) Naphthalene-d8	7.25	136	648388	20.00	ng	-0.01
38) Acenaphthene-d10	9.38	164	273209	20.00	ng	-0.02
63) Phenanthrene-d10	11.20	188	458195	20.00	ng	-0.02
75) Chrysene-d12	14.45	240	304020	20.00	ng	-0.01
86) Perylene-d12	16.08	264	273296	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.32	112	1136615	110.82	ng	0.02
7) Phenol-d6	5.39	99	1416813	117.18	ng	0.00
23) Nitrobenzene-d5	6.41	82	871545	83.00	ng	0.00
41) 2,4,6-Tribromophenol	10.36	330	249520	116.76	ng	-0.01
44) 2-Fluorobiphenyl	8.57	172	1403001	75.56	ng	-0.01
78) Terphenyl-d14	13.18	244	946713	83.23	ng	-0.01
Target Compounds						Qvalue
31) Naphthalene	7.27	128	189666	6.08	ng	99
37) 2-Methylnaphthalene	8.11	142	171361	8.04	ng	98
49) Dimethylphthalate	9.07	163	195457	9.60	ng	98
54) Dibenzofuran	9.63	168	53613	2.23	ng	97
70) Phenanthrene	11.23	178	271422	10.52	ng	98
74) Fluoranthene	12.69	202	331937	12.41	ng	99
77) Pyrene	12.97	202	303499	14.29	ng	97
80) Benzo(a)anthracene	14.44	228	132363	7.49	ng	96
82) Chrysene	14.48	228	150918m	9.35	ng	
83) Bis(2-ethylhexyl)phthalate	14.51	149	41905	3.35	ng	98
85) Indeno(1,2,3-cd)pyrene	17.35	276	76665	4.99	ng	97
87) Benzo(b)fluoranthene	15.68	252	164200m	9.82	ng	
88) Benzo(k)fluoranthene	15.70	252	55717m	3.52	ng	
89) Benzo(a)pyrene	16.02	252	113463	7.29	ng	# 95
91) Benzo(a,h,i)perylene	17.72	276	73858	5.62	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042417\
Data File : BF094579.D
Acq On : 24 Apr 2017 12:26
Operator : SJ/MA
Sample : I2790-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ROLLOFFS-156-VNJ213-155-160-34

Quant Time: Apr 25 02:37:09 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 17 14:59:23 2017
Response via : Initial Calibration

