


Data Path : Z:\HPCHEM1\BNA F\Data\BF042518\
 Data File : BF104780.D
 Acq On : 25 Apr 2018 11:18
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040618.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 25 14:01:50 2018



AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.9	25800	PASS
68	69	0.00	2	1.9	518	PASS
69	198	0.00	100	34.6	27988	PASS
70	69	0.00	2	0.4	103	PASS
127	198	10	80	50.5	40835	PASS
197	198	0.00	2	0.5	373	PASS
198	198	100	100	100.0	80848	PASS
199	198	5	9	6.7	5387	PASS
275	198	10	60	27.6	22293	PASS
365	198	1	100	3.7	2994	PASS
441	443	0.01	100	81.4	12477	PASS
442	198	50	100	99.6	80538	PASS
443	442	15	24	19.0	15330	PASS