



Instrument ID: BNA_F

284 Sheffield Street, Mountainside NJ 07092 (908) 789-8900

Daily Analysis Runlog For Sequence/QC Batch ID #bf042522

STD. NAME	STD REF.#	STD. NAME	STD REF.#
Review By	Jagrut Upadhyay	Review On	4/26/2022 9:50:12 AM
Tune/Reschk	SP5834	Initial Calibration Stds	SP5777,SP5778,SP5779,SP5780,SP5781,SP5782,SP5783
CCC	SP5780	SubDirectory	bf042522
Internal Standard/PEM	S10212,10ul/1000ul sample	HP Acquire Method	BNA_F
ICV/I.BLK	SP5813	HP Processing Method	bf042222
Surrogate Standard	SP5822,SP5811	MS/MSD Standard	SP5854
LCS Standard	SP5854		

Sr#	SampleId	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	BF127869.D	25 Apr 2022 09:44		CG\JU	Ok
2	SSTDCCC040	BF127870.D	25 Apr 2022 10:15		CG\JU	Ok
3	PB144283BL	BF127871.D	25 Apr 2022 10:45		CG\JU	Ok
4	PB144283BS	BF127872.D	25 Apr 2022 11:13		CG\JU	Ok,M
5	N2487-03	BF127873.D	25 Apr 2022 11:44	Double Internal Standard	CG\JU	Not Ok
6	N2494-01	BF127874.D	25 Apr 2022 12:12	Double Internal Standard	CG\JU	Not Ok
7	N2482-17	BF127875.D	25 Apr 2022 12:40		CG\JU	Ok
8	N2526-02	BF127876.D	25 Apr 2022 13:08		CG\JU	Ok
9	N2548-07	BF127877.D	25 Apr 2022 13:36		CG\JU	Ok
10	N2481-04	BF127878.D	25 Apr 2022 14:04		CG\JU	Ok
11	DFTPP	BF127879.D	25 Apr 2022 14:36		CG\JU	Ok
12	SSTDCCC040	BF127880.D	25 Apr 2022 15:04		CG\JU	Ok
13	PB144239BL	BF127881.D	25 Apr 2022 15:33		CG\JU	Ok
14	PB144239BS	BF127882.D	25 Apr 2022 16:01		CG\JU	Ok,M
15	N2550-01	BF127883.D	25 Apr 2022 16:32		CG\JU	Ok
16	N2550-02	BF127884.D	25 Apr 2022 17:01		CG\JU	Ok
17	N2474-02	BF127885.D	25 Apr 2022 17:29		CG\JU	Ok
18	N2474-05	BF127886.D	25 Apr 2022 17:57		CG\JU	Ok
19	N2548-01	BF127887.D	25 Apr 2022 18:26		CG\JU	Ok
20	N2548-01MS	BF127888.D	25 Apr 2022 18:55		CG\JU	Ok,M
21	N2548-01MSD	BF127889.D	25 Apr 2022 19:23		CG\JU	Ok,M



Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID #bf042522

STD. NAME	STD REF.#	STD. NAME	STD REF.#
Review By	Jagrut Upadhyay	Review On	4/26/2022 9:50:12 AM
Tune/Reschk	SP5834	Initial Calibration Stds	SP5777,SP5778,SP5779,SP5780,SP5781,SP5782,SP5783
CCC	SP5780	SubDirectory	bf042522
Internal Standard/PEM	S10212,10ul/1000ul sample	HP Acquire Method	BNA_F
ICV/I.BLK	SP5813	HP Processing Method	bf042222
Surrogate Standard	SP5822,SP5811	MS/MSD Standard	SP5854
LCS Standard	SP5854		

Sr#	SampleId	Data File Name	Date-Time	Comment	Operator	Status
22	N2526-07	BF127890.D	25 Apr 2022 19:52		CG\JU	Ok
23	N2532-01	BF127891.D	25 Apr 2022 20:20		CG\JU	Ok
24	N2548-10	BF127892.D	25 Apr 2022 20:49		CG\JU	Ok
25	N2466-05	BF127893.D	25 Apr 2022 21:18		CG\JU	Ok,M
26	N2548-04	BF127894.D	25 Apr 2022 21:46		CG\JU	Ok
27	N2466-03	BF127895.D	25 Apr 2022 22:14		CG\JU	Ok,M
28	N2466-04	BF127896.D	25 Apr 2022 22:43	Not Match DUP with original(N2466-03)	CG\JU	Not Ok
29	N2541-01	BF127897.D	25 Apr 2022 23:11		CG\JU	Ok,M
30	N2476-01DL	BF127898.D	25 Apr 2022 23:40		CG\JU	Ok,M
31	N2476-05DL	BF127899.D	26 Apr 2022 00:08		CG\JU	Ok,M
32	N2476-03DL	BF127900.D	26 Apr 2022 00:36		CG\JU	Ok,M
33	N2523-01	BF127901.D	26 Apr 2022 01:05		CG\JU	Ok
34	N2523-03	BF127902.D	26 Apr 2022 01:33		CG\JU	Ok,M

M : Manual Integration