

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042522\
 Data File : BF127880.D
 Acq On : 25 Apr 2022 15:04
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 25 16:56:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	40.010	-0.0	93	0.00
3	Pyridine	40.000	39.582	1.0	89	-0.01
4	n-Nitrosodimethylamine	40.000	38.540	3.7	89	-0.01
5 S	2-Fluorophenol	80.000	78.485	1.9	93	0.00
6	Aniline	40.000	39.817	0.5	90	0.00
7 S	Phenol-d6	80.000	78.443	1.9	92	0.00
8	2-Chlorophenol	40.000	39.401	1.5	92	0.00
9	Benzaldehyde	40.000	34.974	12.6	94	0.00
10 C	Phenol	40.000	39.650	0.9	93	0.00
11	bis(2-Chloroethyl)ether	40.000	39.439	1.4	92	0.00
12	1,3-Dichlorobenzene	40.000	38.718	3.2	92	0.00
13 C	1,4-Dichlorobenzene	40.000	39.020	2.4	93	0.00
14	1,2-Dichlorobenzene	40.000	38.808	3.0	93	0.00
15	Benzyl Alcohol	40.000	42.963	-7.4	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.501	-1.3	93	0.00
17	2-Methylphenol	40.000	38.359	4.1	86	0.00
18	Hexachloroethane	40.000	39.219	2.0	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.752	3.1	91	0.00
20	3+4-Methylphenols	40.000	40.266	-0.7	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	38.562	3.6	92	0.00
23 S	Nitrobenzene-d5	80.000	78.125	2.3	90	0.00
24	Nitrobenzene	40.000	39.071	2.3	90	0.00
25	Isophorone	40.000	38.973	2.6	90	0.00
26 C	2-Nitrophenol	40.000	41.376	-3.4	92	0.00
27	2,4-Dimethylphenol	40.000	38.903	2.7	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.860	2.9	90	0.00
29 C	2,4-Dichlorophenol	40.000	39.864	0.3	92	0.00
30	1,2,4-Trichlorobenzene	40.000	38.784	3.0	91	0.00
31	Naphthalene	40.000	38.324	4.2	90	0.00
32	Benzoic acid	40.000	42.860	-7.1	91	0.00
33	4-Chloroaniline	40.000	39.156	2.1	89	0.00
34 C	Hexachlorobutadiene	40.000	39.626	0.9	93	0.00
35	Caprolactam	40.000	38.151	4.6	85	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.129	2.2	89	0.00
37	2-Methylnaphthalene	40.000	38.819	3.0	91	0.00
38	1-Methylnaphthalene	40.000	38.541	3.6	91	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.379	1.6	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.998	-5.0	92	0.00
42 S	2,4,6-Tribromophenol	80.000	80.240	-0.3	93	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.733	0.7	91	0.00
44	2,4,5-Trichlorophenol	40.000	39.966	0.1	89	0.00
45 S	2-Fluorobiphenyl	80.000	79.738	0.3	91	0.00
46	1,1'-Biphenyl	40.000	38.878	2.8	91	0.00
47	2-Chloronaphthalene	40.000	38.646	3.4	90	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042522\
 Data File : BF127880.D
 Acq On : 25 Apr 2022 15:04
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 25 16:56:50 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.318	-0.8	90	0.00
49	Acenaphthylene	40.000	39.381	1.5	91	0.00
50	Dimethylphthalate	40.000	38.541	3.6	89	0.00
51	2,6-Dinitrotoluene	40.000	39.423	1.4	88	0.00
52 C	Acenaphthene	40.000	39.461	1.3	90	0.00
53	3-Nitroaniline	40.000	38.437	3.9	85	0.00
54 P	2,4-Dinitrophenol	40.000	45.247	-13.1	96	0.00
55	Dibenzofuran	40.000	38.890	2.8	90	0.00
56 P	4-Nitrophenol	40.000	39.310	1.7	85	0.00
57	2,4-Dinitrotoluene	40.000	39.743	0.6	87	0.00
58	Fluorene	40.000	38.145	4.6	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.766	0.6	89	0.00
60	Diethylphthalate	40.000	38.293	4.3	87	0.00
61	4-Chlorophenyl-phenylether	40.000	39.239	1.9	92	0.00
62	4-Nitroaniline	40.000	38.257	4.4	84	0.00
63	Azobenzene	40.000	38.495	3.8	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.216	-8.0	85	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.287	1.8	89	0.00
67	4-Bromophenyl-phenylether	40.000	40.605	-1.5	91	0.00
68	Hexachlorobenzene	40.000	39.994	0.0	90	0.00
69	Atrazine	40.000	37.761	5.6	86	0.00
70 C	Pentachlorophenol	40.000	41.804	-4.5	87	0.00
71	Phenanthrene	40.000	38.113	4.7	86	0.00
72	Anthracene	40.000	38.782	3.0	87	0.00
73	Carbazole	40.000	37.715	5.7	85	0.00
74	Di-n-butylphthalate	40.000	38.615	3.5	86	0.00
75 C	Fluoranthene	40.000	37.653	5.9	85	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzidine	40.000	23.228	41.9#	45	0.00
78	Pyrene	40.000	40.411	-1.0	84	0.00
79 S	Terphenyl-d14	80.000	80.233	-0.3	86	0.00
80	Butylbenzylphthalate	40.000	40.606	-1.5	81	0.00
81	Benzo(a)anthracene	40.000	39.260	1.9	80	0.00
82	3,3'-Dichlorobenzidine	40.000	38.941	2.6	81	0.00
83	Chrysene	40.000	39.549	1.1	83	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.120	-0.3	81	0.00
85 c	Di-n-octyl phthalate	40.000	38.681	3.3	77	0.00
86 I	Perylene-d12	20.000	20.000	0.0	84	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.836	-9.6	87	0.00
88	Benzo(b)fluoranthene	40.000	39.163	2.1	82	0.00
89	Benzo(k)fluoranthene	40.000	38.505	3.7	77	0.00
90 C	Benzo(a)pyrene	40.000	39.546	1.1	80	0.00
91	Dibenzo(a,h)anthracene	40.000	44.351	-10.9	87	0.00
92	Benzo(g,h,i)perylene	40.000	45.112	-12.8	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042522\
Data File : BF127880.D
Acq On : 25 Apr 2022 15:04
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 25 16:56:50 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 00:43:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0