

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042523\
 Data File : BF133049.D
 Acq On : 26 Apr 2023 02:35
 Operator : CG\JU
 Sample : 02484-01 5X
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CRCA-1

Quant Time: Apr 26 03:44:53 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 25 18:24:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.740	152	215982	20.000	ng	0.00
21) Naphthalene-d8	8.022	136	805826	20.000	ng	0.00
39) Acenaphthene-d10	9.775	164	400742	20.000	ng	0.00
64) Phenanthrene-d10	11.257	188	607282	20.000	ng	0.00
76) Chrysene-d12	13.892	240	431488	20.000	ng	0.00
86) Perylene-d12	15.315	264	360160	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	43238	3.024	ng	-0.01
7) Phenol-d6	6.369	99	199615	11.248	ng	-0.02
23) Nitrobenzene-d5	7.298	82	146924	9.841	ng	-0.01
42) 2,4,6-Tribromophenol	10.557	330	2105	0.522	ng	-0.01
45) 2-Fluorobiphenyl	9.098	172	303216	12.659	ng	-0.01
79) Terphenyl-d14	12.839	244	254299	10.164	ng	-0.01


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042523\
Data File : BF133049.D
Acq On : 26 Apr 2023 02:35
Operator : CG\JU
Sample : 02484-01 5X
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CRCA-1

Quant Time: Apr 26 03:44:53 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 25 18:24:07 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.740 min Scan# 793
Delta R.T. -0.005 min
Lab File: BF133049.D
Acq: 26 Apr 2023 02:35




Instrument :
BNA_F
ClientSampleId :
CRCA-1

Tgt Ion:152 Resp: 215982
Ion Ratio Lower Upper
152 100
150 156.4 124.8 187.2
115 61.4 49.9 74.9



#5
2-Fluorophenol
Concen: 3.024 ng
RT: 5.340 min Scan# 553
Delta R.T. -0.012 min
Lab File: BF133049.D
Acq: 26 Apr 2023 02:35

Tgt Ion:112 Resp: 43238
Ion Ratio Lower Upper
112 100
64 62.5 46.8 70.2
63 31.7 23.8 35.8








#7
Phenol-d6
Concen: 11.248 ng
RT: 6.369 min Scan# 71
Delta R.T. -0.017 min
Lab File: BF133049.D
Acq: 26 Apr 2023 02:35




Instrument :
BNA_F
ClientSampleId :
CRCA-1


Tgt Ion: 99 Resp: 199615
Ion Ratio Lower Upper
99 100
42 20.2 16.3 24.5
71 32.5 25.6 38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.022 min Scan# 1009
Delta R.T. -0.006 min
Lab File: BF133049.D
Acq: 26 Apr 2023 02:35

Tgt Ion: 136 Resp: 805826
Ion Ratio Lower Upper
136 100
137 11.7 8.6 13.0
54 10.0 7.0 10.4
68 7.4 5.4 8.2





#23

Nitrobenzene-d5

Concen: 9.841 ng

RT: 7.298 min Scan# 81

Delta R.T. -0.012 min

Lab File: BF133049.D

Acq: 26 Apr 2023 02:35

Instrument :

BNA_F

ClientSampleId :

CRCA-1





Tgt Ion: 82 Resp: 146924

Ion Ratio Lower Upper

82 100

128 49.5 37.7 56.5

54 52.5 43.0 64.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.775 min Scan# 1307

Delta R.T. -0.006 min

Lab File: BF133049.D

Acq: 26 Apr 2023 02:35




Tgt Ion: 164 Resp: 400742

Ion Ratio Lower Upper

164 100

162 97.6 78.7 118.1

160 44.0 35.8 53.6





#42

2,4,6-Tribromophenol

Concen: 0.522 ng

RT: 10.557 min Scan# 1444

Delta R.T. -0.012 min

Lab File: BF133049.D

Acq: 26 Apr 2023 02:35

Instrument :

BNA_F

ClientSampleId :

CRCA-1





Tgt Ion:330 Resp: 2105

Ion Ratio Lower Upper

330 100

332 90.7 77.1 115.7

141 48.0 25.8 38.8#



#45

2-Fluorobiphenyl

Concen: 12.659 ng

RT: 9.098 min Scan# 1192

Delta R.T. -0.012 min

Lab File: BF133049.D

Acq: 26 Apr 2023 02:35




Tgt Ion:172 Resp: 303216


Ion Ratio Lower Upper

172 100

171 34.7 28.8 43.2

170 23.5 19.4 29.2





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.257 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF133049.D

Acq: 26 Apr 2023 02:35

Instrument :

BNA_F

ClientSampleId :

CRCA-1





Tgt Ion:188 Resp: 607282

Ion Ratio Lower Upper

188 100

94 10.9 9.7 14.5

80 11.8 10.4 15.6



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.892 min Scan# 2007

Delta R.T. -0.006 min

Lab File: BF133049.D

Acq: 26 Apr 2023 02:35




Tgt Ion:240 Resp: 431488


Ion Ratio Lower Upper

240 100

120 13.2 10.5 15.7

236 25.1 21.5 32.3









#79
 Terphenyl-d14
 Concen: 10.164 ng
 RT: 12.839 min Scan# 1828
 Delta R.T. -0.012 min
 Lab File: BF133049.D
 Acq: 26 Apr 2023 02:35

Instrument :
 BNA_F
 ClientSampleId :
 CRCA-1

Tgt Ion:244 Resp: 254299
 Ion Ratio Lower Upper
 244 100
 212 6.6 5.1 7.7
 122 13.8 8.6 13.0#



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.315 min Scan# 2249
 Delta R.T. 0.000 min
 Lab File: BF133049.D
 Acq: 26 Apr 2023 02:35

Tgt Ion:264 Resp: 360160
 Ion Ratio Lower Upper
 264 100
 260 23.8 18.8 28.2
 265 21.2 17.0 25.4

