


Data Path : Z:\HPCHEM1\BNA F\Data\BF042617\
 Data File : BF094639.D
 Acq On : 26 Apr 2017 11:35
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF041717.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 17 14:59:23 2017



AutoFind: Scans 1641, 1642, 1643; Background Corrected with Scan 1636

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.8	24629	PASS
68	69	0.00	2	1.6	385	PASS
69	198	0.00	100	34.9	24645	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.8	34525	PASS
197	198	0.00	2	0.4	300	PASS
198	198	100	100	100.0	70685	PASS
199	198	5	9	6.5	4618	PASS
275	198	10	60	22.8	16137	PASS
365	198	1	100	2.3	1607	PASS
441	442	0.01	24	15.9	9568	PASS
442	198	50	100	84.9	60002	PASS
443	442	15	24	19.4	11666	PASS