

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042622\
 Data File : BF127904.D
 Acq On : 26 Apr 2022 09:52
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 01:45:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.599	0.577	3.7	103	0.03
3	Pyridine	1.535	1.553	-1.2	106	0.02
4	n-Nitrosodimethylamine	0.743	0.749	-0.8	107	0.03
5 S	2-Fluorophenol	1.256	1.215	3.3	107	0.00
6	Aniline	1.936	2.012	-3.9	109	0.00
7 S	Phenol-d6	1.629	1.635	-0.4	109	0.00
8	2-Chlorophenol	1.303	1.300	0.2	107	0.00
9	Benzaldehyde	1.008	0.908	9.9	111	0.00
10 C	Phenol	1.643	1.655	-0.7	109	0.00
11	bis(2-Chloroethyl)ether	1.363	1.383	-1.5	110	0.00
12	1,3-Dichlorobenzene	1.500	1.456	2.9	106	0.00
13 C	1,4-Dichlorobenzene	1.497	1.457	2.7	107	0.00
14	1,2-Dichlorobenzene	1.386	1.330	4.0	106	0.00
15	Benzyl Alcohol	1.108	1.170	-5.6	108	0.01
16	2,2'-oxybis(1-Chloropropane	2.083	2.159	-3.6	110	0.00
17	2-Methylphenol	1.124	1.231	-9.5	114	0.01
18	Hexachloroethane	0.550	0.558	-1.5	110	0.00
19 P	n-Nitroso-di-n-propylamine	0.997	1.023	-2.6	111	0.00
20	3+4-Methylphenols	1.358	1.367	-0.7	107	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
22	Acetophenone	0.487	0.468	3.9	109	0.00
23 S	Nitrobenzene-d5	0.425	0.426	-0.2	111	0.00
24	Nitrobenzene	0.410	0.403	1.7	109	0.00
25	Isophorone	0.716	0.726	-1.4	112	0.00
26 C	2-Nitrophenol	0.175	0.190	-8.6	116	0.00
27	2,4-Dimethylphenol	0.290	0.292	-0.7	108	0.00
28	bis(2-Chloroethoxy)methane	0.459	0.465	-1.3	113	0.00
29 C	2,4-Dichlorophenol	0.304	0.305	-0.3	111	0.00
30	1,2,4-Trichlorobenzene	0.345	0.337	2.3	110	0.00
31	Naphthalene	1.019	0.979	3.9	109	0.00
32	Benzoic acid	0.208	0.167	19.7	82	0.01
33	4-Chloroaniline	0.403	0.407	-1.0	111	0.00
34 C	Hexachlorobutadiene	0.217	0.218	-0.5	113	0.00
35	Caprolactam	0.098	0.097	1.0	105	0.01
36 C	4-Chloro-3-methylphenol	0.325	0.331	-1.8	111	0.00
37	2-Methylnaphthalene	0.673	0.660	1.9	111	0.00
38	1-Methylnaphthalene	0.654	0.631	3.5	109	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	114	0.00
40	1,2,4,5-Tetrachlorobenzene	0.590	0.572	3.1	110	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.339	-6.3	113	0.00
42 S	2,4,6-Tribromophenol	0.201	0.205	-2.0	115	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.393	-1.3	113	0.00
44	2,4,5-Trichlorophenol	0.422	0.414	1.9	107	0.01
45 S	2-Fluorobiphenyl	1.182	1.174	0.7	111	0.00
46	1,1'-Biphenyl	1.481	1.428	3.6	110	0.00
47	2-Chloronaphthalene	1.128	1.088	3.5	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042622\
 Data File : BF127904.D
 Acq On : 26 Apr 2022 09:52
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 01:45:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.381	0.398	-4.5	113	0.00
49	Acenaphthylene	1.725	1.665	3.5	108	0.00
50	Dimethylphthalate	1.341	1.318	1.7	111	0.00
51	2,6-Dinitrotoluene	0.290	0.292	-0.7	109	0.00
52 C	Acenaphthene	1.160	1.144	1.4	110	0.00
53	3-Nitroaniline	0.322	0.317	1.6	106	0.00
54 P	2,4-Dinitrophenol	0.153	0.189	-23.5	128	0.00
55	Dibenzofuran	1.673	1.595	4.7	108	0.00
56 P	4-Nitrophenol	0.246	0.233	5.3	100	0.01
57	2,4-Dinitrotoluene	0.385	0.379	1.6	105	0.00
58	Fluorene	1.249	1.178	5.7	108	0.00
59	2,3,4,6-Tetrachlorophenol	0.355	0.352	0.8	109	0.01
60	Diethylphthalate	1.315	1.264	3.9	107	0.00
61	4-Chlorophenyl-phenylether	0.620	0.605	2.4	112	0.00
62	4-Nitroaniline	0.315	0.305	3.2	104	0.00
63	Azobenzene	1.447	1.404	3.0	108	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	107	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.136	-14.3	108	0.01
66 c	n-Nitrosodiphenylamine	0.622	0.620	0.3	107	0.00
67	4-Bromophenyl-phenylether	0.222	0.226	-1.8	109	0.00
68	Hexachlorobenzene	0.234	0.237	-1.3	109	0.01
69	Atrazine	0.192	0.188	2.1	106	0.00
70 C	Pentachlorophenol	0.137	0.141	-2.9	102	0.01
71	Phenanthrene	1.047	1.004	4.1	103	0.00
72	Anthracene	1.043	1.002	3.9	103	0.00
73	Carbazole	1.005	0.940	6.5	101	0.00
74	Di-n-butylphthalate	1.130	1.125	0.4	106	0.00
75 C	Fluoranthene	1.102	1.025	7.0	100	0.01
76 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
77	Benzidine	0.378	0.417	-10.3	100	0.00
78	Pyrene	1.614	1.625	-0.7	98	0.00
79 S	Terphenyl-d14	1.094	1.092	0.2	101	0.00
80	Butylbenzylphthalate	0.633	0.690	-9.0	103	0.00
81	Benzo(a)anthracene	1.333	1.306	2.0	95	0.00
82	3,3'-Dichlorobenzidine	0.423	0.395	6.6	91	0.00
83	Chrysene	1.273	1.223	3.9	95	0.01
84	Bis(2-ethylhexyl)phthalate	0.839	0.913	-8.8	104	0.00
85 c	Di-n-octyl phthalate	1.300	1.395	-7.3	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	96	0.01
87	Indeno(1,2,3-cd)pyrene	1.335	1.340	-0.4	91	0.02
88	Benzo(b)fluoranthene	1.256	1.226	2.4	94	0.01
89	Benzo(k)fluoranthene	1.266	1.242	1.9	91	0.00
90 C	Benzo(a)pyrene	1.042	1.032	1.0	92	0.02
91	Dibenzo(a,h)anthracene	1.071	1.084	-1.2	91	0.02
92	Benzo(g,h,i)perylene	1.128	1.117	1.0	89	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042622\
Data File : BF127904.D
Acq On : 26 Apr 2022 09:52
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 27 01:45:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 00:43:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0