

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042622\
 Data File : BF127914.D
 Acq On : 26 Apr 2022 15:39
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 01:47:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	113	0.00
2	1,4-Dioxane	40.000	38.804	3.0	106	0.05
3	Pyridine	40.000	40.452	-1.1	108	0.05
4	n-Nitrosodimethylamine	40.000	40.639	-1.6	111	0.04
5 S	2-Fluorophenol	80.000	77.303	3.4	109	0.00
6	Aniline	40.000	41.139	-2.8	111	0.00
7 S	Phenol-d6	80.000	79.740	0.3	111	0.00
8	2-Chlorophenol	40.000	39.832	0.4	110	0.00
9	Benzaldehyde	40.000	35.092	12.3	111	0.00
10 C	Phenol	40.000	40.110	-0.3	111	0.01
11	bis(2-Chloroethyl)ether	40.000	40.296	-0.7	112	0.00
12	1,3-Dichlorobenzene	40.000	38.762	3.1	109	0.00
13 C	1,4-Dichlorobenzene	40.000	38.708	3.2	109	0.00
14	1,2-Dichlorobenzene	40.000	37.301	6.7	106	0.00
15	Benzyl Alcohol	40.000	45.591	-14.0	119	0.01
16	2,2'-oxybis(1-Chloropropane	40.000	40.335	-0.8	109	0.00
17	2-Methylphenol	40.000	42.816	-7.0	114	0.01
18	Hexachloroethane	40.000	40.496	-1.2	113	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.110	-2.8	114	0.00
20	3+4-Methylphenols	40.000	39.202	2.0	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	114	0.00
22	Acetophenone	40.000	38.032	4.9	110	0.00
23 S	Nitrobenzene-d5	80.000	80.549	-0.7	113	0.00
24	Nitrobenzene	40.000	39.653	0.9	111	0.00
25	Isophorone	40.000	40.361	-0.9	113	0.00
26 C	2-Nitrophenol	40.000	43.544	-8.9	117	0.00
27	2,4-Dimethylphenol	40.000	40.508	-1.3	110	0.01
28	bis(2-Chloroethoxy)methane	40.000	40.146	-0.4	113	0.00
29 C	2,4-Dichlorophenol	40.000	40.000	0.0	112	0.00
30	1,2,4-Trichlorobenzene	40.000	39.019	2.5	111	0.00
31	Naphthalene	40.000	38.260	4.4	109	0.00
32	Benzoic acid	40.000	41.509	-3.8	106	0.02
33	4-Chloroaniline	40.000	39.541	1.1	110	0.00
34 C	Hexachlorobutadiene	40.000	40.107	-0.3	114	0.00
35	Caprolactam	40.000	40.337	-0.8	109	0.02
36 C	4-Chloro-3-methylphenol	40.000	40.669	-1.7	113	0.00
37	2-Methylnaphthalene	40.000	38.585	3.5	110	0.00
38	1-Methylnaphthalene	40.000	38.523	3.7	110	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	115	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.540	3.7	110	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.128	-7.8	116	0.00
42 S	2,4,6-Tribromophenol	80.000	82.172	-2.7	117	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.801	-2.0	115	0.00
44	2,4,5-Trichlorophenol	40.000	40.649	-1.6	112	0.01
45 S	2-Fluorobiphenyl	80.000	79.488	0.6	112	0.00
46	1,1'-Biphenyl	40.000	38.330	4.2	110	0.00
47	2-Chloronaphthalene	40.000	38.309	4.2	110	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042622\
 Data File : BF127914.D
 Acq On : 26 Apr 2022 15:39
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 27 01:47:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Apr 24 00:43:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.781	-4.5	114	0.00
49	Acenaphthylene	40.000	38.392	4.0	109	0.00
50	Dimethylphthalate	40.000	39.259	1.9	112	0.00
51	2,6-Dinitrotoluene	40.000	40.243	-0.6	110	0.00
52 C	Acenaphthene	40.000	39.697	0.8	112	0.00
53	3-Nitroaniline	40.000	39.577	1.1	108	0.01
54 P	2,4-Dinitrophenol	40.000	52.928	-32.3#	138	0.00
55	Dibenzofuran	40.000	37.674	5.8	107	0.00
56 P	4-Nitrophenol	40.000	37.956	5.1	101	0.01
57	2,4-Dinitrotoluene	40.000	39.484	1.3	106	0.00
58	Fluorene	40.000	37.539	6.2	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.768	0.6	110	0.01
60	Diethylphthalate	40.000	38.890	2.8	109	0.00
61	4-Chlorophenyl-phenylether	40.000	38.465	3.8	111	0.00
62	4-Nitroaniline	40.000	39.803	0.5	108	0.01
63	Azobenzene	40.000	38.702	3.2	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.842	-17.1	113	0.01
66 c	n-Nitrosodiphenylamine	40.000	39.183	2.0	108	0.00
67	4-Bromophenyl-phenylether	40.000	39.858	0.4	110	0.00
68	Hexachlorobenzene	40.000	40.303	-0.8	111	0.00
69	Atrazine	40.000	38.345	4.1	107	0.01
70 C	Pentachlorophenol	40.000	41.482	-3.7	106	0.00
71	Phenanthrene	40.000	37.663	5.8	104	0.00
72	Anthracene	40.000	38.116	4.7	105	0.00
73	Carbazole	40.000	37.353	6.6	103	0.00
74	Di-n-butylphthalate	40.000	39.460	1.3	108	0.00
75 C	Fluoranthene	40.000	37.182	7.0	103	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	44.913	-12.3	104	0.00
78	Pyrene	40.000	40.544	-1.4	101	0.00
79 S	Terphenyl-d14	80.000	81.140	-1.4	105	0.00
80	Butylbenzylphthalate	40.000	44.659	-11.6	108	0.00
81	Benzo(a)anthracene	40.000	39.456	1.4	98	0.00
82	3,3'-Dichlorobenzidine	40.000	38.565	3.6	96	0.00
83	Chrysene	40.000	38.534	3.7	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.831	-7.1	104	0.00
85 c	Di-n-octyl phthalate	40.000	41.999	-5.0	101	0.00
86 I	Perylene-d12	20.000	20.000	0.0	95	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.446	1.4	88	0.02
88	Benzo(b)fluoranthene	40.000	40.284	-0.7	95	0.00
89	Benzo(k)fluoranthene	40.000	38.116	4.7	87	0.00
90 C	Benzo(a)pyrene	40.000	39.163	2.1	90	0.01
91	Dibenzo(a,h)anthracene	40.000	40.277	-0.7	89	0.02
92	Benzo(g,h,i)perylene	40.000	39.451	1.4	87	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042622\
Data File : BF127914.D
Acq On : 26 Apr 2022 15:39
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 27 01:47:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Apr 24 00:43:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0