

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042716\
 Data File : BF086462.D
 Acq On : 28 Apr 2016 9:30
 Operator : UM/SJ
 Sample : H2736-08
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RICHMOND-AMBOY-SB05(0.5-10.0)

Manual Integrations
 APPROVED

sohil
 4/28/2016 2:14:11 PM

Quant Time: Apr 28 11:43:27 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 27 14:20:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	127170	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	473842	20.00	ng	-0.01
38) Acenaphthene-d10	9.85	164	170940	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	274191	20.00	ng	0.00
75) Chrysene-d12	13.95	240	287502	20.00	ng	-0.01
86) Perylene-d12	15.37	264	207727	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	1028462	139.50	ng	0.01
7) Phenol-d6	6.44	99	1366724	132.82	ng	0.00
23) Nitrobenzene-d5	7.37	82	833842	94.85	ng	-0.01
41) 2,4,6-Tribromophenol	10.64	330	221051	122.62	ng	0.00
44) 2-Fluorobiphenyl	9.17	172	1285287	153.81	ng	0.00
78) Terphenyl-d14	12.91	244	860083	65.91	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.56	163	253705	19.71	ng	99
74) Fluoranthene	12.53	202	65567	4.51	ng	95
77) Pyrene	12.76	202	61648	2.98	ng	98
80) Benzo(a)anthracene	13.94	228	33929	2.21	ng	99
83) Bis(2-ethylhexyl)phthalate	13.95	149	37133	3.33	ng	# 97
87) Benzo(b)fluoranthene	14.97	252	27716m	2.27	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042716\
Data File : BF086462.D
Acq On : 28 Apr 2016 9:30
Operator : UM/SJ
Sample : H2736-08
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RICHMOND-AMBOY-SB05(0.5-10.0)

Manual Integrations
APPROVED

sohil
4/28/2016 2:14:11 PM

Quant Time: Apr 28 11:43:27 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 27 14:20:52 2016
Response via : Initial Calibration

