

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042717\
 Data File : BF094676.D
 Acq On : 27 Apr 2017 21:03
 Operator : SJ/MA
 Sample : I2870-07
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PURGE-WATER-DRUMS(COMP)

Quant Time: Apr 28 05:52:29 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 17 14:59:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.77	152	105424	20.00	ng	0.00
21) Naphthalene-d8	7.26	136	434738	20.00	ng	0.00
38) Acenaphthene-d10	9.40	164	206712	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	361379	20.00	ng	0.00
75) Chrysene-d12	14.47	240	279667	20.00	ng	0.00
86) Perylene-d12	16.09	264	205017	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.32	112	435301	68.50	ng	0.02
7) Phenol-d6	5.39	99	334946	44.71	ng	0.00
23) Nitrobenzene-d5	6.42	82	689433	97.92	ng	0.00
41) 2,4,6-Tribromophenol	10.38	330	245525	151.85	ng	0.00
44) 2-Fluorobiphenyl	8.59	172	1298592	92.43	ng	0.00
78) Terphenyl-d14	13.20	244	1032370	98.67	ng	0.00
Target Compounds						Qvalue
8) 2-Chlorophenol	5.54	128	21905	3.03	ng	96
12) 1,3-Dichlorobenzene	5.70	146	21107	2.63	ng	96
13) 1,4-Dichlorobenzene	5.79	146	85195	10.57	ng	99
14) 1,2-Dichlorobenzene	5.96	146	28640	3.86	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042717\
Data File : BF094676.D
Acq On : 27 Apr 2017 21:03
Operator : SJ/MA
Sample : I2870-07
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PURGE-WATER-DRUMS(COMP)

Quant Time: Apr 28 05:52:29 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 17 14:59:23 2017
Response via : Initial Calibration

