

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042816\
 Data File : BF086479.D
 Acq On : 28 Apr 2016 22:51
 Operator : UM/SJ
 Sample : H2654-08
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 26WARD-2A-CS-9(0-6FTBGS)

Manual Integrations
 APPROVED

sohil
 4/29/2016 4:52:51 PM

Quant Time: Apr 29 01:54:19 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 27 14:20:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.78	152	169771	20.00	ng	-0.02
21) Naphthalene-d8	8.08	136	693590	20.00	ng	-0.02
38) Acenaphthene-d10	9.82	164	304153	20.00	ng	-0.02
63) Phenanthrene-d10	11.31	188	489161	20.00	ng	-0.01
75) Chrysene-d12	13.94	240	316297	20.00	ng	-0.02
86) Perylene-d12	15.33	264	336236	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.39	112	1135987	115.42	ng	0.00
7) Phenol-d6	6.43	99	1555090	113.20	ng	-0.01
23) Nitrobenzene-d5	7.36	82	1020499	79.31	ng	-0.02
41) 2,4,6-Tribromophenol	10.61	330	289262	90.18	ng	-0.02
44) 2-Fluorobiphenyl	9.15	172	1305442	72.68	ng	-0.02
78) Terphenyl-d14	12.89	244	861954	60.04	ng	-0.02
Target Compounds						
49) Dimethylphthalate	9.55	163	169882	7.42	ng	98
70) Phenanthrene	11.33	178	135123	5.26	ng	99
74) Fluoranthene	12.51	202	236692	9.13	ng	97
77) Pyrene	12.74	202	204518	8.97	ng	98
80) Benzo(a)anthracene	13.93	228	113601	6.74	ng	98
82) Chrysene	13.96	228	76510m	4.17	ng	
83) Bis(2-ethylhexyl)phthalate	13.93	149	35689	2.91	ng	# 98
85) Indeno(1,2,3-cd)pyrene	16.68	276	55699	4.10	ng	96
87) Benzo(b)fluoranthene	14.95	252	128034m	6.47	ng	
88) Benzo(k)fluoranthene	14.96	252	52207m	2.53	ng	
89) Benzo(a)pyrene	15.28	252	91477	4.90	ng	# 92
91) Benzo(a,h,i)perylene	17.08	276	49175	3.21	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042816\
Data File : BF086479.D
Acq On : 28 Apr 2016 22:51
Operator : UM/SJ
Sample : H2654-08
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
26WARD-2A-CS-9(0-6FTBGS)

Manual Integrations
APPROVED

sohil
4/29/2016 4:52:51 PM

Quant Time: Apr 29 01:54:19 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 27 14:20:52 2016
Response via : Initial Calibration

