

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042821\
 Data File : BF123752.D
 Acq On : 28 Apr 2021 9:19
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 28 11:18:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 16:30:39 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	40.091	-0.2	100	0.00
3	Pyridine	40.000	40.547	-1.4	101	0.01
4	n-Nitrosodimethylamine	40.000	40.534	-1.3	99	0.01
5 S	2-Fluorophenol	80.000	77.253	3.4	97	0.00
6	Aniline	40.000	39.188	2.0	97	0.00
7 S	Phenol-d6	80.000	77.487	3.1	99	0.00
8	2-Chlorophenol	40.000	38.652	3.4	97	0.00
9	Benzaldehyde	40.000	38.384	4.0	100	0.00
10 C	Phenol	40.000	38.632	3.4	97	0.00
11	bis(2-Chloroethyl)ether	40.000	38.492	3.8	96	0.00
12	1,3-Dichlorobenzene	40.000	39.202	2.0	98	0.00
13 C	1,4-Dichlorobenzene	40.000	37.959	5.1	96	0.00
14	1,2-Dichlorobenzene	40.000	36.700	8.2	98	0.00
15	Benzyl Alcohol	40.000	39.742	0.6	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.605	3.5	96	0.00
17	2-Methylphenol	40.000	39.058	2.4	98	0.00
18	Hexachloroethane	40.000	39.153	2.1	96	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.637	3.4	98	0.00
20	3+4-Methylphenols	40.000	37.481	6.3	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	38.153	4.6	98	0.00
23 S	Nitrobenzene-d5	80.000	78.018	2.5	97	0.00
24	Nitrobenzene	40.000	39.225	1.9	98	0.00
25	Isophorone	40.000	38.155	4.6	95	0.00
26 C	2-Nitrophenol	40.000	39.721	0.7	96	0.00
27	2,4-Dimethylphenol	40.000	38.776	3.1	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.322	4.2	95	0.00
29 C	2,4-Dichlorophenol	40.000	38.822	2.9	97	0.00
30	1,2,4-Trichlorobenzene	40.000	38.165	4.6	97	0.00
31	Naphthalene	40.000	38.523	3.7	97	0.00
32	Benzoic acid	40.000	34.419	14.0	89	0.00
33	4-Chloroaniline	40.000	38.868	2.8	98	0.00
34 C	Hexachlorobutadiene	40.000	38.318	4.2	94	0.00
35	Caprolactam	40.000	38.468	3.8	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.607	1.0	97	0.00
37	2-Methylnaphthalene	40.000	38.292	4.3	97	0.00
38	1-Methylnaphthalene	40.000	38.499	3.8	98	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.004	2.5	97	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.641	3.4	94	0.00
42 S	2,4,6-Tribromophenol	80.000	78.143	2.3	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.970	0.1	96	0.00
44	2,4,5-Trichlorophenol	40.000	38.936	2.7	94	0.00
45 S	2-Fluorobiphenyl	80.000	75.453	5.7	97	0.00
46	1,1'-Biphenyl	40.000	37.876	5.3	95	0.00
47	2-Chloronaphthalene	40.000	38.224	4.4	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042821\
 Data File : BF123752.D
 Acq On : 28 Apr 2021 9:19
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 28 11:18:57 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 16:30:39 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.408	-1.0	99	0.00
49	Acenaphthylene	40.000	38.885	2.8	97	0.00
50	Dimethylphthalate	40.000	38.135	4.7	96	0.00
51	2,6-Dinitrotoluene	40.000	39.393	1.5	97	0.00
52 C	Acenaphthene	40.000	39.027	2.4	98	0.00
53	3-Nitroaniline	40.000	38.637	3.4	95	0.00
54 P	2,4-Dinitrophenol	40.000	37.909	5.2	87	0.00
55	Dibenzofuran	40.000	38.330	4.2	95	0.00
56 P	4-Nitrophenol	40.000	39.617	1.0	93	0.00
57	2,4-Dinitrotoluene	40.000	38.794	3.0	95	0.00
58	Fluorene	40.000	38.101	4.7	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.995	0.0	96	0.00
60	Diethylphthalate	40.000	37.979	5.1	95	0.00
61	4-Chlorophenyl-phenylether	40.000	37.749	5.6	95	0.00
62	4-Nitroaniline	40.000	38.788	3.0	94	0.00
63	Azobenzene	40.000	38.589	3.5	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.936	0.2	91	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.506	3.7	94	0.00
67	4-Bromophenyl-phenylether	40.000	39.787	0.5	97	0.00
68	Hexachlorobenzene	40.000	38.833	2.9	97	0.00
69	Atrazine	40.000	38.642	3.4	95	0.00
70 C	Pentachlorophenol	40.000	41.323	-3.3	99	0.00
71	Phenanthrene	40.000	38.840	2.9	96	0.00
72	Anthracene	40.000	38.978	2.6	97	0.00
73	Carbazole	40.000	38.962	2.6	97	0.00
74	Di-n-butylphthalate	40.000	38.912	2.7	94	0.00
75 C	Fluoranthene	40.000	38.403	4.0	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	96	0.00
77	Benzidine	40.000	44.038	-10.1	98	0.00
78	Pyrene	40.000	39.465	1.3	97	0.00
79 S	Terphenyl-d14	80.000	76.191	4.8	95	0.00
80	Butylbenzylphthalate	40.000	39.455	1.4	95	0.00
81	Benzo(a)anthracene	40.000	39.722	0.7	98	0.00
82	3,3'-Dichlorobenzidine	40.000	40.546	-1.4	98	0.00
83	Chrysene	40.000	39.448	1.4	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.058	-0.1	98	0.00
85 c	Di-n-octyl phthalate	40.000	40.486	-1.2	97	0.00
86 I	Perylene-d12	20.000	20.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.231	1.9	107	0.00
88	Benzo(b)fluoranthene	40.000	38.013	5.0	99	0.00
89	Benzo(k)fluoranthene	40.000	39.733	0.7	99	0.00
90 C	Benzo(a)pyrene	40.000	39.651	0.9	101	0.00
91	Dibenzo(a,h)anthracene	40.000	38.986	2.5	106	0.00
92	Benzo(g,h,i)perylene	40.000	38.521	3.7	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042821\
Data File : BF123752.D
Acq On : 28 Apr 2021 9:19
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 28 11:18:57 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 16:30:39 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0