

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042822\
 Data File : BF127979.D
 Acq On : 28 Apr 2022 14:23
 Operator : CG\JU
 Sample : N2482-04
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COP-1R

Quant Time: Apr 28 17:24:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 05:13:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.928	152	102463	20.000	ng	0.00
21) Naphthalene-d8	8.210	136	396198	20.000	ng	0.00
39) Acenaphthene-d10	9.969	164	223508	20.000	ng	0.00
64) Phenanthrene-d10	11.463	188	393620	20.000	ng	0.00
76) Chrysene-d12	14.110	240	219856	20.000	ng	0.00
86) Perylene-d12	15.615	264	226080	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.557	112	672987	104.616	ng	0.00
7) Phenol-d6	6.551	99	681127	81.600	ng	-0.01
23) Nitrobenzene-d5	7.498	82	1005082	119.292	ng	0.00
42) 2,4,6-Tribromophenol	10.769	330	381435	169.528	ng	0.00
45) 2-Fluorobiphenyl	9.292	172	1560492	118.114	ng	0.00
79) Terphenyl-d14	13.057	244	1573638	130.881	ng	0.00
Target Compounds						
35) Caprolactam	8.622	113	5593	2.878	ng	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042822\
Data File : BF127979.D
Acq On : 28 Apr 2022 14:23
Operator : CG\JU
Sample : N2482-04
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COP-1R

Quant Time: Apr 28 17:24:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 28 05:13:03 2022
Response via : Initial Calibration

