

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042822\
 Data File : BF127982.D
 Acq On : 28 Apr 2022 15:48
 Operator : CG\JU
 Sample : N2577-03
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GT-06-(8.5-10)

Quant Time: Apr 28 17:25:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 05:13:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.934	152	81330	20.000	ng	0.00
21) Naphthalene-d8	8.222	136	282525	20.000	ng	# 0.00
39) Acenaphthene-d10	9.975	164	175739	20.000	ng	0.00
64) Phenanthrene-d10	11.463	188	339486	20.000	ng	0.00
76) Chrysene-d12	14.110	240	191046	20.000	ng	0.00
86) Perylene-d12	15.616	264	189555	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.551	112	600406	117.585	ng	0.00
7) Phenol-d6	6.551	99	785020	118.483	ng	-0.01
23) Nitrobenzene-d5	7.498	82	552086	91.890	ng	0.00
42) 2,4,6-Tribromophenol	10.769	330	209060	118.172	ng	0.00
45) 2-Fluorobiphenyl	9.292	172	817843	78.729	ng	0.00
79) Terphenyl-d14	13.051	244	953703	91.282	ng	0.00
Target Compounds						
27) 2,4-Dimethylphenol	7.875	122	15691	3.829	ng	# 47
37) 2-Methylnaphthalene	8.934	142	382465	40.218	ng	98
38) 1-Methylnaphthalene	9.034	142	352760	38.163	ng	99
58) Fluorene	10.522	166	32244	2.939	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042822\
Data File : BF127982.D
Acq On : 28 Apr 2022 15:48
Operator : CG\JU
Sample : N2577-03
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GT-06-(8.5-10)

Quant Time: Apr 28 17:25:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 28 05:13:03 2022
Response via : Initial Calibration

