

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042915\
 Data File : BF078810.D
 Acq On : 29 Apr 2015 13:43
 Operator : TP/IZ
 Sample : G2009-02DL 10X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DCW-WW-119-42315DL

Quant Time: Apr 29 14:36:05 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 29 11:41:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.38	152	74729	20.00	ng	-0.01
21) Naphthalene-d8	8.96	136	317061	20.00	ng	-0.01
38) Acenaphthene-d10	11.14	164	142626	20.00	ng	0.00
63) Phenanthrene-d10	12.98	188	286185	20.00	ng	-0.01
75) Chrysene-d12	16.30	240	284800	20.00	ng	-0.01
86) Perylene-d12	18.11	264	279145	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.67	112	18780	4.62	ng	-0.01
7) Phenol-d6	6.92	99	14685	2.70	ng	-0.01
23) Nitrobenzene-d5	8.07	82	38326	8.70	ng	-0.01
41) 2,4,6-Tribromophenol	12.11	330	12634	13.81	ng	-0.01
44) 2-Fluorobiphenyl	10.31	172	95507	9.17	ng	-0.01
78) Terphenyl-d14	14.98	244	92506	7.74	ng	-0.01
Target Compounds						
33) 4-Chloroaniline	9.03	127	192285	30.36	ng	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042915\
Data File : BF078810.D
Acq On : 29 Apr 2015 13:43
Operator : TP/IZ
Sample : G2009-02DL 10X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DCW-WW-119-42315DL

Quant Time: Apr 29 14:36:05 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 29 11:41:09 2015
Response via : Initial Calibration

