

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042916\
 Data File : BF086497.D
 Acq On : 29 Apr 2016 17:35
 Operator : UM/SJ
 Sample : H2784-02
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP2(2A-2B-2C-2D)

Manual Integrations
 APPROVED

sohil
 4/30/2016 11:23:12 AM

Quant Time: Apr 30 01:21:42 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 27 14:20:52 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	173676	20.00	ng	-0.03
21) Naphthalene-d8	8.06	136	717714	20.00	ng	-0.03
38) Acenaphthene-d10	9.81	164	299910	20.00	ng	-0.03
63) Phenanthrene-d10	11.30	188	468464	20.00	ng	-0.02
75) Chrysene-d12	13.93	240	334650	20.00	ng	-0.03
86) Perylene-d12	15.33	264	278007	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	1322253	131.32	ng	-0.01
7) Phenol-d6	6.43	99	2043276	145.40	ng	-0.01
23) Nitrobenzene-d5	7.34	82	1176552	88.36	ng	-0.03
41) 2,4,6-Tribromophenol	10.60	330	347209	109.78	ng	-0.03
44) 2-Fluorobiphenyl	9.15	172	1717606	105.25	ng	-0.02
78) Terphenyl-d14	12.89	244	1130477	74.42	ng	-0.02
Target Compounds						
49) Dimethylphthalate	9.54	163	179679	7.96	ng	99
70) Phenanthrene	11.32	178	72480	2.95	ng	98
74) Fluoranthene	12.50	202	160618	6.47	ng	99
77) Pyrene	12.73	202	160737	6.67	ng	99
80) Benzo(a)anthracene	13.92	228	71424	4.00	ng	92
82) Chrysene	13.95	228	92151m	4.75	ng	
83) Bis(2-ethylhexyl)phthalate	13.92	149	28650	2.21	ng	# 100
85) Indeno(1,2,3-cd)pyrene	16.68	276	77684	5.41	ng	99
87) Benzo(b)fluoranthene	14.93	252	102614m	6.27	ng	
88) Benzo(k)fluoranthene	14.95	252	57264m	3.36	ng	
89) Benzo(a)pyrene	15.28	252	83262	5.39	ng	# 85
90) Dibenzo(a,h)anthracene	16.69	278	33091	2.62	ng	# 72
91) Benzo(a,h,i)perylene	17.09	276	138992	10.98	ng	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF042916\
Data File : BF086497.D
Acq On : 29 Apr 2016 17:35
Operator : UM/SJ
Sample : H2784-02
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
COMP2(2A-2B-2C-2D)

Manual Integrations
APPROVED

sohil
4/30/2016 11:23:12 AM

Quant Time: Apr 30 01:21:42 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF042716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 27 14:20:52 2016
Response via : Initial Calibration

