

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF042919\
 Data File : BF114026.D
 Acq On : 29 Apr 2019 16:51
 Operator : JU/SJ
 Sample : K2554-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RT3860

Manual Integrations
 APPROVED

Sohil
 4/30/2019 6:42:17 PM

Quant Time: Apr 30 05:53:46 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 23 03:32:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	128642	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	489316	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	259290	20.00	ng	-0.01
64) Phenanthrene-d10	11.41	188	517497	20.00	ng	0.00
76) Chrysene-d12	14.04	240	355128	20.00	ng	-0.02
87) Perylene-d12	15.54	264	382042	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	804113	106.91	ng	0.00
7) Phenol-d6	6.52	99	1057149	112.03	ng	0.00
23) Nitrobenzene-d5	7.45	82	629388	79.42	ng	0.00
42) 2,4,6-Tribromophenol	10.72	330	351702	111.34	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	1240531	74.83	ng	-0.01
79) Terphenyl-d14	12.99	244	1381924	79.78	ng	-0.01
Target Compounds						Qvalue
50) Dimethylphthalate	9.63	163	196328	10.470	ng	99
58) Fluorene	10.47	166	82526	4.910	ng	100
71) Phenanthrene	11.43	178	674248	25.925	ng	99
72) Anthracene	11.48	178	143366	5.459	ng	98
75) Fluoranthene	12.62	202	1296019	45.676	ng	99
78) Pyrene	12.85	202	951988	38.333	ng	99
81) Benzo(a)anthracene	14.04	228	249485	11.924	ng	98
83) Chrysene	14.07	228	204527	10.037	ng	97
86) Indeno(1,2,3-cd)pyrene	17.02	276	54761	2.837	ng	96
88) Benzo(b)fluoranthene	15.10	252	168412m	6.967	ng	
89) Benzo(k)fluoranthene	15.13	252	58675m	2.822	ng	
90) Benzo(a)pyrene	15.47	252	94895	4.539	ng	99
92) Benzo(g,h,i)perylene	17.46	276	46807	2.363	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF042919\
Data File : BF114026.D
Acq On : 29 Apr 2019 16:51
Operator : JU/SJ
Sample : K2554-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RT3860

Manual Integrations
APPROVED

Sohil
4/30/2019 6:42:17 PM

Quant Time: Apr 30 05:53:46 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 23 03:32:35 2019
Response via : Initial Calibration

