

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF042919\
 Data File : BF114035.D
 Acq On : 29 Apr 2019 20:46
 Operator : JU/SJ
 Sample : K2578-01 5X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WATER-COMP

Quant Time: Apr 30 07:33:04 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 23 03:32:35 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	116134	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	437034	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	205596	20.00	ng	0.00
64) Phenanthrene-d10	11.41	188	340086	20.00	ng	0.00
76) Chrysene-d12	14.06	240	258945	20.00	ng	0.00
87) Perylene-d12	15.54	264	278033	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	45856	6.75	ng	0.00
7) Phenol-d6	6.52	99	40678	4.78	ng	0.00
23) Nitrobenzene-d5	7.45	82	107141	15.14	ng	-0.01
42) 2,4,6-Tribromophenol	10.71	330	32638	13.03	ng	-0.01
45) 2-Fluorobiphenyl	9.25	172	244314	18.59	ng	-0.01
79) Terphenyl-d14	13.00	244	221221	17.51	ng	0.00
Target Compounds						
22) Acetophenone	7.30	105	265824	27.333	ng	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF042919\
Data File : BF114035.D
Acq On : 29 Apr 2019 20:46
Operator : JU/SJ
Sample : K2578-01 5X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WATER-COMP

Quant Time: Apr 30 07:33:04 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 23 03:32:35 2019
Response via : Initial Calibration

