

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042921\
 Data File : BF123763.D
 Acq On : 29 Apr 2021 14:40
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 29 15:08:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 29 11:58:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	0.00
2	1,4-Dioxane	40.000	39.402	1.5	89	-0.02
3	Pyridine	40.000	39.678	0.8	89	-0.02
4	n-Nitrosodimethylamine	40.000	40.418	-1.0	89	-0.02
5 S	2-Fluorophenol	80.000	78.520	1.9	89	0.00
6	Aniline	40.000	38.857	2.9	87	0.00
7 S	Phenol-d6	80.000	77.450	3.2	89	0.00
8	2-Chlorophenol	40.000	39.020	2.4	88	0.00
9	Benzaldehyde	40.000	38.903	2.7	92	0.00
10 C	Phenol	40.000	38.530	3.7	88	0.00
11	bis(2-Chloroethyl)ether	40.000	38.443	3.9	87	0.00
12	1,3-Dichlorobenzene	40.000	38.824	2.9	88	0.00
13 C	1,4-Dichlorobenzene	40.000	38.751	3.1	88	0.00
14	1,2-Dichlorobenzene	40.000	37.515	6.2	90	0.00
15	Benzyl Alcohol	40.000	39.610	1.0	89	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.852	0.4	90	0.00
17	2-Methylphenol	40.000	39.547	1.1	90	0.00
18	Hexachloroethane	40.000	40.521	-1.3	90	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.521	1.2	90	0.00
20	3+4-Methylphenols	40.000	36.678	8.3	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	88	0.00
22	Acetophenone	40.000	39.277	1.8	90	0.00
23 S	Nitrobenzene-d5	80.000	79.490	0.6	88	0.00
24	Nitrobenzene	40.000	40.038	-0.1	89	0.00
25	Isophorone	40.000	39.923	0.2	89	0.00
26 C	2-Nitrophenol	40.000	41.239	-3.1	89	0.00
27	2,4-Dimethylphenol	40.000	40.119	-0.3	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.158	-0.4	89	0.00
29 C	2,4-Dichlorophenol	40.000	39.912	0.2	89	0.00
30	1,2,4-Trichlorobenzene	40.000	39.913	0.2	90	0.00
31	Naphthalene	40.000	39.350	1.6	89	0.00
32	Benzoic acid	40.000	35.770	10.6	79	0.00
33	4-Chloroaniline	40.000	40.420	-1.1	90	0.00
34 C	Hexachlorobutadiene	40.000	39.754	0.6	87	0.00
35	Caprolactam	40.000	38.381	4.0	84	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.789	0.5	87	0.00
37	2-Methylnaphthalene	40.000	39.522	1.2	89	0.00
38	1-Methylnaphthalene	40.000	39.501	1.2	90	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.838	0.4	89	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.990	-7.5	90	0.00
42 S	2,4,6-Tribromophenol	80.000	81.288	-1.6	90	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.097	-0.2	87	0.00
44	2,4,5-Trichlorophenol	40.000	39.798	0.5	87	0.00
45 S	2-Fluorobiphenyl	80.000	78.276	2.2	90	0.00
46	1,1'-Biphenyl	40.000	38.718	3.2	88	0.00
47	2-Chloronaphthalene	40.000	39.034	2.4	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042921\
 Data File : BF123763.D
 Acq On : 29 Apr 2021 14:40
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 29 15:08:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 29 11:58:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.506	1.2	87	0.00
49	Acenaphthylene	40.000	39.795	0.5	89	0.00
50	Dimethylphthalate	40.000	39.296	1.8	89	0.00
51	2,6-Dinitrotoluene	40.000	41.295	-3.2	91	0.00
52 C	Acenaphthene	40.000	39.605	1.0	89	0.00
53	3-Nitroaniline	40.000	38.968	2.6	86	0.00
54 P	2,4-Dinitrophenol	40.000	37.749	5.6	78	0.00
55	Dibenzofuran	40.000	39.376	1.6	88	0.00
56 P	4-Nitrophenol	40.000	39.594	1.0	84	0.00
57	2,4-Dinitrotoluene	40.000	39.380	1.5	87	0.00
58	Fluorene	40.000	39.257	1.9	89	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.325	-0.8	87	0.00
60	Diethylphthalate	40.000	39.776	0.6	89	0.00
61	4-Chlorophenyl-phenylether	40.000	39.544	1.1	90	0.00
62	4-Nitroaniline	40.000	38.261	4.3	84	0.00
63	Azobenzene	40.000	39.449	1.4	88	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.769	-4.4	85	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.044	-0.1	88	0.00
67	4-Bromophenyl-phenylether	40.000	41.026	-2.6	89	0.00
68	Hexachlorobenzene	40.000	40.724	-1.8	91	0.00
69	Atrazine	40.000	39.152	2.1	86	0.00
70 C	Pentachlorophenol	40.000	39.254	1.9	84	0.00
71	Phenanthrene	40.000	39.149	2.1	87	0.00
72	Anthracene	40.000	39.717	0.7	89	0.00
73	Carbazole	40.000	39.064	2.3	87	0.00
74	Di-n-butylphthalate	40.000	40.518	-1.3	88	0.00
75 C	Fluoranthene	40.000	38.487	3.8	85	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzydine	40.000	43.420	-8.6	83	0.00
78	Pyrene	40.000	41.471	-3.7	88	0.00
79 S	Terphenyl-d14	80.000	80.796	-1.0	86	0.00
80	Butylbenzylphthalate	40.000	40.720	-1.8	84	0.00
81	Benzo(a)anthracene	40.000	38.576	3.6	82	0.00
82	3,3'-Dichlorobenzidine	40.000	38.717	3.2	81	0.00
83	Chrysene	40.000	38.907	2.7	84	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.438	1.4	83	0.00
85 c	Di-n-octyl phthalate	40.000	38.424	3.9	79	0.00
86 I	Perylene-d12	20.000	20.000	0.0	80	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.610	1.0	86	0.00
88	Benzo(b)fluoranthene	40.000	39.664	0.8	82	0.00
89	Benzo(k)fluoranthene	40.000	39.535	1.2	78	0.00
90 C	Benzo(a)pyrene	40.000	39.617	1.0	80	0.00
91	Dibenzo(a,h)anthracene	40.000	39.560	1.1	86	0.00
92	Benzo(g,h,i)perylene	40.000	39.099	2.3	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042921\
Data File : BF123763.D
Acq On : 29 Apr 2021 14:40
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 29 15:08:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 29 11:58:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0