

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042922\
 Data File : BF127991.D
 Acq On : 29 Apr 2022 13:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 29 17:55:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 05:13:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	87	0.00
2	1,4-Dioxane	40.000	39.406	1.5	83	0.00
3	Pyridine	40.000	41.820	-4.6	86	0.00
4	n-Nitrosodimethylamine	40.000	42.391	-6.0	89	0.00
5 S	2-Fluorophenol	80.000	79.075	1.2	86	0.00
6	Aniline	40.000	42.426	-6.1	88	0.00
7 S	Phenol-d6	80.000	82.404	-3.0	88	-0.01
8	2-Chlorophenol	40.000	40.845	-2.1	86	-0.01
9	Benzaldehyde	40.000	34.365	14.1	84	0.00
10 C	Phenol	40.000	40.609	-1.5	86	0.00
11	bis(2-Chloroethyl)ether	40.000	40.997	-2.5	87	0.00
12	1,3-Dichlorobenzene	40.000	39.799	0.5	86	0.00
13 C	1,4-Dichlorobenzene	40.000	39.990	0.0	87	0.00
14	1,2-Dichlorobenzene	40.000	39.292	1.8	86	0.00
15	Benzyl Alcohol	40.000	44.902	-12.3	90	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.939	-2.3	85	0.00
17	2-Methylphenol	40.000	42.776	-6.9	88	0.00
18	Hexachloroethane	40.000	41.509	-3.8	89	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.107	-5.3	90	0.00
20	3+4-Methylphenols	40.000	41.442	-3.6	87	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	84	0.00
22	Acetophenone	40.000	41.724	-4.3	88	0.00
23 S	Nitrobenzene-d5	80.000	87.507	-9.4	90	0.00
24	Nitrobenzene	40.000	43.312	-8.3	89	0.00
25	Isophorone	40.000	43.258	-8.1	89	0.00
26 C	2-Nitrophenol	40.000	44.936	-12.3	89	0.00
27	2,4-Dimethylphenol	40.000	42.965	-7.4	86	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.854	-4.6	87	0.00
29 C	2,4-Dichlorophenol	40.000	42.446	-6.1	88	0.00
30	1,2,4-Trichlorobenzene	40.000	41.198	-3.0	86	0.00
31	Naphthalene	40.000	40.701	-1.8	85	0.00
32	Benzoic acid	40.000	45.248	-13.1	85	0.00
33	4-Chloroaniline	40.000	41.836	-4.6	85	0.00
34 C	Hexachlorobutadiene	40.000	43.547	-8.9	91	0.00
35	Caprolactam	40.000	42.575	-6.4	84	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.632	-9.1	89	0.00
37	2-Methylnaphthalene	40.000	41.706	-4.3	87	-0.01
38	1-Methylnaphthalene	40.000	41.112	-2.8	86	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	86	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.891	-2.2	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.270	-8.2	87	-0.01
42 S	2,4,6-Tribromophenol	80.000	87.902	-9.9	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.157	-5.4	89	0.00
44	2,4,5-Trichlorophenol	40.000	41.235	-3.1	85	0.00
45 S	2-Fluorobiphenyl	80.000	86.092	-7.6	91	0.00
46	1,1'-Biphenyl	40.000	40.526	-1.3	87	0.00
47	2-Chloronaphthalene	40.000	39.908	0.2	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042922\
 Data File : BF127991.D
 Acq On : 29 Apr 2022 13:11
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 29 17:55:20 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 05:13:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.211	-10.5	91	0.00
49	Acenaphthylene	40.000	40.224	-0.6	85	0.00
50	Dimethylphthalate	40.000	41.011	-2.5	87	0.00
51	2,6-Dinitrotoluene	40.000	42.134	-5.3	86	0.00
52 C	Acenaphthene	40.000	41.258	-3.1	87	0.00
53	3-Nitroaniline	40.000	40.544	-1.4	83	0.00
54 P	2,4-Dinitrophenol	40.000	45.004	-12.5	88	0.00
55	Dibenzofuran	40.000	40.518	-1.3	87	0.00
56 P	4-Nitrophenol	40.000	39.116	2.2	78	0.00
57	2,4-Dinitrotoluene	40.000	42.697	-6.7	86	0.00
58	Fluorene	40.000	40.294	-0.7	87	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.159	-5.4	87	0.00
60	Diethylphthalate	40.000	41.199	-3.0	87	0.00
61	4-Chlorophenyl-phenylether	40.000	40.851	-2.1	89	0.00
62	4-Nitroaniline	40.000	41.046	-2.6	83	0.00
63	Azobenzene	40.000	41.178	-2.9	87	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	85	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.248	-18.1	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.005	-0.0	86	0.00
67	4-Bromophenyl-phenylether	40.000	41.095	-2.7	88	0.00
68	Hexachlorobenzene	40.000	41.547	-3.9	89	0.00
69	Atrazine	40.000	39.087	2.3	84	0.00
70 C	Pentachlorophenol	40.000	42.066	-5.2	83	0.00
71	Phenanthrene	40.000	39.324	1.7	85	0.00
72	Anthracene	40.000	39.768	0.6	85	0.00
73	Carbazole	40.000	38.649	3.4	83	0.00
74	Di-n-butylphthalate	40.000	40.628	-1.6	86	0.00
75 C	Fluoranthene	40.000	39.375	1.6	85	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	83	0.00
77	Benzydine	40.000	50.236	-25.6#	97	0.00
78	Pyrene	40.000	40.769	-1.9	84	0.00
79 S	Terphenyl-d14	80.000	83.420	-4.3	90	0.00
80	Butylbenzylphthalate	40.000	44.346	-10.9	89	0.00
81	Benzo(a)anthracene	40.000	41.719	-4.3	85	0.00
82	3,3'-Dichlorobenzidine	40.000	39.932	0.2	83	0.00
83	Chrysene	40.000	40.570	-1.4	85	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.763	-9.4	88	0.00
85 c	Di-n-octyl phthalate	40.000	43.984	-10.0	87	0.00
86 I	Perylene-d12	20.000	20.000	0.0	81	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.065	2.3	75	-0.01
88	Benzo(b)fluoranthene	40.000	44.082	-10.2	89	0.00
89	Benzo(k)fluoranthene	40.000	40.647	-1.6	79	0.00
90 C	Benzo(a)pyrene	40.000	42.108	-5.3	83	0.00
91	Dibenzo(a,h)anthracene	40.000	39.671	0.8	75	-0.01
92	Benzo(g,h,i)perylene	40.000	37.764	5.6	71	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042922\
Data File : BF127991.D
Acq On : 29 Apr 2022 13:11
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 29 17:55:20 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 28 05:13:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0