

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042922\
 Data File : BF127999.D
 Acq On : 29 Apr 2022 17:16
 Operator : CG\JU
 Sample : N2602-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 10-16-SB3-BK-PRISON

Quant Time: Apr 29 17:58:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Apr 28 05:13:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.922	152	112370	20.000	ng	-0.01
21) Naphthalene-d8	8.204	136	436896	20.000	ng	-0.01
39) Acenaphthene-d10	9.963	164	251668	20.000	ng	-0.01
64) Phenanthrene-d10	11.457	188	445123	20.000	ng	-0.01
76) Chrysene-d12	14.098	240	241694	20.000	ng	-0.02
86) Perylene-d12	15.610	264	237121	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.546	112	635100	90.022	ng	0.00
7) Phenol-d6	6.546	99	834445	91.154	ng	-0.02
23) Nitrobenzene-d5	7.487	82	597115	64.269	ng	-0.02
42) 2,4,6-Tribromophenol	10.757	330	234748	92.659	ng	-0.01
45) 2-Fluorobiphenyl	9.281	172	1005995	67.624	ng	-0.01
79) Terphenyl-d14	13.045	244	989783	74.883	ng	-0.01


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042922\
Data File : BF127999.D
Acq On : 29 Apr 2022 17:16
Operator : CG\JU
Sample : N2602-05
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
10-16-SB3-BK-PRISON

Quant Time: Apr 29 17:58:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Apr 28 05:13:03 2022
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.922 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF127999.D
Acq: 29 Apr 2022 17:16




Instrument :
BNA_F
ClientSampleId :
10-16-SB3-BK-PRISON

Tgt Ion:152 Resp: 112370
Ion Ratio Lower Upper
152 100
150 158.7 125.3 187.9
115 65.6 48.8 73.2



#5
2-Fluorophenol
Concen: 90.022 ng
RT: 5.546 min Scan# 554
Delta R.T. -0.006 min
Lab File: BF127999.D
Acq: 29 Apr 2022 17:16

Tgt Ion:112 Resp: 635100
Ion Ratio Lower Upper
112 100
64 66.9 49.5 74.3
63 36.6 26.3 39.5





#7

Phenol-d6

Concen: 91.154 ng

RT: 6.546 min Scan# 71

Delta R.T. -0.017 min

Lab File: BF127999.D

Acq: 29 Apr 2022 17:16

Instrument :

BNA_F

ClientSampleId :

10-16-SB3-BK-PRISON





Tgt Ion: 99 Resp: 834445

Ion Ratio Lower Upper

99 100

42 22.5 16.6 24.8

71 39.3 30.0 45.0



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.204 min Scan# 1006

Delta R.T. -0.012 min

Lab File: BF127999.D

Acq: 29 Apr 2022 17:16

Tgt Ion: 136 Resp: 436896




Ion Ratio Lower Upper

136 100

137 10.9 8.6 13.0

54 9.0 6.8 10.2

68 6.2 5.1 7.7





#23

Nitrobenzene-d5

Concen: 64.269 ng

RT: 7.487 min Scan# 81

Delta R.T. -0.017 min

Lab File: BF127999.D

Acq: 29 Apr 2022 17:16

Instrument :

BNA_F

ClientSampleId :

10-16-SB3-BK-PRISON





Tgt Ion: 82 Resp: 597115

Ion Ratio Lower Upper

82 100

128 37.4 30.0 45.0

54 51.6 41.3 61.9



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.963 min Scan# 1305

Delta R.T. -0.012 min

Lab File: BF127999.D

Acq: 29 Apr 2022 17:16




Tgt Ion: 164 Resp: 251668


Ion Ratio Lower Upper

164 100

162 95.3 76.0 114.0


160 42.3 34.8 52.2








#42
2,4,6-Tribromophenol
Concen: 92.659 ng
RT: 10.757 min Scan# 1440
Delta R.T. -0.012 min
Lab File: BF127999.D
Acq: 29 Apr 2022 17:16

Instrument :
BNA_F
ClientSampleId :
10-16-SB3-BK-PRISON







Tgt Ion:330 Resp: 234748
Ion Ratio Lower Upper
330 100
332 95.8 79.0 118.4
141 34.5 28.9 43.3



#45
2-Fluorobiphenyl
Concen: 67.624 ng
RT: 9.281 min Scan# 1189
Delta R.T. -0.012 min
Lab File: BF127999.D
Acq: 29 Apr 2022 17:16

Tgt Ion:172 Resp: 1005995
Ion Ratio Lower Upper
172 100
171 35.2 28.6 42.8
170 23.4 19.1 28.7









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.457 min Scan# 11
Delta R.T. -0.012 min
Lab File: BF127999.D
Acq: 29 Apr 2022 17:16




Instrument :
BNA_F
ClientSampleId :
10-16-SB3-BK-PRISON


Tgt Ion:188 Resp: 445123
Ion Ratio Lower Upper
188 100
94 8.4 8.0 12.0
80 9.1 8.9 13.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.098 min Scan# 2008
Delta R.T. -0.017 min
Lab File: BF127999.D
Acq: 29 Apr 2022 17:16


Tgt Ion:240 Resp: 241694
Ion Ratio Lower Upper
240 100
120 10.3 7.0 10.6
236 25.6 20.3 30.5








#79
 Terphenyl-d14
 Concen: 74.883 ng
 RT: 13.045 min Scan# 1829
 Delta R.T. -0.012 min
 Lab File: BF127999.D
 Acq: 29 Apr 2022 17:16

Instrument :
 BNA_F
 ClientSampleId :
 10-16-SB3-BK-PRISON



Tgt Ion:244 Resp: 989783
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.2 9.4
 122 9.5 8.6 12.8



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.610 min Scan# 2265
 Delta R.T. -0.012 min
 Lab File: BF127999.D
 Acq: 29 Apr 2022 17:16

Tgt Ion:264 Resp: 237121
 Ion Ratio Lower Upper
 264 100
 260 23.9 20.1 30.1
 265 22.1 17.5 26.3

