




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042924\
Data File : BF137745.D
Acq On : 29 Apr 2024 09:35
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Apr 29 11:50:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration



TIC: BF137745.D\data.ms

(70) Pentachlorophenol (C)




11.398min (+ 0.000) 334435.42 ng

response 505508

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.00	62.52
264.00	64.20	62.74
0.00	0.00	0.00

```
Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042924\  
Data File : BF137745.D  
Acq On    : 29 Apr 2024  09:35  
Operator  : CG\JU  
Sample    : DFTPP  
Misc      :  
ALS Vial  : 1    Sample Multiplier: 1
```

Quant Time: Apr 29 11:50:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration



TIC: BF137745.D\data.ms

(77) Benzidine

12.945min (+ 0.006) 436241.71 ng m

response **2109158**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.30	14.62
183.00	11.70	12.18
0.00	0.00	0.00