

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042924\
 Data File : BF137746.D
 Acq On : 29 Apr 2024 10:05
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 29 19:28:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 03:57:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	87	0.00
2	1,4-Dioxane	40.000	37.724	5.7	85	-0.01
3	Pyridine	40.000	37.425	6.4	84	0.00
4	n-Nitrosodimethylamine	40.000	34.595	13.5	78	0.00
5 S	2-Fluorophenol	80.000	74.595	6.8	85	0.00
6	Aniline	40.000	36.126	9.7	81	0.00
7 S	Phenol-d6	80.000	73.594	8.0	83	0.00
8	2-Chlorophenol	40.000	38.168	4.6	86	0.00
9	Benzaldehyde	40.000	34.352	14.1	84	0.00
10 C	Phenol	40.000	36.999	7.5	84	0.00
11	bis(2-Chloroethyl)ether	40.000	36.042	9.9	81	0.00
12	1,3-Dichlorobenzene	40.000	37.966	5.1	86	0.00
13 C	1,4-Dichlorobenzene	40.000	38.231	4.4	87	0.00
14	1,2-Dichlorobenzene	40.000	38.319	4.2	87	0.00
15	Benzyl Alcohol	40.000	36.398	9.0	81	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.302	14.2	77	0.00
17	2-Methylphenol	40.000	37.097	7.3	84	0.00
18	Hexachloroethane	40.000	37.543	6.1	85	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	34.141	14.6	78	0.00
20	3+4-Methylphenols	40.000	37.010	7.5	82	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	83	0.00
22	Acetophenone	40.000	37.450	6.4	82	0.00
23 S	Nitrobenzene-d5	80.000	75.813	5.2	82	0.00
24	Nitrobenzene	40.000	37.837	5.4	82	0.00
25	Isophorone	40.000	35.776	10.6	79	0.00
26 C	2-Nitrophenol	40.000	42.538	-6.3	89	0.00
27	2,4-Dimethylphenol	40.000	37.488	6.3	84	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.473	8.8	80	0.00
29 C	2,4-Dichlorophenol	40.000	39.273	1.8	86	0.00
30	1,2,4-Trichlorobenzene	40.000	39.000	2.5	85	0.00
31	Naphthalene	40.000	38.365	4.1	84	0.00
32	Benzoic acid	40.000	40.466	-1.2	85	0.00
33	4-Chloroaniline	40.000	38.074	4.8	83	0.00
34 C	Hexachlorobutadiene	40.000	39.664	0.8	86	0.00
35	Caprolactam	40.000	37.096	7.3	80	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.660	5.9	82	0.00
37	2-Methylnaphthalene	40.000	37.870	5.3	84	0.00
38	1-Methylnaphthalene	40.000	37.506	6.2	83	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	82	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.355	1.6	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.233	-0.6	84	0.00
42 S	2,4,6-Tribromophenol	80.000	79.492	0.6	88	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.436	3.9	82	0.00
44	2,4,5-Trichlorophenol	40.000	38.557	3.6	84	0.00
45 S	2-Fluorobiphenyl	80.000	73.549	8.1	85	0.00
46	1,1'-Biphenyl	40.000	38.468	3.8	84	0.00
47	2-Chloronaphthalene	40.000	38.301	4.2	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042924\
 Data File : BF137746.D
 Acq On : 29 Apr 2024 10:05
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 29 19:28:14 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Apr 24 03:57:12 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.853	2.9	81	0.00
49	Acenaphthylene	40.000	37.900	5.3	82	0.00
50	Dimethylphthalate	40.000	36.648	8.4	81	0.00
51	2,6-Dinitrotoluene	40.000	39.524	1.2	84	0.00
52 C	Acenaphthene	40.000	37.870	5.3	83	0.00
53	3-Nitroaniline	40.000	38.640	3.4	81	0.00
54 P	2,4-Dinitrophenol	40.000	39.713	0.7	91	0.00
55	Dibenzofuran	40.000	37.706	5.7	83	0.00
56 P	4-Nitrophenol	40.000	37.220	7.0	77	0.00
57	2,4-Dinitrotoluene	40.000	40.576	-1.4	85	0.00
58	Fluorene	40.000	37.239	6.9	83	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.033	4.9	83	0.00
60	Diethylphthalate	40.000	36.147	9.6	80	0.00
61	4-Chlorophenyl-phenylether	40.000	38.199	4.5	85	0.00
62	4-Nitroaniline	40.000	38.286	4.3	80	0.00
63	Azobenzene	40.000	35.306	11.7	77	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	81	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.133	-5.3	87	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.590	3.5	83	0.00
67	4-Bromophenyl-phenylether	40.000	39.464	1.3	85	0.00
68	Hexachlorobenzene	40.000	40.102	-0.3	86	0.00
69	Atrazine	40.000	36.432	8.9	77	0.00
70 C	Pentachlorophenol	40.000	38.488	3.8	80	0.00
71	Phenanthrene	40.000	38.120	4.7	82	0.00
72	Anthracene	40.000	38.075	4.8	82	0.00
73	Carbazole	40.000	37.481	6.3	80	0.00
74	Di-n-butylphthalate	40.000	35.930	10.2	78	0.00
75 C	Fluoranthene	40.000	36.135	9.7	78	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	71	0.00
77	Benzydine	40.000	44.201	-10.5	83	0.00
78	Pyrene	40.000	41.118	-2.8	78	0.00
79 S	Terphenyl-d14	80.000	81.640	-2.1	79	0.00
80	Butylbenzylphthalate	40.000	40.148	-0.4	73	0.00
81	Benzo(a)anthracene	40.000	38.135	4.7	71	0.00
82	3,3'-Dichlorobenzidine	40.000	38.490	3.8	71	0.00
83	Chrysene	40.000	37.706	5.7	70	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	36.649	8.4	67	0.00
85 c	Di-n-octyl phthalate	40.000	35.685	10.8	63	0.00
86 I	Perylene-d12	20.000	20.000	0.0	77	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	46.967	-17.4	99	0.00
88	Benzo(b)fluoranthene	40.000	35.928	10.2	70	0.00
89	Benzo(k)fluoranthene	40.000	39.025	2.4	78	0.00
90 C	Benzo(a)pyrene	40.000	38.831	2.9	77	0.00
91	Dibenzo(a,h)anthracene	40.000	47.221	-18.1	100	0.00
92	Benzo(g,h,i)perylene	40.000	48.586	-21.5	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF042924\
Data File : BF137746.D
Acq On : 29 Apr 2024 10:05
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 29 19:28:14 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Apr 24 03:57:12 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0