

Data Path : Z:\HPCHEM1\BNA F\DATA\BF043018\
 Data File : BF104939.D
 Acq On : 30 Apr 2018 18:07
 Operator : JU/SJ
 Sample : J2609-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-01D

Manual Integrations
 APPROVED

Sohil
 5/1/2018 7:05:41 PM

Quant Time: May 01 05:46:59 2018
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 30 15:20:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	122878	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	502170	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	220066	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	360375	20.00	ng	0.00
75) Chrysene-d12	13.97	240	228471	20.00	ng	0.00
86) Perylene-d12	15.44	264	214661	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.41	112	688961	85.73	ng	0.00
7) Phenol-d6	6.44	99	844523	85.53	ng	0.00
23) Nitrobenzene-d5	7.36	82	517539	67.30	ng	-0.01
41) 2,4,6-Tribromophenol	10.63	330	186677	81.78	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	776204	55.72	ng	0.00
78) Terphenyl-d14	12.91	244	540761	53.96	ng	0.00
Target Compounds						
10) Phenol	6.45	94	21323	2.035	ng	Qvalue # 78
49) Dimethylphthalate	9.55	163	101783	5.865	ng	99
70) Phenanthrene	11.35	178	49913	2.487	ng	98
74) Fluoranthene	12.54	202	61780	2.946	ng	97
77) Pyrene	12.77	202	47382	2.798	ng	96
87) Benzo(b)fluoranthene	15.02	252	33536m	2.474	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF043018\
Data File : BF104939.D
Acq On : 30 Apr 2018 18:07
Operator : JU/SJ
Sample : J2609-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-01D

Manual Integrations
APPROVED

Sohil
5/1/2018 7:05:41 PM

Quant Time: May 01 05:46:59 2018
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF040618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 30 15:20:08 2018
Response via : Initial Calibration

