


Data Path : Z:\HPCHEM1\BNA F\Data\BF043018\  
 Data File : BF104926.D  
 Acq On : 30 Apr 2018 12:14  
 Operator : JU/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF040618.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Mon Apr 30 15:20:08 2018



AutoFind: Scans 1611, 1612, 1613; Background Corrected with Scan 1606

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 29.9      | 21601   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.8       | 186     | PASS             |
| 69          | 198          | 0.00         | 100          | 32.4      | 23438   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 48.0      | 34711   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 317     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 72274   | PASS             |
| 199         | 198          | 5            | 9            | 6.5       | 4713    | PASS             |
| 275         | 198          | 10           | 60           | 27.5      | 19872   | PASS             |
| 365         | 198          | 1            | 100          | 3.3       | 2353    | PASS             |
| 441         | 443          | 0.01         | 100          | 82.9      | 10902   | PASS             |
| 442         | 198          | 50           | 100          | 98.4      | 71144   | PASS             |
| 443         | 442          | 15           | 24           | 18.5      | 13150   | PASS             |