

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043021\
 Data File : BF123765.D
 Acq On : 30 Apr 2021 8:43
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 30 11:16:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 11:10:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	38.842	2.9	93	0.00
3	Pyridine	40.000	40.829	-2.1	97	0.00
4	n-Nitrosodimethylamine	40.000	40.941	-2.4	96	0.00
5 S	2-Fluorophenol	80.000	76.886	3.9	93	0.00
6	Aniline	40.000	38.784	3.0	93	0.00
7 S	Phenol-d6	80.000	76.680	4.1	94	0.00
8	2-Chlorophenol	40.000	39.368	1.6	95	0.00
9	Benzaldehyde	40.000	38.987	2.5	98	0.00
10 C	Phenol	40.000	38.916	2.7	94	0.00
11	bis(2-Chloroethyl)ether	40.000	38.268	4.3	92	0.00
12	1,3-Dichlorobenzene	40.000	39.122	2.2	94	0.00
13 C	1,4-Dichlorobenzene	40.000	38.917	2.7	94	0.00
14	1,2-Dichlorobenzene	40.000	37.156	7.1	95	0.00
15	Benzyl Alcohol	40.000	38.953	2.6	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.108	2.2	94	0.00
17	2-Methylphenol	40.000	38.957	2.6	94	0.00
18	Hexachloroethane	40.000	39.088	2.3	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.544	3.6	94	0.00
20	3+4-Methylphenols	40.000	36.891	7.8	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	93	0.00
22	Acetophenone	40.000	39.322	1.7	96	0.00
23 S	Nitrobenzene-d5	80.000	79.513	0.6	94	0.00
24	Nitrobenzene	40.000	39.293	1.8	93	0.00
25	Isophorone	40.000	39.265	1.8	93	0.00
26 C	2-Nitrophenol	40.000	40.251	-0.6	92	0.00
27	2,4-Dimethylphenol	40.000	38.857	2.9	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.668	3.3	91	0.00
29 C	2,4-Dichlorophenol	40.000	39.109	2.2	93	0.00
30	1,2,4-Trichlorobenzene	40.000	38.768	3.1	93	0.00
31	Naphthalene	40.000	39.211	2.0	94	0.00
32	Benzoic acid	40.000	35.135	12.2	83	0.00
33	4-Chloroaniline	40.000	38.790	3.0	92	0.00
34 C	Hexachlorobutadiene	40.000	40.215	-0.5	94	0.00
35	Caprolactam	40.000	39.895	0.3	92	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.718	0.7	92	0.00
37	2-Methylnaphthalene	40.000	39.304	1.7	94	0.00
38	1-Methylnaphthalene	40.000	39.049	2.4	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.306	1.7	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.686	-6.7	94	0.00
42 S	2,4,6-Tribromophenol	80.000	78.334	2.1	92	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.174	-0.4	92	0.00
44	2,4,5-Trichlorophenol	40.000	39.352	1.6	91	0.00
45 S	2-Fluorobiphenyl	80.000	76.577	4.3	94	0.00
46	1,1'-Biphenyl	40.000	38.695	3.3	93	0.00
47	2-Chloronaphthalene	40.000	38.974	2.6	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043021\
 Data File : BF123765.D
 Acq On : 30 Apr 2021 8:43
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 30 11:16:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 11:10:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.422	1.4	93	0.00
49	Acenaphthylene	40.000	39.518	1.2	94	0.00
50	Dimethylphthalate	40.000	38.988	2.5	94	0.00
51	2,6-Dinitrotoluene	40.000	40.309	-0.8	95	0.00
52 C	Acenaphthene	40.000	38.713	3.2	93	0.00
53	3-Nitroaniline	40.000	39.005	2.5	91	0.00
54 P	2,4-Dinitrophenol	40.000	38.331	4.2	84	0.00
55	Dibenzofuran	40.000	38.959	2.6	93	0.00
56 P	4-Nitrophenol	40.000	39.445	1.4	89	0.00
57	2,4-Dinitrotoluene	40.000	39.214	2.0	92	0.00
58	Fluorene	40.000	38.889	2.8	94	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.537	1.2	91	0.00
60	Diethylphthalate	40.000	39.439	1.4	94	0.00
61	4-Chlorophenyl-phenylether	40.000	38.669	3.3	94	0.00
62	4-Nitroaniline	40.000	38.771	3.1	90	0.00
63	Azobenzene	40.000	38.963	2.6	92	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.981	-2.5	88	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.720	0.7	92	0.00
67	4-Bromophenyl-phenylether	40.000	40.720	-1.8	94	0.00
68	Hexachlorobenzene	40.000	40.716	-1.8	96	0.00
69	Atrazine	40.000	39.917	0.2	93	0.00
70 C	Pentachlorophenol	40.000	40.076	-0.2	91	0.00
71	Phenanthrene	40.000	39.651	0.9	93	0.00
72	Anthracene	40.000	39.712	0.7	94	0.00
73	Carbazole	40.000	39.447	1.4	93	0.00
74	Di-n-butylphthalate	40.000	39.674	0.8	91	0.00
75 C	Fluoranthene	40.000	39.300	1.8	92	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	94	0.00
77	Benzydine	40.000	34.841	12.9	76	0.00
78	Pyrene	40.000	38.822	2.9	94	0.00
79 S	Terphenyl-d14	80.000	77.179	3.5	94	0.00
80	Butylbenzylphthalate	40.000	39.071	2.3	92	0.00
81	Benzo(a)anthracene	40.000	39.077	2.3	95	0.00
82	3,3'-Dichlorobenzidine	40.000	39.204	2.0	93	0.00
83	Chrysene	40.000	38.879	2.8	95	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.887	2.8	93	0.00
85 c	Di-n-octyl phthalate	40.000	38.879	2.8	91	0.00
86 I	Perylene-d12	20.000	20.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.746	5.6	96	0.00
88	Benzo(b)fluoranthene	40.000	38.773	3.1	94	0.00
89	Benzo(k)fluoranthene	40.000	41.495	-3.7	96	0.00
90 C	Benzo(a)pyrene	40.000	39.514	1.2	94	0.00
91	Dibenzo(a,h)anthracene	40.000	38.055	4.9	97	0.00
92	Benzo(g,h,i)perylene	40.000	37.288	6.8	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043021\
Data File : BF123765.D
Acq On : 30 Apr 2021 8:43
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 30 11:16:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 30 11:10:41 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0