

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043021\
 Data File : BF123775.D
 Acq On : 30 Apr 2021 17:54
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 30 18:14:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 14:46:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	39.414	1.5	94	-0.01
3	Pyridine	40.000	39.770	0.6	94	-0.01
4	n-Nitrosodimethylamine	40.000	41.641	-4.1	97	-0.01
5 S	2-Fluorophenol	80.000	77.284	3.4	93	0.00
6	Aniline	40.000	39.815	0.5	94	0.00
7 S	Phenol-d6	80.000	78.044	2.4	95	0.00
8	2-Chlorophenol	40.000	39.607	1.0	94	0.00
9	Benzaldehyde	40.000	39.107	2.2	97	0.00
10 C	Phenol	40.000	39.092	2.3	94	0.00
11	bis(2-Chloroethyl)ether	40.000	39.005	2.5	93	0.00
12	1,3-Dichlorobenzene	40.000	39.353	1.6	94	0.00
13 C	1,4-Dichlorobenzene	40.000	38.819	3.0	93	0.00
14	1,2-Dichlorobenzene	40.000	37.547	6.1	95	0.00
15	Benzyl Alcohol	40.000	40.070	-0.2	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.515	-1.3	96	0.00
17	2-Methylphenol	40.000	39.862	0.3	95	0.00
18	Hexachloroethane	40.000	40.021	-0.1	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.672	0.8	95	0.00
20	3+4-Methylphenols	40.000	38.092	4.8	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	39.373	1.6	96	0.00
23 S	Nitrobenzene-d5	80.000	80.087	-0.1	95	0.00
24	Nitrobenzene	40.000	39.744	0.6	94	0.00
25	Isophorone	40.000	40.575	-1.4	96	0.00
26 C	2-Nitrophenol	40.000	41.346	-3.4	95	0.00
27	2,4-Dimethylphenol	40.000	39.475	1.3	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.655	0.9	93	0.00
29 C	2,4-Dichlorophenol	40.000	39.310	1.7	94	0.00
30	1,2,4-Trichlorobenzene	40.000	40.375	-0.9	97	0.00
31	Naphthalene	40.000	39.259	1.9	94	0.00
32	Benzoic acid	40.000	35.727	10.7	84	0.00
33	4-Chloroaniline	40.000	39.411	1.5	94	0.00
34 C	Hexachlorobutadiene	40.000	39.911	0.2	93	0.00
35	Caprolactam	40.000	40.322	-0.8	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.574	-1.4	94	0.00
37	2-Methylnaphthalene	40.000	39.488	1.3	95	0.00
38	1-Methylnaphthalene	40.000	39.735	0.7	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.811	0.5	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.954	-7.4	96	0.00
42 S	2,4,6-Tribromophenol	80.000	81.284	-1.6	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.484	1.3	92	0.00
44	2,4,5-Trichlorophenol	40.000	40.094	-0.2	94	0.00
45 S	2-Fluorobiphenyl	80.000	76.905	3.9	95	0.00
46	1,1'-Biphenyl	40.000	38.780	3.0	94	0.00
47	2-Chloronaphthalene	40.000	39.146	2.1	94	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043021\
 Data File : BF123775.D
 Acq On : 30 Apr 2021 17:54
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Apr 30 18:14:11 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Apr 30 14:46:08 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.839	0.4	95	0.00
49	Acenaphthylene	40.000	39.338	1.7	95	0.00
50	Dimethylphthalate	40.000	39.970	0.1	98	0.00
51	2,6-Dinitrotoluene	40.000	40.727	-1.8	97	0.00
52 C	Acenaphthene	40.000	39.470	1.3	96	0.00
53	3-Nitroaniline	40.000	39.704	0.7	94	0.00
54 P	2,4-Dinitrophenol	40.000	38.546	3.6	85	0.00
55	Dibenzofuran	40.000	39.182	2.0	94	0.00
56 P	4-Nitrophenol	40.000	39.591	1.0	90	0.00
57	2,4-Dinitrotoluene	40.000	40.464	-1.2	96	0.00
58	Fluorene	40.000	39.518	1.2	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.475	-1.2	94	0.00
60	Diethylphthalate	40.000	39.894	0.3	96	0.00
61	4-Chlorophenyl-phenylether	40.000	39.075	2.3	96	0.00
62	4-Nitroaniline	40.000	39.883	0.3	94	0.00
63	Azobenzene	40.000	39.557	1.1	95	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	94	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.479	-3.7	93	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.287	1.8	94	0.00
67	4-Bromophenyl-phenylether	40.000	40.488	-1.2	97	0.00
68	Hexachlorobenzene	40.000	41.191	-3.0	101	0.00
69	Atrazine	40.000	39.323	1.7	94	0.00
70 C	Pentachlorophenol	40.000	39.084	2.3	91	0.00
71	Phenanthrene	40.000	38.854	2.9	95	0.00
72	Anthracene	40.000	39.153	2.1	96	0.00
73	Carbazole	40.000	38.965	2.6	95	0.00
74	Di-n-butylphthalate	40.000	40.053	-0.1	95	0.00
75 C	Fluoranthene	40.000	39.190	2.0	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	43.945	-9.9	93	0.00
78	Pyrene	40.000	40.942	-2.4	96	0.00
79 S	Terphenyl-d14	80.000	80.391	-0.5	95	0.00
80	Butylbenzylphthalate	40.000	40.980	-2.4	93	0.00
81	Benzo(a)anthracene	40.000	39.051	2.4	92	0.00
82	3,3'-Dichlorobenzidine	40.000	39.719	0.7	92	0.00
83	Chrysene	40.000	38.497	3.8	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.120	-0.3	93	0.00
85 c	Di-n-octyl phthalate	40.000	39.288	1.8	89	0.00
86 I	Perylene-d12	20.000	20.000	0.0	87	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.720	-1.8	96	0.00
88	Benzo(b)fluoranthene	40.000	39.294	1.8	88	0.00
89	Benzo(k)fluoranthene	40.000	41.346	-3.4	89	0.00
90 C	Benzo(a)pyrene	40.000	39.634	0.9	87	0.00
91	Dibenzo(a,h)anthracene	40.000	39.944	0.1	94	0.00
92	Benzo(g,h,i)perylene	40.000	41.097	-2.7	95	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043021\
Data File : BF123775.D
Acq On : 30 Apr 2021 17:54
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Apr 30 18:14:11 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Apr 30 14:46:08 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0