

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
 Data File : BF137764.D
 Acq On : 30 Apr 2024 11:08
 Operator : CG\JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: May 01 02:55:00 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 02:54:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.057	152	258875	20.000	ng	0.00
21) Naphthalene-d8	8.339	136	1035499	20.000	ng	0.00
39) Acenaphthene-d10	10.104	164	613407	20.000	ng	0.00
64) Phenanthrene-d10	11.598	188	1097512	20.000	ng	0.00
76) Chrysene-d12	14.251	240	922774	20.000	ng	0.00
86) Perylene-d12	15.827	264	879851	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.445	70	31383	2.678	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
Data File : BF137764.D
Acq On : 30 Apr 2024 11:08
Operator : CG\JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: May 01 02:55:00 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 02:54:23 2024
Response via : Initial Calibration

