

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
 Data File : BF137773.D
 Acq On : 30 Apr 2024 15:52
 Operator : CG\JU
 Sample : PB160553TB
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB160553TB

Quant Time: May 01 03:28:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.057	152	251347	20.000	ng	0.00
21) Naphthalene-d8	8.339	136	981552	20.000	ng	0.00
39) Acenaphthene-d10	10.104	164	562072	20.000	ng	0.00
64) Phenanthrene-d10	11.598	188	1043850	20.000	ng	0.00
76) Chrysene-d12	14.251	240	866522	20.000	ng	# 0.00
86) Perylene-d12	15.821	264	747506	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.692	112	1834846	122.224	ng	0.02
7) Phenol-d6	6.669	99	2314294	121.540	ng	0.00
23) Nitrobenzene-d5	7.616	82	1423865	84.279	ng	0.00
42) 2,4,6-Tribromophenol	10.898	330	751187	130.894	ng	0.00
45) 2-Fluorobiphenyl	9.416	172	2860486	82.565	ng	0.00
79) Terphenyl-d14	13.192	244	3743574	72.401	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
Data File : BF137773.D
Acq On : 30 Apr 2024 15:52
Operator : CG\JU
Sample : PB160553TB
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB160553TB

Quant Time: May 01 03:28:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration

