

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
 Data File : BF137772.D
 Acq On : 30 Apr 2024 15:18
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043024

Quant Time: May 01 03:27:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	94	0.00
2	1,4-Dioxane	40.000	39.362	1.6	94	0.00
3	Pyridine	40.000	40.698	-1.7	96	0.00
4	n-Nitrosodimethylamine	40.000	38.883	2.8	93	0.00
5 S	2-Fluorophenol	80.000	78.649	1.7	95	0.00
6	Aniline	40.000	39.784	0.5	93	0.00
7 S	Phenol-d6	80.000	77.397	3.3	93	0.00
8	2-Chlorophenol	40.000	39.189	2.0	94	0.00
9	Benzaldehyde	40.000	38.924	2.7	95	0.00
10 C	Phenol	40.000	39.423	1.4	95	0.00
11	bis(2-Chloroethyl)ether	40.000	39.271	1.8	94	0.00
12	1,3-Dichlorobenzene	40.000	39.446	1.4	95	0.00
13 C	1,4-Dichlorobenzene	40.000	39.206	2.0	94	0.00
14	1,2-Dichlorobenzene	40.000	39.339	1.7	94	0.00
15	Benzyl Alcohol	40.000	40.206	-0.5	93	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.253	1.9	93	0.00
17	2-Methylphenol	40.000	39.628	0.9	94	0.00
18	Hexachloroethane	40.000	40.397	-1.0	95	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.370	1.6	94	0.00
20	3+4-Methylphenols	40.000	39.528	1.2	93	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	95	0.00
22	Acetophenone	40.000	39.103	2.2	93	0.00
23 S	Nitrobenzene-d5	80.000	80.368	-0.5	95	0.00
24	Nitrobenzene	40.000	39.506	1.2	93	0.00
25	Isophorone	40.000	39.534	1.2	95	0.00
26 C	2-Nitrophenol	40.000	43.012	-7.5	96	0.00
27	2,4-Dimethylphenol	40.000	39.851	0.4	95	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.495	1.3	94	0.00
29 C	2,4-Dichlorophenol	40.000	40.067	-0.2	94	0.00
30	1,2,4-Trichlorobenzene	40.000	39.448	1.4	94	0.00
31	Naphthalene	40.000	39.245	1.9	94	0.00
32	Benzoic acid	40.000	43.716	-9.3	101	0.00
33	4-Chloroaniline	40.000	39.510	1.2	94	0.00
34 C	Hexachlorobutadiene	40.000	39.616	1.0	95	0.00
35	Caprolactam	40.000	41.442	-3.6	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.131	-0.3	95	0.00
37	2-Methylnaphthalene	40.000	39.766	0.6	95	0.00
38	1-Methylnaphthalene	40.000	39.341	1.6	94	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.434	1.4	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.272	-3.2	95	0.00
42 S	2,4,6-Tribromophenol	80.000	81.140	-1.4	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.807	0.5	95	0.00
44	2,4,5-Trichlorophenol	40.000	40.306	-0.8	95	0.00
45 S	2-Fluorobiphenyl	80.000	76.875	3.9	96	0.00
46	1,1'-Biphenyl	40.000	39.273	1.8	96	0.00
47	2-Chloronaphthalene	40.000	39.345	1.6	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
 Data File : BF137772.D
 Acq On : 30 Apr 2024 15:18
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF043024

Quant Time: May 01 03:27:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.318	-3.3	96	0.00
49	Acenaphthylene	40.000	39.435	1.4	96	0.00
50	Dimethylphthalate	40.000	40.062	-0.2	97	0.00
51	2,6-Dinitrotoluene	40.000	41.382	-3.5	98	0.00
52 C	Acenaphthene	40.000	39.545	1.1	96	0.00
53	3-Nitroaniline	40.000	41.431	-3.6	98	0.00
54 P	2,4-Dinitrophenol	40.000	45.402	-13.5	106	0.00
55	Dibenzofuran	40.000	39.624	0.9	97	0.00
56 P	4-Nitrophenol	40.000	42.621	-6.6	100	0.00
57	2,4-Dinitrotoluene	40.000	42.370	-5.9	99	0.00
58	Fluorene	40.000	39.097	2.3	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.452	-1.1	96	0.00
60	Diethylphthalate	40.000	40.141	-0.4	98	0.00
61	4-Chlorophenyl-phenylether	40.000	39.822	0.4	98	0.00
62	4-Nitroaniline	40.000	40.919	-2.3	96	0.00
63	Azobenzene	40.000	39.859	0.4	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.718	-9.3	104	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.878	2.8	97	0.00
67	4-Bromophenyl-phenylether	40.000	39.820	0.4	99	0.00
68	Hexachlorobenzene	40.000	39.276	1.8	98	0.00
69	Atrazine	40.000	39.972	0.1	100	0.00
70 C	Pentachlorophenol	40.000	41.273	-3.2	99	0.00
71	Phenanthrene	40.000	38.456	3.9	97	0.00
72	Anthracene	40.000	38.927	2.7	98	0.00
73	Carbazole	40.000	39.010	2.5	99	0.00
74	Di-n-butylphthalate	40.000	40.505	-1.3	101	0.00
75 C	Fluoranthene	40.000	38.943	2.6	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	39.925	0.2	86	0.00
78	Pyrene	40.000	38.903	2.7	98	0.00
79 S	Terphenyl-d14	80.000	78.298	2.1	101	0.00
80	Butylbenzylphthalate	40.000	43.513	-8.8	105	0.00
81	Benzo(a)anthracene	40.000	39.569	1.1	101	0.00
82	3,3'-Dichlorobenzidine	40.000	38.865	2.8	100	0.00
83	Chrysene	40.000	39.193	2.0	101	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.600	-6.5	107	0.00
85 c	Di-n-octyl phthalate	40.000	43.485	-8.7	111	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.757	-1.9	102	0.00
88	Benzo(b)fluoranthene	40.000	40.406	-1.0	102	0.00
89	Benzo(k)fluoranthene	40.000	41.257	-3.1	101	0.00
90 C	Benzo(a)pyrene	40.000	40.373	-0.9	102	0.00
91	Dibenzo(a,h)anthracene	40.000	40.714	-1.8	101	0.00
92	Benzo(g,h,i)perylene	40.000	40.624	-1.6	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF043024\
Data File : BF137772.D
Acq On : 30 Apr 2024 15:18
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF043024

Quant Time: May 01 03:27:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0