


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050115\
Data File : BF078882.D
Acq On : 1 May 2015 9:54
Operator : TP/IZ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF043015.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Apr 30 19:07:47 2015



AutoFind: Scans 1061, 1062, 1063; Background Corrected with Scan 1058

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.6	11330	PASS
68	69	0.00	2	2.0	195	PASS
69	198	0.00	100	36.7	9759	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	50.9	13542	PASS
197	198	0.00	2	0.5	125	PASS
198	198	100	100	100.0	26584	PASS
199	198	5	9	6.7	1787	PASS
275	198	10	60	25.0	6637	PASS
365	198	1	100	3.1	831	PASS
441	442	0.01	24	15.7	3925	PASS
442	198	50	100	94.3	25079	PASS
443	442	15	24	19.1	4790	PASS