

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050117\
 Data File : BF094764.D
 Acq On : 1 May 2017 22:15
 Operator : SJ/MA
 Sample : I2914-14
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 02 01:41:37 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Apr 17 14:59:23 2017
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	5.77	152	108079	20.00	ng	0.00
21) Naphthalene-d8	7.25	136	426761	20.00	ng	0.00
38) Acenaphthene-d10	9.39	164	169343	20.00	ng	0.00
63) Phenanthrene-d10	11.21	188	256603	20.00	ng	0.00
75) Chrysene-d12	14.47	240	210434	20.00	ng	0.00
86) Perylene-d12	16.09	264	156395	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	4.34	112	793153	121.75	ng	0.03
7) Phenol-d6	5.40	99	943524	122.86	ng	0.00
23) Nitrobenzene-d5	6.41	82	566171	81.92	ng	0.00
41) 2,4,6-Tribromophenol	10.37	330	184821	139.53	ng	0.00
44) 2-Fluorobiphenyl	8.58	172	1106752	96.16	ng	0.00
78) Terphenyl-d14	13.19	244	780212	99.10	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propylamine	6.41	70	73517	14.45	ng	# 78
50) 2,6-Dinitrotoluene	9.39	165	21316	7.53	ng	# 31
56) 2,4-Dinitrotoluene	9.39	165	21316	6.13	ng	# 31

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF050117\
Data File : BF094764.D
Acq On : 1 May 2017 22:15
Operator : SJ/MA
Sample : I2914-14
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: May 02 01:41:37 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF041717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Apr 17 14:59:23 2017
Response via : Initial Calibration






#1



1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 5.77 min Scan# 651
Delta R.T. 0.00 min
Lab File: BF094764.D
Acq: 1 May 2017 22:15

Instrument :
BNA_F
ClientSampleId :

Tot Ion:152 Resp: 108079
Ion Ratio Lower Upper
152 100
150 160.0 126.2 189.2
115 59.9 52.2 78.2




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 121.75 ng
RT: 4.34 min Scan# 408
Delta R.T. 0.03 min
Lab File: BF094764.D
Acq: 1 May 2017 22:15

Tgt Ion:112 Resp: 793153
Ion Ratio Lower Upper
112 100
64 56.5 46.0 69.0
63 28.4 23.4 35.0



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 122.86 ng

RT: 5.40 min Scan# 589

Delta R.T. 0.01 min

Lab File: BF094764.D

Acq: 1 May 2017 22:15

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 99 Resp: 943524

Ion Ratio Lower Upper

99 100

42 17.6 13.5 20.3




71 33.4 24.5 36.7



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#19

n-Nitroso-di-n-propylamine

Concen: 14.45 ng

RT: 6.41 min Scan# 761

Delta R.T. 0.16 min

Lab File: BF094764.D

Acq: 1 May 2017 22:15

Tgt Ion: 70 Resp: 73517


Ion Ratio Lower Upper

70 100

42 46.7 47.2 70.8#

101 0.0 7.2 10.8#

130 2.3 15.9 23.9#






Abundance Ion 70.00 (69.70 to 70.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 101.00 (100.70 to 101.70): BF0

Ion 130.00 (129.70 to 130.70): BF0





#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.25 min Scan# 904

Delta R.T. -0.01 min

Lab File: BF094764.D

Acq: 1 May 2017 22:15

Instrument :

BNA_F

ClientSampleId :

Tot Ion:136 Resp: 426761


Ion Ratio Lower Upper

136 100

137 10.5 8.5 12.7

54 6.7 4.9 7.3

68 5.0 4.1 6.1






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0



#23

Nitrobenzene-d5

Concen: 81.92 ng

RT: 6.41 min Scan# 761

Delta R.T. -0.00 min

Lab File: BF094764.D

Acq: 1 May 2017 22:15


Tgt Ion: 82 Resp: 566171

Ion Ratio Lower Upper

82 100

128 46.5 34.0 51.0



54 49.5 42.2 63.2



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0









#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.39 min Scan# 1267
 Delta R.T. -0.01 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion:164 Resp: 169343
 Ion Ratio Lower Upper
 164 100
 162 105.8 82.6 123.8
 160 46.6 36.2 54.2




Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E






#41
 2,4,6-Tribromophenol
 Concen: 139.53 ng
 RT: 10.37 min Scan# 1434
 Delta R.T. -0.00 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15

Tgt Ion:330 Resp: 184821
 Ion Ratio Lower Upper
 330 100
 332 94.7 76.6 115.0
 141 38.4 31.4 47.2



Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E









#44
 2-Fluorobiphenyl
 Concen: 96.16 ng
 RT: 8.58 min Scan# 1129
 Delta R.T. -0.01 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15

Instrument :
 BNA_F
 ClientSampled :

Tot Ion:172 Resp: 1106752
 Ion Ratio Lower Upper
 172 100
 171 35.9 29.6 44.4
 170 23.8 19.8 29.8




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#50
 2,6-Dinitrotoluene
 Concen: 7.53 ng
 RT: 9.39 min Scan# 1267
 Delta R.T. 0.23 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15

Tgt Ion:165 Resp: 21316
 Ion Ratio Lower Upper
 165 100
 63 0.0 34.3 51.5#
 89 0.0 37.1 55.7#



Abundance Ion 165.00 (164.70 to 165.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 89.00 (88.70 to 89.70): BF0









#56
 2,4-Dinitrotoluene
 Concen: 6.13 ng
 RT: 9.39 min Scan# 1267
 Delta R.T. -0.27 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:	165	Resp:	21316
Ion	Ratio	Lower	Upper
165	100		
63	0.0	25.4	38.0#
89	0.0	46.4	69.6#
182	0.0	1.8	2.6#



#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.21 min Scan# 1577
 Delta R.T. -0.01 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15

Tgt Ion:	188	Resp:	256603
Ion	Ratio	Lower	Upper
188	100		
94	9.2	9.4	14.2#
80	9.5	10.2	15.2#





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.47 min Scan# 2130

Delta R.T. -0.00 min

Lab File: BF094764.D

Acq: 1 May 2017 22:15

Instrument :

BNA_F

ClientSampled :


Tot Ion:240 Resp: 210434

Ion Ratio Lower Upper

240 100

120 12.6 5.8 8.8#




236 25.4 20.5 30.7



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 99.10 ng

RT: 13.19 min Scan# 1913

Delta R.T. -0.01 min

Lab File: BF094764.D

Acq: 1 May 2017 22:15


Tgt Ion:244 Resp: 780212

Ion Ratio Lower Upper

244 100

212 7.1 6.0 9.0



122 14.4 10.3 15.5




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E






#86
 Pervlene-d12
 Concen: 20.00 ng
 RT: 16.09 min Scan# 2407
 Delta R.T. 0.01 min
 Lab File: BF094764.D
 Acq: 1 May 2017 22:15

Instrument :
 BNA_F
 ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	24.4	19.5	29.3
265	21.3	17.2	25.8



Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

