


Data Path : Z:\HPCHEM1\BNA F\DATA\BF050117\
 Data File : BF094743.D
 Acq On : 1 May 2017 11:11
 Operator : SJ/MA
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF041717.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 17 14:59:23 2017



AutoFind: Scans 1642, 1643, 1644; Background Corrected with Scan 1637

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.8	26026	PASS
68	69	0.00	2	2.5	647	FAIL*
69	198	0.00	100	35.7	25978	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.8	35509	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	72768	PASS
199	198	5	9	6.7	4857	PASS
275	198	10	60	22.8	16589	PASS
365	198	1	100	2.1	1550	PASS
441	442	0.01	24	16.0	9101	PASS
442	198	50	100	78.0	56736	PASS
443	442	15	24	19.8	11246	PASS