

Data Path : Z:\HPCHEM1\BNA\_F\Data\BF050118\  
 Data File : BF104976.D  
 Acq On : 1 May 2018 11:50  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: May 01 16:20:13 2018  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF040618.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue May 01 16:13:08 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                    | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|------|-----------------------------|-------|-------|--------|-------|----------|
| 1 I  | 1,4-Dichlorobenzene-d4      | 1.000 | 1.000 | 0.0    | 87    | 0.00     |
| 2    | 1,4-Dioxane                 | 0.609 | 0.534 | 12.3   | 75    | 0.00     |
| 3    | Pyridine                    | 1.714 | 1.497 | 12.7   | 76    | 0.00     |
| 4    | n-Nitrosodimethylamine      | 0.599 | 0.498 | 16.9   | 72    | 0.00     |
| 5 S  | 2-Fluorophenol              | 1.308 | 1.235 | 5.6    | 83    | 0.00     |
| 6    | Aniline                     | 2.233 | 2.095 | 6.2    | 81    | 0.00     |
| 7 S  | Phenol-d6                   | 1.607 | 1.568 | 2.4    | 86    | 0.00     |
| 8    | 2-Chlorophenol              | 1.381 | 1.410 | -2.1   | 89    | 0.00     |
| 9    | Benzaldehyde                | 0.802 | 0.856 | -6.7   | 89    | 0.00     |
| 10 C | Phenol                      | 1.705 | 1.726 | -1.2   | 90    | 0.00     |
| 11   | bis(2-Chloroethyl)ether     | 1.361 | 1.377 | -1.2   | 88    | 0.00     |
| 12   | 1,3-Dichlorobenzene         | 1.564 | 1.590 | -1.7   | 89    | 0.00     |
| 13 C | 1,4-Dichlorobenzene         | 1.559 | 1.603 | -2.8   | 90    | 0.00     |
| 14   | 1,2-Dichlorobenzene         | 1.445 | 1.509 | -4.4   | 90    | 0.00     |
| 15   | Benzyl Alcohol              | 1.221 | 1.127 | 7.7    | 81    | 0.00     |
| 16   | 2,2'-oxybis(1-Chloropropane | 1.827 | 1.567 | 14.2   | 75    | 0.00     |
| 17   | 2-Methylphenol              | 1.200 | 1.222 | -1.8   | 89    | 0.00     |
| 18   | Hexachloroethane            | 0.556 | 0.574 | -3.2   | 90    | 0.00     |
| 19 P | n-Nitroso-di-n-propylamine  | 1.050 | 0.948 | 9.7    | 82    | 0.00     |
| 20   | 3+4-Methylphenols           | 1.488 | 1.473 | 1.0    | 87    | 0.00     |
| 21 I | Naphthalene-d8              | 1.000 | 1.000 | 0.0    | 92    | 0.00     |
| 22   | Acetophenone                | 0.492 | 0.464 | 5.7    | 88    | 0.00     |
| 23 S | Nitrobenzene-d5             | 0.306 | 0.332 | -8.5   | 97    | 0.00     |
| 24   | Nitrobenzene                | 0.338 | 0.341 | -0.9   | 89    | 0.00     |
| 25   | Isophorone                  | 0.648 | 0.606 | 6.5    | 87    | 0.00     |
| 26 C | 2-Nitrophenol               | 0.138 | 0.189 | -37.0# | 120   | 0.00     |
| 27   | 2,4-Dimethylphenol          | 0.286 | 0.262 | 8.4    | 83    | 0.00     |
| 28   | bis(2-Chloroethoxy)methane  | 0.396 | 0.376 | 5.1    | 88    | 0.00     |
| 29 C | 2,4-Dichlorophenol          | 0.273 | 0.276 | -1.1   | 92    | 0.00     |
| 30   | 1,2,4-Trichlorobenzene      | 0.299 | 0.307 | -2.7   | 95    | 0.00     |
| 31   | Naphthalene                 | 0.996 | 0.977 | 1.9    | 91    | 0.00     |
| 32   | Benzoic acid                | 0.228 | 0.235 | -3.1   | 89    | 0.00     |
| 33   | 4-Chloroaniline             | 0.427 | 0.423 | 0.9    | 92    | 0.00     |
| 34 C | Hexachlorobutadiene         | 0.164 | 0.173 | -5.5   | 98    | 0.00     |
| 35   | Caprolactam                 | 0.095 | 0.090 | 5.3    | 85    | 0.00     |
| 36 C | 4-Chloro-3-methylphenol     | 0.292 | 0.290 | 0.7    | 91    | 0.00     |
| 37   | 2-Methylnaphthalene         | 0.644 | 0.636 | 1.2    | 92    | 0.00     |
| 38 I | Acenaphthene-d10            | 1.000 | 1.000 | 0.0    | 91    | 0.00     |
| 39   | 1,2,4,5-Tetrachlorobenzene  | 0.573 | 0.611 | -6.6   | 100   | 0.00     |
| 40 P | Hexachlorocyclopentadiene   | 0.321 | 0.358 | -11.5  | 101   | 0.00     |
| 41 S | 2,4,6-Tribromophenol        | 0.207 | 0.211 | -1.9   | 93    | 0.00     |
| 42 C | 2,4,6-Trichlorophenol       | 0.397 | 0.402 | -1.3   | 91    | 0.00     |
| 43   | 2,4,5-Trichlorophenol       | 0.445 | 0.457 | -2.7   | 94    | 0.00     |
| 44 S | 2-Fluorobiphenyl            | 1.266 | 1.261 | 0.4    | 93    | 0.00     |

Data Path : Z:\HPCHEM1\BNA\_F\Data\BF050118\  
 Data File : BF104976.D  
 Acq On : 1 May 2018 11:50  
 Operator : JU/SJ  
 Sample : SSTDCCC040  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 LabSampleId :  
 SSTDCCC040

Quant Time: May 01 16:20:13 2018  
 Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF040618.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue May 01 16:13:08 2018  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound                   | AvgRF | CCRF  | %Dev   | Area% | Dev(min) |
|------|----------------------------|-------|-------|--------|-------|----------|
| 45   | 1,1'-Biphenyl              | 1.680 | 1.686 | -0.4   | 94    | 0.00     |
| 46   | 2-Chloronaphthalene        | 1.327 | 1.339 | -0.9   | 94    | 0.00     |
| 47   | 2-Nitroaniline             | 0.328 | 0.385 | -17.4  | 101   | 0.00     |
| 48   | Acenaphthylene             | 2.082 | 1.982 | 4.8    | 88    | 0.00     |
| 49   | Dimethylphthalate          | 1.577 | 1.595 | -1.1   | 95    | 0.00     |
| 50   | 2,6-Dinitrotoluene         | 0.292 | 0.340 | -16.4  | 99    | 0.00     |
| 51 C | Acenaphthene               | 1.270 | 1.182 | 6.9    | 87    | 0.00     |
| 52   | 3-Nitroaniline             | 0.342 | 0.391 | -14.3  | 98    | 0.00     |
| 53 P | 2,4-Dinitrophenol          | 0.089 | 0.135 | -51.7# | 142   | 0.00     |
| 54   | Dibenzofuran               | 1.770 | 1.624 | 8.2    | 85    | 0.00     |
| 55 P | 4-Nitrophenol              | 0.268 | 0.275 | -2.6   | 87    | 0.00     |
| 56   | 2,4-Dinitrotoluene         | 0.383 | 0.411 | -7.3   | 94    | 0.00     |
| 57   | Fluorene                   | 1.289 | 1.349 | -4.7   | 96    | 0.00     |
| 58   | 2,3,4,6-Tetrachlorophenol  | 0.332 | 0.327 | 1.5    | 91    | 0.00     |
| 59   | Diethylphthalate           | 1.571 | 1.487 | 5.3    | 89    | 0.00     |
| 60   | 4-Chlorophenyl-phenylether | 0.574 | 0.617 | -7.5   | 98    | 0.00     |
| 61   | 4-Nitroaniline             | 0.334 | 0.395 | -18.3  | 99    | 0.00     |
| 62   | Azobenzene                 | 1.501 | 1.318 | 12.2   | 81    | 0.00     |
| 63 I | Phenanthrene-d10           | 1.000 | 1.000 | 0.0    | 90    | 0.00     |
| 64   | 4,6-Dinitro-2-methylphenol | 0.087 | 0.134 | -54.0# | 132   | 0.00     |
| 65 c | n-Nitrosodiphenylamine     | 0.693 | 0.680 | 1.9    | 90    | 0.00     |
| 66   | 4-Bromophenyl-phenylether  | 0.213 | 0.217 | -1.9   | 93    | 0.00     |
| 67   | Hexachlorobenzene          | 0.239 | 0.247 | -3.3   | 95    | 0.00     |
| 68   | Atrazine                   | 0.209 | 0.198 | 5.3    | 87    | 0.00     |
| 69 C | Pentachlorophenol          | 0.148 | 0.148 | 0.0    | 86    | 0.00     |
| 70   | Phenanthrene               | 1.114 | 1.058 | 5.0    | 87    | 0.00     |
| 71   | Anthracene                 | 1.132 | 1.091 | 3.6    | 89    | 0.00     |
| 72   | Carbazole                  | 1.106 | 1.044 | 5.6    | 87    | 0.00     |
| 73   | Di-n-butylphthalate        | 1.351 | 1.259 | 6.8    | 86    | 0.00     |
| 74 C | Fluoranthene               | 1.164 | 1.075 | 7.6    | 85    | 0.00     |
| 75 I | Chrysene-d12               | 1.000 | 1.000 | 0.0    | 81    | 0.00     |
| 76   | Benzidine                  | 0.692 | 0.615 | 11.1   | 72    | 0.00     |
| 77   | Pyrene                     | 1.482 | 1.539 | -3.8   | 85    | 0.00     |
| 78 S | Terphenyl-d14              | 0.877 | 0.906 | -3.3   | 87    | 0.00     |
| 79   | Butylbenzylphthalate       | 0.698 | 0.717 | -2.7   | 83    | 0.00     |
| 80   | Benzo(a)anthracene         | 1.195 | 1.247 | -4.4   | 87    | 0.00     |
| 81   | 3,3'-Dichlorobenzidine     | 0.445 | 0.418 | 6.1    | 74    | 0.00     |
| 82   | Chrysene                   | 1.227 | 1.179 | 3.9    | 78    | 0.00     |
| 83   | Bis(2-ethylhexyl)phthalate | 0.898 | 0.931 | -3.7   | 85    | 0.00     |
| 84 c | Di-n-octyl phthalate       | 1.501 | 1.384 | 7.8    | 73    | 0.00     |
| 85   | Indeno(1,2,3-cd)pyrene     | 1.043 | 0.928 | 11.0   | 75    | 0.00     |
| 86 I | Perylene-d12               | 1.000 | 1.000 | 0.0    | 68    | 0.00     |
| 87   | Benzo(b)fluoranthene       | 1.263 | 1.353 | -7.1   | 73    | 0.00     |

Data Path : Z:\HPCHEM1\BNA\_F\Data\BF050118\  
Data File : BF104976.D  
Acq On : 1 May 2018 11:50  
Operator : JU/SJ  
Sample : SSTDCCC040  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_F  
LabSampleId :  
SSTDCCC040

Quant Time: May 01 16:20:13 2018  
Quant Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF040618.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue May 01 16:13:08 2018  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

|      | Compound               | AvgRF | CCRF  | %Dev  | Area% | Dev(min) |
|------|------------------------|-------|-------|-------|-------|----------|
| 88   | Benzo(k)fluoranthene   | 1.193 | 1.166 | 2.3   | 68    | 0.00     |
| 89 C | Benzo(a)pyrene         | 1.130 | 1.148 | -1.6  | 70    | 0.00     |
| 90   | Dibenzo(a,h)anthracene | 0.928 | 0.964 | -3.9  | 74    | 0.00     |
| 91   | Benzo(g,h,i)perylene   | 0.899 | 0.989 | -10.0 | 80    | 0.00     |

(#) = Out of Range

SPCC's out = 0 CCC's out = 1