


Data Path : Z:\HPCHEM1\BNA F\Data\BF050216\
 Data File : BF086613.D
 Acq On : 2 May 2016 13:12
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF042716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 27 14:20:52 2016



AutoFind: Scans 826, 827, 828; Background Corrected with Scan 823

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	60.6	46265	PASS
68	69	0.00	2	1.8	641	PASS
69	198	0.00	100	46.0	35080	PASS
70	69	0.00	2	0.5	178	PASS
127	198	10	80	54.1	41278	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	76296	PASS
199	198	5	9	6.8	5181	PASS
275	198	10	60	24.7	18852	PASS
365	198	1	100	3.5	2636	PASS
441	442	0.01	24	15.8	10192	PASS
442	198	50	100	84.6	64574	PASS
443	442	15	24	18.2	11776	PASS