

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050218\
 Data File : BF105028.D
 Acq On : 2 May 2018 14:03
 Operator : JU/SJ
 Sample : J2652-12MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS

Quant Time: May 03 04:22:08 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 02 17:25:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	91290	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	371150	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	171182	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	284058	20.00	ng	0.00
75) Chrysene-d12	14.04	240	198832	20.00	ng	0.00
86) Perylene-d12	15.61	264	195608	20.00	ng	0.02

System Monitoring Compounds

5) 2-Fluorophenol	5.41	112	146579	24.55	ng	0.00
7) Phenol-d6	6.43	99	453744	61.85	ng	0.00
23) Nitrobenzene-d5	7.36	82	501011	88.15	ng	0.00
41) 2,4,6-Tribromophenol	10.63	330	39099	22.02	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	813786	75.10	ng	0.00
78) Terphenyl-d14	12.92	244	681427	78.13	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.53	88	92087	33.102	ng	# 86
3) Pyridine	3.26	79	261254	33.402	ng	85
4) n-Nitrosodimethylamine	3.23	42	101840	37.248	ng	# 74
6) Aniline	6.45	93	193206	18.957	ng	# 22
8) 2-Chlorophenol	6.58	128	143069	22.704	ng	93
9) Benzaldehyde	6.35	77	178298	54.158	ng	93
10) Phenol	6.45	94	267015	34.306	ng	# 72
11) bis(2-Chloroethyl)ether	6.53	93	257733	41.495	ng	91
12) 1,3-Dichlorobenzene	6.73	146	289580	40.564	ng	99
13) 1,4-Dichlorobenzene	6.80	146	295658	41.547	ng	99
14) 1,2-Dichlorobenzene	6.96	146	276589	41.941	ng	97
15) Benzyl Alcohol	6.94	79	216412	38.842	ng	98
16) 2,2'-oxybis(1-Chloropropan	7.06	45	278932	33.440	ng	# 3
17) 2-Methylphenol	7.06	107	215157	39.279	ng	# 88
18) Hexachloroethane	7.30	117	92369	36.405	ng	96
19) n-Nitroso-di-n-propylamine	7.21	70	176928	36.910	ng	92
20) 3+4-Methylphenols	7.22	107	277541	40.854	ng	95
22) Acetophenone	7.20	105	377239	41.285	ng	94
24) Nitrobenzene	7.38	77	274313	43.712	ng	96
25) Isophorone	7.62	82	490654	40.822	ng	98
26) 2-Nitrophenol	7.69	139	50842	19.901	ng	92
27) 2,4-Dimethylphenol	7.74	122	240675	45.371	ng	99
28) bis(2-Chloroethoxy)methane	7.83	93	321417	43.737	ng	99
29) 2,4-Dichlorophenol	7.95	162	111519	22.033	ng	97
30) 1,2,4-Trichlorobenzene	8.02	180	238261	42.894	ng	99
31) Naphthalene	8.10	128	818267	44.275	ng	100
32) Benzoic acid	7.74	122	240675	56.864	ng	# 13
33) 4-Chloroaniline	8.16	127	155032	19.580	ng	96
34) Hexachlorobutadiene	8.20	225	127442	41.887	ng	99
35) Caprolactam	8.51	113	73886	41.846	ng	# 78
36) 4-Chloro-3-methylphenol	8.65	107	225735	41.594	ng	99
37) 2-Methylnaphthalene	8.79	142	530753	44.397	ng	99
39) 1,2,4,5-Tetrachlorobenzene	8.96	216	223744	45.660	ng	99
40) Hexachlorocyclopentadiene	8.93	237	18029	6.572	ng	96


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF050218\
 Data File : BF105028.D
 Acq On : 2 May 2018 14:03
 Operator : JU/SJ
 Sample : J2652-12MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS

Quant Time: May 03 04:22:08 2018
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF040618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 02 17:25:18 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	9.07	196	34788	10.227	ng	99
43) 2,4,5-Trichlorophenol	9.12	196	56290	14.788	ng	96
45) 1,1'-Biphenyl	9.25	154	609235	42.365	ng	99
46) 2-Chloronaphthalene	9.28	162	478527	42.128	ng	99
47) 2-Nitroaniline	9.39	65	139200	49.555	ng	93
48) Acenaphthylene	9.70	152	718092	40.294	ng	99
49) Dimethylphthalate	9.56	163	664368	49.216	ng	99
50) 2,6-Dinitrotoluene	9.63	165	123286	49.357	ng	92
51) Acenaphthene	9.87	154	434697	39.977	ng	98
52) 3-Nitroaniline	9.80	138	112913	38.613	ng	94
54) Dibenzofuran	10.04	168	629108	41.516	ng	99
55) 4-Nitrophenol	9.99	139	10455	4.551	ng	96
56) 2,4-Dinitrotoluene	10.03	165	148640	45.366	ng	95
57) Fluorene	10.39	166	516645	46.841	ng	100
58) 2,3,4,6-Tetrachlorophenol	10.16	232	24025	8.460	ng	# 73
59) Diethylphthalate	10.25	149	534126	39.729	ng	98
60) 4-Chlorophenyl-phenylether	10.37	204	227224	46.258	ng	94
61) 4-Nitroaniline	10.42	138	135401	47.356	ng	96
62) Azobenzene	10.53	77	463917	36.118	ng	94
64) 4,6-Dinitro-2-methylphenol	10.45	198	1771	8.431	ng	95
65) n-Nitrosodiphenylamine	10.49	169	438402	44.512	ng	100
66) 4-Bromophenyl-phenylether	10.86	248	135461	44.833	ng	96
67) Hexachlorobenzene	10.93	284	145925	42.922	ng	# 90
68) Atrazine	11.03	200	140110	47.129	ng	98
69) Pentachlorophenol	11.13	266	21401	10.175	ng	97
70) Phenanthrene	11.35	178	898875	56.822	ng	99
71) Anthracene	11.40	178	729988	45.397	ng	98
72) Carbazole	11.56	167	620817	39.509	ng	100
73) Di-n-butylphthalate	11.88	149	799588	41.657	ng	99
74) Fluoranthene	12.54	202	895807	54.200	ng	99
76) Benzidine	12.67	184	302824	47.204	ng	98
77) Pyrene	12.78	202	878494	59.607	ng	99
79) Butylbenzylphthalate	13.40	149	293170	42.223	ng	100
80) Benzo(a)anthracene	14.03	228	678495	57.120	ng	99
81) 3,3'-Dichlorobenzidine	13.99	252	150958	34.085	ng	99
82) Chrysene	14.07	228	630245	51.655	ng	100
83) Bis(2-ethylhexyl)phthalate	14.00	149	421017	47.142	ng	100
84) Di-n-octyl phthalate	14.70	149	692496	46.405	ng	# 94
85) Indeno(1,2,3-cd)pyrene	17.11	276	506257	48.845	ng	95
87) Benzo(b)fluoranthene	15.18	252	719164	58.212	ng	98
88) Benzo(k)fluoranthene	15.21	252	495449	42.478	ng	99
89) Benzo(a)pyrene	15.55	252	588812	53.267	ng	98
90) Dibenzo(a,h)anthracene	17.12	278	383362	42.227	ng	97
91) Benzo(g,h,i)perylene	17.56	276	408653	46.461	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.79 min Scan# 799
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS




Tgt Ion: 152 Resp: 91290
Ion Ratio Lower Upper
152 100
150 157.2 124.6 186.8
115 64.0 50.1 75.1




#2

1,4-Dioxane
Concen: 33.102 ng
RT: 2.53 min Scan# 75
Delta R.T. 0.04 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion: 88 Resp: 92087
Ion Ratio Lower Upper
88 100
58 65.7 62.6 93.8
43 24.2 24.6 37.0#





#3

Pyridine

Concen: 33.402 ng

RT: 3.26 min Scan# 199

Delta R.T. 0.04 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 79 Resp: 261254

Ion Ratio Lower Upper




79 100

52 61.2 59.8 89.6

51 30.2 29.8 44.8



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 52.00 (51.70 to 52.70): BF1
Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 37.248 ng

RT: 3.23 min Scan# 195

Delta R.T. 0.02 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion: 42 Resp: 101840

Ion Ratio Lower Upper



42 100


74 163.0 104.9 157.3#

44 6.8 6.9 10.3#



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 24.551 ng

RT: 5.41 min Scan# 565

Delta R.T. 0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 112 Resp: 146579

Ion Ratio Lower Upper

112 100

64 54.3 47.9 71.9




63 27.1 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 18.957 ng

RT: 6.45 min Scan# 742

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion: 93 Resp: 193206

Ion Ratio Lower Upper

93 100

66 89.9 32.2 48.2#



65 54.3 16.4 24.6#




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 61.855 ng

RT: 6.43 min Scan# 739

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 99 Resp: 453744

Ion Ratio Lower Upper

99 100

42 15.3 14.5 21.7




71 33.0 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 22.704 ng

RT: 6.58 min Scan# 764

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion: 128 Resp: 143069

Ion Ratio Lower Upper

128 100

130 32.5 13.0 53.0



64 42.0 29.4 69.4




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 54.158 ng

RT: 6.35 min Scan# 724

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 77 Resp: 178298

Ion Ratio Lower Upper

77 100

105 109.0 81.1 121.1




106 100.0 74.5 114.5



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 34.306 ng

RT: 6.45 min Scan# 742

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion: 94 Resp: 267015

Ion Ratio Lower Upper

94 100

65 35.6 7.9 47.9



66 58.9 16.2 56.2#




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1









#11
bis(2-Chloroethyl)ether
Concen: 41.495 ng
RT: 6.53 min Scan# 755
Delta R.T. -0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 93 Resp: 257733
Ion Ratio Lower Upper
93 100
63 67.4 58.6 98.6
95 32.4 11.8 51.8



#12
1,3-Dichlorobenzene
Concen: 40.564 ng
RT: 6.73 min Scan# 789
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion: 146 Resp: 289580
Ion Ratio Lower Upper
146 100
148 63.6 51.8 77.8
75 30.1 24.2 36.4









#13
 1,4-Dichlorobenzene
 Concen: 41.547 ng
 RT: 6.80 min Scan# 802
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:146 Resp: 295658
 Ion Ratio Lower Upper
 146 100
 148 62.3 50.8 76.2
 111 43.6 34.2 51.2



#14
 1,2-Dichlorobenzene
 Concen: 41.941 ng
 RT: 6.96 min Scan# 829
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:146 Resp: 276589
 Ion Ratio Lower Upper
 146 100
 148 64.1 50.6 75.8
 111 42.0 36.0 54.0









#15
Benzyl Alcohol
Concen: 38.842 ng
RT: 6.94 min Scan# 825
Delta R.T. -0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 79 Resp: 216412
Ion Ratio Lower Upper
79 100
108 78.7 65.1 97.7
77 62.8 50.6 75.8



#16
2,2'-oxybis(1-Chloropropane)
Concen: 33.440 ng
RT: 7.06 min Scan# 846
Delta R.T. -0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion: 45 Resp: 278932
Ion Ratio Lower Upper
45 100
77 50.7 0.0 32.9#
79 46.0 0.0 29.6#









#17
2-Methylphenol
Concen: 39.279 ng
RT: 7.06 min Scan# 845
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:	107	Resp:	215157
Ion	Ratio	Lower	Upper
107	100		
108	110.1	91.7	137.5
77	57.8	33.8	50.8#
79	55.1	33.8	50.8#




Abundance Ion 107.00 (106.70 to 107.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF1
Ion 79.00 (78.70 to 79.70): BF1






#18
Hexachloroethane
Concen: 36.405 ng
RT: 7.30 min Scan# 886
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:	117	Resp:	92369
Ion	Ratio	Lower	Upper
117	100		
119	97.2	75.8	113.8
201	100.3	75.7	113.5



Abundance Ion 117.00 (116.70 to 117.70): E
Ion 119.00 (118.70 to 119.70): E
Ion 201.00 (200.70 to 201.70): E





#19

n-Nitroso-di-n-propylamine

Concen: 36.910 ng

RT: 7.21 min Scan# 871

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS

Tgt Ion: 70 Resp: 176928


Ion Ratio Lower Upper

70 100

42 50.8 47.5 71.3

101 9.6 7.4 11.2

130 22.1 16.6 25.0





Abundance Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): BF1

Ion 130.00 (129.70 to 130.70): BF1



#20

3+4-Methylphenols

Concen: 40.854 ng

RT: 7.22 min Scan# 872

Delta R.T. 0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Tgt Ion: 107 Resp: 277541



Ion Ratio Lower Upper

107 100

108 85.7 68.2 108.2

77 40.2 26.7 66.7

79 24.9 6.3 46.3






Abundance Ion 107.00 (106.70 to 107.70): BF1

Ion 108.00 (107.70 to 108.70): BF1

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.07 min Scan# 1018

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:136 Resp: 371150



Ion Ratio Lower Upper

136 100

137 11.0 8.9 13.3

54 6.4 6.6 9.8#

68 5.3 5.0 7.4





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 41.285 ng

RT: 7.20 min Scan# 870

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Tgt Ion:105 Resp: 377239



Ion Ratio Lower Upper

105 100

71 4.9 4.4 6.6

51 22.6 22.3 33.5

120 21.5 16.9 25.3





Abundance Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 88.145 ng
 RT: 7.36 min Scan# 897
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 82 Resp: 501011
 Ion Ratio Lower Upper
 82 100
 128 48.4 36.1 54.1
 54 43.9 41.4 62.2



#24
 Nitrobenzene
 Concen: 43.712 ng
 RT: 7.38 min Scan# 900
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion: 77 Resp: 274313
 Ion Ratio Lower Upper
 77 100
 123 50.8 37.9 56.9
 65 13.1 11.0 16.4





#25

Isophorone

Concen: 40.822 ng

RT: 7.62 min Scan# 940

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS





Tgt Ion: 82 Resp: 490654

Ion Ratio Lower Upper

82 100

95 7.2 5.2 7.8

138 17.7 14.9 22.3



#26

2-Nitrophenol

Concen: 19.901 ng

RT: 7.69 min Scan# 953

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03




Tgt Ion: 139 Resp: 50842


Ion Ratio Lower Upper

139 100

109 29.4 22.7 34.1

65 42.5 40.5 60.7









#27
2,4-Dimethylphenol
Concen: 45.371 ng
RT: 7.74 min Scan# 961
Delta R.T. 0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:122 Resp: 240675
Ion Ratio Lower Upper
122 100
107 115.3 91.2 136.8
121 59.8 47.2 70.8




Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E






#28
bis(2-Chloroethoxy)methane
Concen: 43.737 ng
RT: 7.83 min Scan# 976
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion: 93 Resp: 321417
Ion Ratio Lower Upper
93 100
95 33.0 26.3 39.5
123 11.6 9.8 14.6



Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 95.00 (94.70 to 95.70): BF1
Ion 123.00 (122.70 to 123.70): E









#29
2,4-Dichlorophenol
Concen: 22.033 ng
RT: 7.95 min Scan# 997
Delta R.T. 0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion	Resp	Lower	Upper
162	111519		
164	64.7	42.7	82.7
98	35.7	18.1	58.1




Abundance Ion 162.00 (161.70 to 162.70): E
Ion 164.00 (163.70 to 164.70): E
Ion 98.00 (97.70 to 98.70): BF1






#30
1,2,4-Trichlorobenzene
Concen: 42.894 ng
RT: 8.02 min Scan# 1008
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion	Resp	Lower	Upper
180	238261		
182	94.4	76.8	115.2
145	32.9	26.5	39.7



Abundance Ion 180.00 (179.70 to 180.70): E
Ion 182.00 (181.70 to 182.70): E
Ion 145.00 (144.70 to 145.70): E





#31

Naphthalene

Concen: 44.275 ng

RT: 8.10 min Scan# 1022

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:128 Resp: 818267

Ion Ratio Lower Upper

128 100

129 11.2 8.8 13.2




127 13.7 10.9 16.3



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#32

Benzoic acid

Concen: 56.864 ng

RT: 7.74 min Scan# 961

Delta R.T. -0.16 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:122 Resp: 240675

Ion Ratio Lower Upper

122 100

105 3.6 101.1 141.1#



77 31.5 70.7 110.7#




Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF1





#33

4-Chloroaniline

Concen: 19.580 ng

RT: 8.16 min Scan# 1032

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:127 Resp: 155032


Ion Ratio Lower Upper

127 100

129 32.2 25.9 38.9

65 26.9 25.0 37.4

92 18.0 15.2 22.8






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF1

Ion 92.00 (91.70 to 92.70): BF1



#34

Hexachlorobutadiene

Concen: 41.887 ng

RT: 8.20 min Scan# 1040

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:225 Resp: 127442

Ion Ratio Lower Upper

225 100

223 62.6 50.6 76.0



227 64.2 51.9 77.9




Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E









#35
 Caprolactam
 Concen: 41.846 ng
 RT: 8.51 min Scan# 1092
 Delta R.T. -0.04 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:113 Resp: 73886
 Ion Ratio Lower Upper
 113 100
 55 147.1 163.3 203.3#
 56 113.4 115.7 155.7#



#36
 4-Chloro-3-methylphenol
 Concen: 41.594 ng
 RT: 8.65 min Scan# 1116
 Delta R.T. 0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:107 Resp: 225735
 Ion Ratio Lower Upper
 107 100
 144 25.7 20.5 30.7
 142 78.1 62.0 93.0









#37
2-Methylnaphthalene
Concen: 44.397 ng
RT: 8.79 min Scan# 1140
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion	Ratio	Lower	Upper
142	100		
141	89.8	70.8	106.2
115	33.3	26.1	39.1




Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 115.00 (114.70 to 115.70): E






#38
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.83 min Scan# 1317
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.9	86.1	129.1
160	47.1	38.3	57.5



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#39

1,2,4,5-Tetrachlorobenzene

Concen: 45.660 ng

RT: 8.96 min Scan# 1168

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:216 Resp: 223744


Ion Ratio Lower Upper

216 100

214 78.6 63.0 94.4

179 22.6 18.1 27.1

108 18.1 17.2 25.8






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E



#40

Hexachlorocyclopentadiene

Concen: 6.572 ng

RT: 8.93 min Scan# 1164

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:237 Resp: 18029

Ion Ratio Lower Upper

237 100

235 61.2 44.8 84.8



272 14.1 0.0 33.3




Abundance Ion 236.80 (236.50 to 237.50): E

Ion 234.90 (234.60 to 235.60): E

Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 22.019 ng
 RT: 10.63 min Scan# 1452
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:330 Resp: 39099
 Ion Ratio Lower Upper
 330 100
 332 97.3 77.0 115.6
 141 36.9 29.9 44.9



#42
 2,4,6-Trichlorophenol
 Concen: 10.227 ng
 RT: 9.07 min Scan# 1188
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:196 Resp: 34788
 Ion Ratio Lower Upper
 196 100
 198 95.3 75.7 113.5
 200 31.1 24.5 36.7









#43
 2,4,5-Trichlorophenol
 Concen: 14.788 ng
 RT: 9.12 min Scan# 1196
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:	196	Resp:	56290
Ion	Ratio	Lower	Upper
196	100		
198	96.8	74.6	112.0
97	50.3	42.9	64.3
132	30.3	26.3	39.5




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 75.100 ng
 RT: 9.15 min Scan# 1201
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:	172	Resp:	813786
Ion	Ratio	Lower	Upper
172	100		
171	36.5	29.7	44.5
170	24.5	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#45
 1,1'-Biphenyl
 Concen: 42.365 ng
 RT: 9.25 min Scan# 1218
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:	154	Resp:	609235
Ion	Ratio	Lower	Upper
154	100		
153	41.1	21.3	61.3
76	13.2	0.0	34.0




Abundance Ion 154.00 (153.70 to 154.70): E
 Ion 153.00 (152.70 to 153.70): E
 Ion 76.00 (75.70 to 76.70): BF1






#46
 2-Chloronaphthalene
 Concen: 42.128 ng
 RT: 9.28 min Scan# 1223
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:	162	Resp:	478527
Ion	Ratio	Lower	Upper
162	100		
127	41.3	33.9	50.9
164	33.0	26.5	39.7



Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E









#47
 2-Nitroaniline
 Concen: 49.555 ng
 RT: 9.39 min Scan# 1241
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 65 Resp: 139200
 Ion Ratio Lower Upper
 65 100
 92 75.0 55.8 83.6
 138 122.0 91.2 136.8



#48
 Acenaphthylene
 Concen: 40.294 ng
 RT: 9.70 min Scan# 1294
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion: 152 Resp: 718092
 Ion Ratio Lower Upper
 152 100
 151 20.3 16.3 24.5
 153 12.9 10.7 16.1





#49

Dimethylphthalate

Concen: 49.216 ng

RT: 9.56 min Scan# 1270

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:163 Resp: 664368

Ion Ratio Lower Upper

163 100

194 3.4 2.7 4.1




164 9.8 8.2 12.2



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 49.357 ng

RT: 9.63 min Scan# 1282

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:165 Resp: 123286

Ion Ratio Lower Upper

165 100

63 47.8 44.7 67.1



89 51.7 44.0 66.0




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1





#51

Acenaphthene

Concen: 39.977 ng

RT: 9.87 min Scan# 1323

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:154 Resp: 434697

Ion Ratio Lower Upper




154 100

153 115.7 91.2 136.8

152 55.8 43.8 65.8



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 38.613 ng

RT: 9.80 min Scan# 1311

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:138 Resp: 112913

Ion Ratio Lower Upper



138 100


108 10.7 9.4 14.0

92 104.5 89.4 134.2



Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1









#54
Dibenzofuran
Concen: 41.516 ng
RT: 10.04 min Scan# 1352
Delta R.T. -0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:168 Resp: 629108
Ion Ratio Lower Upper
168 100
139 38.8 30.1 45.1
169 13.7 10.9 16.3




Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E






#55
4-Nitrophenol
Concen: 4.551 ng
RT: 9.99 min Scan# 1344
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:139 Resp: 10455
Ion Ratio Lower Upper
139 100
109 68.2 48.4 88.4
65 86.2 72.6 112.6



Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1






#56
2,4-Dinitrotoluene
Concen: 45.366 ng
RT: 10.03 min Scan# 1351
Delta R.T. -0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion	Ratio	Lower	Upper
165	100		
63	50.6	36.4	54.6
89	75.7	57.8	86.6
182	2.2	1.6	2.4






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1


Ion 89.00 (88.70 to 89.70): BF1

Ion 182.00 (181.70 to 182.70): E



#57
Fluorene
Concen: 46.841 ng
RT: 10.39 min Scan# 1411
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03



Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.7	79.8	119.8
167	12.9	10.6	16.0




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#58
 2,3,4,6-Tetrachlorophenol
 Concen: 8.460 ng
 RT: 10.16 min Scan# 1373
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 232 Resp: 24025
 Ion Ratio Lower Upper
 232 100
 131 49.7 43.8 65.8
 130 2.2 2.1 3.1
 166 0.0 27.8 41.6#



#59
 Diethylphthalate
 Concen: 39.729 ng
 RT: 10.25 min Scan# 1388
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion: 149 Resp: 534126
 Ion Ratio Lower Upper
 149 100
 177 21.4 16.1 24.1
 150 12.6 10.1 15.1









#60
 4-Chlorophenyl-phenylether
 Concen: 46.258 ng
 RT: 10.37 min Scan# 1409
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 204 Resp: 227224
 Ion Ratio Lower Upper
 204 100
 206 32.5 26.5 39.7
 141 61.9 54.6 81.8



#61
 4-Nitroaniline
 Concen: 47.356 ng
 RT: 10.42 min Scan# 1416
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion: 138 Resp: 135401
 Ion Ratio Lower Upper
 138 100
 92 47.2 30.2 70.2
 108 68.8 52.5 92.5





#62

Azobenzene

Concen: 36.118 ng

RT: 10.53 min Scan# 1436

Delta R.T. -0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS

Tgt Ion: 77 Resp: 463917


Ion Ratio Lower Upper

77 100

182 25.5 1.7 41.7

105 23.9 3.0 43.0

51 25.8 9.9 49.9






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF1



#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.32 min Scan# 1570

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion: 188 Resp: 284058

Ion Ratio Lower Upper

188 100

94 9.6 8.5 12.7



80 11.0 9.2 13.8




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#64
4,6-Dinitro-2-methylphenol
Concen: 8.431 ng
RT: 10.45 min Scan# 1422
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:198 Resp: 1771
Ion Ratio Lower Upper
198 100
51 53.1 37.7 77.7
105 58.5 36.3 76.3



#65
n-Nitrosodiphenylamine
Concen: 44.512 ng
RT: 10.49 min Scan# 1429
Delta R.T. -0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:169 Resp: 438402
Ion Ratio Lower Upper
169 100
168 67.7 54.4 81.6
167 37.1 29.8 44.6









#66
 4-Bromophenyl-phenylether
 Concen: 44.833 ng
 RT: 10.86 min Scan# 1492
 Delta R.T. -0.01 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:	248	Resp:	135461
Ion	Ratio	Lower	Upper
248	100		
250	97.6	76.9	115.3
141	87.6	74.4	111.6



#67
 Hexachlorobenzene
 Concen: 42.922 ng
 RT: 10.93 min Scan# 1504
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:	284	Resp:	145925
Ion	Ratio	Lower	Upper
284	100		
142	33.3	34.6	52.0#
249	29.8	24.8	37.2









#68
 Atrazine
 Concen: 47.129 ng
 RT: 11.03 min Scan# 1520
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS

Tgt Ion: 200 Resp: 140110
 Ion Ratio Lower Upper
 200 100
 173 27.0 8.6 48.6
 215 44.5 25.1 65.1




Abundance Ion 200.00 (199.70 to 200.70): E
 Ion 173.00 (172.70 to 173.70): E
 Ion 215.00 (214.70 to 215.70): E






#69
 Pentachlorophenol
 Concen: 10.175 ng
 RT: 11.13 min Scan# 1538
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion: 266 Resp: 21401
 Ion Ratio Lower Upper
 266 100
 268 63.1 51.1 76.7
 264 58.5 49.5 74.3



Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E









#70
Phenanthrene
Concen: 56.822 ng
RT: 11.35 min Scan# 1575
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:178 Resp: 898875
Ion Ratio Lower Upper
178 100
176 20.1 15.8 23.8
179 15.8 13.0 19.6



#71
Anthracene
Concen: 45.397 ng
RT: 11.40 min Scan# 1584
Delta R.T. -0.00 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:178 Resp: 729988
Ion Ratio Lower Upper
178 100
176 18.5 15.4 23.2
179 15.3 12.6 19.0









#72
 Carbazole
 Concen: 39.509 ng
 RT: 11.56 min Scan# 1611
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:167 Resp: 620817
 Ion Ratio Lower Upper
 167 100
 166 22.3 17.6 26.4
 139 13.6 10.9 16.3



#73
 Di-n-butylphthalate
 Concen: 41.657 ng
 RT: 11.88 min Scan# 1665
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion:149 Resp: 799588
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.8 11.6
 104 4.7 3.8 5.8





#74

Fluoranthene

Concen: 54.200 ng

RT: 12.54 min Scan# 1778

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS




Tgt Ion:	202	Resp:	895807
Ion	Ratio	Lower	Upper
202	100		
101	13.2	0.0	34.1
203	17.9	0.0	38.1



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#75

Chrysene-d12

Concen: 20.000 ng


RT: 14.04 min Scan# 2032

Delta R.T. 0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03



Tgt Ion:	240	Resp:	198832
Ion	Ratio	Lower	Upper
240	100		
120	11.2	9.5	14.3
236	25.4	20.7	31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#76

Benzidine

Concen: 47.204 ng

RT: 12.67 min Scan# 1799

Delta R.T. -0.00 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:184 Resp: 302824

Ion Ratio Lower Upper

184 100

185 14.3 12.6 18.8




183 11.8 9.3 13.9



Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E



#77

Pyrene

Concen: 59.607 ng

RT: 12.78 min Scan# 1818

Delta R.T. 0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:202 Resp: 878494

Ion Ratio Lower Upper

202 100

200 21.4 17.4 26.2



203 17.4 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#78
 Terphenyl-d14
 Concen: 78.133 ng
 RT: 12.92 min Scan# 1841
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03




Instrument :
 BNA_F
 ClientSampleId :
 AMENDED-COMPOSITE-P5-WC MS


Tgt Ion: 244 Resp: 681427
 Ion Ratio Lower Upper
 244 100
 212 7.1 5.4 8.0
 122 12.7 11.0 16.4



#79
 Butylbenzylphthalate
 Concen: 42.223 ng
 RT: 13.40 min Scan# 1924
 Delta R.T. -0.00 min
 Lab File: BF105028.D
 Acq: 2 May 2018 14:03

Tgt Ion: 149 Resp: 293170
 Ion Ratio Lower Upper
 149 100
 91 70.8 56.8 85.2
 206 17.7 14.1 21.1









#80
Benzo(a)anthracene
Concen: 57.120 ng
RT: 14.03 min Scan# 2030
Delta R.T. 0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:228 Resp: 678495
Ion Ratio Lower Upper
228 100
226 28.8 22.6 33.8
229 20.2 15.8 23.6



#81
3,3'-Dichlorobenzidine
Concen: 34.085 ng
RT: 13.99 min Scan# 2023
Delta R.T. 0.01 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:252 Resp: 150958
Ion Ratio Lower Upper
252 100
254 63.9 50.4 75.6
126 13.0 11.3 16.9





#82

Chrysene

Concen: 51.655 ng

RT: 14.07 min Scan# 2037

Delta R.T. 0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03

Instrument :

BNA_F

ClientSampleId :

AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:228 Resp: 630245

Ion Ratio Lower Upper

228 100

226 31.0 25.0 37.6




229 20.4 16.2 24.4



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#83

Bis(2-ethylhexyl)phthalate

Concen: 47.142 ng

RT: 14.00 min Scan# 2026

Delta R.T. 0.01 min

Lab File: BF105028.D

Acq: 2 May 2018 14:03


Tgt Ion:149 Resp: 421017

Ion Ratio Lower Upper

149 100

167 26.5 21.3 31.9



279 3.6 3.0 4.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#84
Di-n-octyl phthalate
Concen: 46.405 ng
RT: 14.70 min Scan# 2144
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion	Ratio	Lower	Upper
149	100		
167	1.5	1.2	1.8
43	7.7	8.4	12.6



#85
Indeno(1,2,3-cd)pyrene
Concen: 48.845 ng
RT: 17.11 min Scan# 2554
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion	Ratio	Lower	Upper
276	100		
138	26.4	25.0	37.6
277	25.0	20.1	30.1









#86
Perylene-d12
Concen: 20.000 ng
RT: 15.61 min Scan# 2299
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:264 Resp: 195608
Ion Ratio Lower Upper
264 100
260 23.4 19.2 28.8
265 21.8 17.5 26.3




Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E






#87
Benzo(b)fluoranthene
Concen: 58.212 ng
RT: 15.18 min Scan# 2226
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:252 Resp: 719164
Ion Ratio Lower Upper
252 100
253 22.2 17.0 25.6
125 12.2 9.0 13.6



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#88
Benzo(k)fluoranthene
Concen: 42.478 ng
RT: 15.21 min Scan# 2231
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03




Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS


Tgt Ion:252 Resp: 495449
Ion Ratio Lower Upper
252 100
253 22.3 17.9 26.9
125 11.9 10.2 15.2



#89
Benzo(a)pyrene
Concen: 53.267 ng
RT: 15.55 min Scan# 2289
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:252 Resp: 588812
Ion Ratio Lower Upper
252 100
253 22.1 17.1 25.7
125 12.9 9.8 14.6









#90
Dibenzo(a,h)anthracene
Concen: 42.227 ng
RT: 17.12 min Scan# 2556
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Instrument :
BNA_F
ClientSampleId :
AMENDED-COMPOSITE-P5-WC MS

Tgt Ion:278 Resp: 383362
Ion Ratio Lower Upper
278 100
139 17.4 15.9 23.9
279 24.0 19.0 28.4



#91
Benzo(g,h,i)perylene
Concen: 46.461 ng
RT: 17.56 min Scan# 2630
Delta R.T. 0.02 min
Lab File: BF105028.D
Acq: 2 May 2018 14:03

Tgt Ion:276 Resp: 408653
Ion Ratio Lower Upper
276 100
277 24.2 17.9 26.9
138 22.8 21.8 32.8

