

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
 Data File : BF133182.D
 Acq On : 02 May 2023 11:39
 Operator : CG\JU
 Sample : 02521-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 02 12:08:58 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 02 10:52:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	217825	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	867706	20.000	ng	0.00
39) Acenaphthene-d10	9.769	164	448217	20.000	ng	0.00
64) Phenanthrene-d10	11.251	188	782429	20.000	ng	0.00
76) Chrysene-d12	13.886	240	465239	20.000	ng	0.00
86) Perylene-d12	15.304	264	390223	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.339	112	965783	66.977	ng	0.00
7) Phenol-d6	6.363	99	732248	40.913	ng	-0.01
23) Nitrobenzene-d5	7.298	82	1397459	86.931	ng	0.00
42) 2,4,6-Tribromophenol	10.557	330	579848	128.636	ng	0.00
45) 2-Fluorobiphenyl	9.092	172	2288051	85.403	ng	0.00
79) Terphenyl-d14	12.839	244	2436044	90.306	ng	0.00
Target Compounds						
					Qvalue	
9) Benzaldehyde	6.275	77	2428	Below Cal	#	80
18) Hexachloroethane	7.269	117	14273	2.632	ng	# 1
19) n-Nitroso-di-n-propyla...	7.298	70	187243	16.805	ng	# 74
25) Isophorone	7.298	82	1391994	45.606	ng	# 63
51) 2,6-Dinitrotoluene	9.769	165	59775	8.498	ng	# 25
57) 2,4-Dinitrotoluene	9.769	165	59775	6.663	ng	# 23
67) 4-Bromophenyl-phenylether	10.557	248	35886	4.229	ng	# 1
77) Benzidine	12.839	184	29373	3.733	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
Data File : BF133182.D
Acq On : 02 May 2023 11:39
Operator : CG\JU
Sample : 02521-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: May 02 12:08:58 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 02 10:52:52 2023
Response via : Initial Calibration

