

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
 Data File : BF133194.D
 Acq On : 02 May 2023 17:52
 Operator : CG\JU
 Sample : 02571-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: May 02 18:19:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue May 02 10:52:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	208861	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	816971	20.000	ng	0.00
39) Acenaphthene-d10	9.763	164	414273	20.000	ng	0.00
64) Phenanthrene-d10	11.245	188	661508	20.000	ng	0.00
76) Chrysene-d12	13.880	240	364665	20.000	ng	-0.01
86) Perylene-d12	15.304	264	376699	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.346	112	1369736	99.068	ng	0.00
7) Phenol-d6	6.369	99	1705607	99.387	ng	0.00
23) Nitrobenzene-d5	7.292	82	1011948	66.859	ng	0.00
42) 2,4,6-Tribromophenol	10.557	330	381936	91.673	ng	0.00
45) 2-Fluorobiphenyl	9.092	172	1681101	67.890	ng	0.00
79) Terphenyl-d14	12.833	244	1443227	68.257	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.269	77	499	Below Cal	#	71
19) n-Nitroso-di-n-propyla...	7.292	70	135210	12.656	ng	# 77
25) Isophorone	7.292	82	1006384	35.020	ng	# 63
51) 2,6-Dinitrotoluene	9.763	165	54272	8.348	ng	# 25
57) 2,4-Dinitrotoluene	9.763	165	54272	6.546	ng	# 24
63) Azobenzene	10.480	77	205886	7.123	ng	# 1
67) 4-Bromophenyl-phenylether	10.551	248	24938	3.476	ng	# 1
77) Benzidine	12.833	184	17755	2.879	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
Data File : BF133194.D
Acq On : 02 May 2023 17:52
Operator : CG\JU
Sample : 02571-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: May 02 18:19:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 02 10:52:52 2023
Response via : Initial Calibration

