

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
 Data File : BF133198.D
 Acq On : 02 May 2023 19:54
 Operator : CG\JU
 Sample : 02572-01 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RBR-251671

Quant Time: May 03 05:55:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 04:52:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	228479	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	878554	20.000	ng	0.00
39) Acenaphthene-d10	9.769	164	400448	20.000	ng	0.00
64) Phenanthrene-d10	11.251	188	596603	20.000	ng	0.00
76) Chrysene-d12	13.892	240	368800	20.000	ng	0.00
86) Perylene-d12	15.315	264	319837	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.334	112	636063	42.054	ng	0.00
7) Phenol-d6	6.369	99	761549	40.566	ng	0.00
23) Nitrobenzene-d5	7.292	82	437603	26.886	ng	0.00
42) 2,4,6-Tribromophenol	10.551	330	135806	33.722	ng	-0.01
45) 2-Fluorobiphenyl	9.086	172	822050	34.344	ng	-0.01
79) Terphenyl-d14	12.833	244	658225	30.781	ng	0.00
Target Compounds						
					Qvalue	
9) Benzaldehyde	6.275	77	1869	Below Cal	#	79
19) n-Nitroso-di-n-propyla...	7.292	70	57812	4.947	ng	# 76
25) Isophorone	7.292	82	436303	14.118	ng	# 63
31) Naphthalene	8.033	128	426086	9.700	ng	99
33) 4-Chloroaniline	8.033	127	55877	3.148	ng	# 32
37) 2-Methylnaphthalene	8.727	142	80158	2.669	ng	99
49) Acenaphthylene	9.627	152	98687	2.657	ng	99
50) Dimethylphthalate	9.769	163	81859	2.932	ng	# 1
51) 2,6-Dinitrotoluene	9.769	165	55209	8.785	ng	# 25
52) Acenaphthene	9.798	154	208159	8.369	ng	99
55) Dibenzofuran	9.969	168	264787	7.804	ng	99
56) 4-Nitrophenol	9.969	139	100507	17.366	ng	# 17
58) Fluorene	10.316	166	299288	12.040	ng	99
63) Azobenzene	10.474	77	73678	2.637	ng	# 1
70) Pentachlorophenol	11.063	266	41139	8.712	ng	96
71) Phenanthrene	11.280	178	2335128	72.823	ng	99
72) Anthracene	11.327	178	882669	26.898	ng	100
73) Carbazole	11.480	167	432364	14.797	ng	99
75) Fluoranthene	12.474	202	3197435	99.356	ng	97
78) Pyrene	12.698	202	3014476	95.351	ng	97
81) Benzo(a)anthracene	13.880	228	1246455	49.148	ng	97
83) Chrysene	13.921	228	1379979	54.426	ng	97
87) Indeno(1,2,3-cd)pyrene	16.703	276	386657	21.253	ng	93
88) Benzo(b)fluoranthene	14.915	252	1806245	88.769	ng	# 96
89) Benzo(k)fluoranthene	14.915	252	1806245	89.549	ng	# 96
90) Benzo(a)pyrene	15.262	252	673682	36.375	ng	98
91) Dibenzo(a,h)anthracene	16.709	278	95226	6.276	ng	# 88
92) Benzo(g,h,i)perylene	17.121	276	344860	23.021	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
Data File : BF133198.D
Acq On : 02 May 2023 19:54
Operator : CG\JU
Sample : 02572-01 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
RBR-251671

Quant Time: May 03 05:55:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 04:52:44 2023
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.728 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:152 Resp: 228479
Ion Ratio Lower Upper
152 100
150 157.2 124.8 187.2
115 62.7 49.9 74.9



#5
2-Fluorophenol
Concen: 42.054 ng
RT: 5.334 min Scan# 552
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:112 Resp: 636063
Ion Ratio Lower Upper
112 100
64 56.3 46.8 70.2
63 28.9 23.8 35.8









#7
Phenol-d6
Concen: 40.566 ng
RT: 6.369 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion: 99 Resp: 761549
Ion Ratio Lower Upper
99 100
42 17.9 16.3 24.5
71 30.0 25.6 38.4



#19
n-Nitroso-di-n-propylamine
Concen: 4.947 ng
RT: 7.292 min Scan# 885
Delta R.T. 0.141 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion: 70 Resp: 57812
Ion Ratio Lower Upper
70 100
42 50.2 51.8 77.6#
101 0.0 8.0 12.0#
130 2.4 16.8 25.2#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.016 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Instrument :
BNA_F
ClientSampleId :
RBR-251671

Tgt Ion:136 Resp: 878554




Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	13.0
54	8.4	7.0	10.4
68	6.6	5.4	8.2




#23
Nitrobenzene-d5
Concen: 26.886 ng
RT: 7.292 min Scan# 885
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion: 82 Resp: 437603

Ion	Ratio	Lower	Upper
82	100		
128	49.5	37.7	56.5
54	51.5	43.0	64.4









#25
Isophorone
Concen: 14.118 ng
RT: 7.292 min Scan# 81
Delta R.T. -0.265 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion: 82 Resp: 436303
Ion Ratio Lower Upper
82 100
95 0.1 5.4 8.0#
138 0.0 16.0 24.0#



#31
Naphthalene
Concen: 9.700 ng
RT: 8.033 min Scan# 1011
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:128 Resp: 426086
Ion Ratio Lower Upper
128 100
129 11.5 8.8 13.2
127 13.3 10.7 16.1









#33
4-Chloroaniline
Concen: 3.148 ng
RT: 8.033 min Scan# 1109
Delta R.T. -0.059 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:	127	Resp:	55877
Ion	Ratio	Lower	Upper
127	100		
129	86.3	25.8	38.8#
65	0.0	23.6	35.4#
92	0.3	15.3	22.9#



#37
2-Methylnaphthalene
Concen: 2.669 ng
RT: 8.727 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:	142	Resp:	80158
Ion	Ratio	Lower	Upper
142	100		
141	85.8	69.4	104.0









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.769 min Scan# 1
Delta R.T. 0.000 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:164 Resp: 400448
Ion Ratio Lower Upper
164 100
162 98.1 78.7 118.1
160 44.1 35.8 53.6



#42
2,4,6-Tribromophenol
Concen: 33.722 ng
RT: 10.551 min Scan# 1439
Delta R.T. -0.012 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:330 Resp: 135806
Ion Ratio Lower Upper
330 100
332 98.6 77.1 115.7
141 34.6 25.8 38.8





#45

2-Fluorobiphenyl

Concen: 34.344 ng

RT: 9.086 min Scan# 1192

Delta R.T. -0.012 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:172 Resp: 822050

Ion Ratio Lower Upper

172 100

171 35.2 28.8 43.2

170 23.3 19.4 29.2



#49

Acenaphthylene

Concen: 2.657 ng

RT: 9.627 min Scan# 1282

Delta R.T. -0.006 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:152 Resp: 98687


Ion Ratio Lower Upper

152 100

151 20.1 16.6 24.8

153 14.3 11.5 17.3









#50
Dimethylphthalate
Concen: 2.932 ng
RT: 9.769 min Scan# 1306
Delta R.T. 0.271 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:163 Resp: 81859
Ion Ratio Lower Upper
163 100
194 0.0 3.1 4.7#
164 476.9 8.5 12.7#



#51
2,6-Dinitrotoluene
Concen: 8.785 ng
RT: 9.769 min Scan# 1306
Delta R.T. 0.212 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:165 Resp: 55209
Ion Ratio Lower Upper
165 100
63 1.4 45.9 68.9#
89 1.7 44.0 66.0#









#52
Acenaphthene
Concen: 8.369 ng
RT: 9.798 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:154 Resp: 208159
Ion Ratio Lower Upper
154 100
153 110.6 89.0 133.6
152 50.9 40.4 60.6



#55
Dibenzofuran
Concen: 7.804 ng
RT: 9.969 min Scan# 1340
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:168 Resp: 264787
Ion Ratio Lower Upper
168 100
139 37.7 29.8 44.8
169 13.6 10.8 16.2









#56
4-Nitrophenol
Concen: 17.366 ng
RT: 9.969 min Scan# 1340
Delta R.T. 0.076 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:139 Resp: 100507
Ion Ratio Lower Upper
139 100
109 1.6 39.4 79.4#
65 1.5 57.5 97.5#



#58
Fluorene
Concen: 12.040 ng
RT: 10.316 min Scan# 1399
Delta R.T. -0.006 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:166 Resp: 299288
Ion Ratio Lower Upper
166 100
165 97.3 78.6 117.8
167 13.3 11.0 16.4









#63
Azobenzene
Concen: 2.637 ng
RT: 10.474 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion: 77	Resp: 73678
Ion Ratio	Lower Upper
77 100	
182 114.3	5.9 45.9#
105 148.8	2.3 42.3#
51 40.9	11.1 51.1



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.251 min Scan# 1558
Delta R.T. 0.000 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion: 188	Resp: 596603
Ion Ratio	Lower Upper
188 100	
94 10.8	9.7 14.5
80 11.4	10.4 15.6





#70

Pentachlorophenol

Concen: 8.712 ng

RT: 11.063 min Scan# 1526

Delta R.T. 0.000 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:266 Resp: 41139

Ion Ratio Lower Upper

266 100

268 61.5 48.8 73.2

264 57.8 50.6 76.0



#71

Phenanthrene

Concen: 72.823 ng

RT: 11.280 min Scan# 1563

Delta R.T. 0.000 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:178 Resp: 2335128


Ion Ratio Lower Upper

178 100

176 20.2 15.9 23.9

179 16.1 12.5 18.7





#72

Anthracene

Concen: 26.898 ng

RT: 11.327 min Scan# 1571

Delta R.T. 0.000 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:178 Resp: 882669

Ion Ratio Lower Upper

178 100

176 19.3 15.4 23.0

179 15.8 12.4 18.6



#73

Carbazole

Concen: 14.797 ng

RT: 11.480 min Scan# 1597

Delta R.T. -0.006 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:167 Resp: 432364


Ion Ratio Lower Upper

167 100

166 21.3 17.4 26.0

139 13.0 10.6 16.0





#75

Fluoranthene

Concen: 99.356 ng

RT: 12.474 min Scan# 1766

Delta R.T. 0.012 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:202 Resp: 3197435

Ion Ratio Lower Upper

202 100

101 13.2 0.0 35.3

203 18.6 0.0 38.1



#76

Chrysene-d12

Concen: 20.000 ng

RT: 13.892 min Scan# 2007

Delta R.T. 0.000 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:240 Resp: 368800


Ion Ratio Lower Upper

240 100

120 14.0 10.5 15.7

236 26.4 21.5 32.3





#78

Pyrene

Concen: 95.351 ng

RT: 12.698 min Scan# 1803

Delta R.T. 0.006 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:202 Resp: 3014476

Ion Ratio Lower Upper

202 100

200 22.6 17.1 25.7

203 19.0 14.1 21.1



#79

Terphenyl-d14

Concen: 30.781 ng

RT: 12.833 min Scan# 1827

Delta R.T. -0.006 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:244 Resp: 658225


Ion Ratio Lower Upper

244 100

212 6.7 5.1 7.7

122 14.6 8.6 13.0#









#81
Benzo(a)anthracene
Concen: 49.148 ng
RT: 13.880 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54




Instrument :
BNA_F
ClientSampleId :
RBR-251671


Tgt Ion:228 Resp: 1246455
Ion Ratio Lower Upper
228 100
226 29.6 22.2 33.4
229 21.6 16.0 24.0



#83
Chrysene
Concen: 54.426 ng
RT: 13.921 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BF133198.D
Acq: 02 May 2023 19:54

Tgt Ion:228 Resp: 1379979
Ion Ratio Lower Upper
228 100
226 31.1 24.7 37.1
229 23.1 16.1 24.1





#86

Perylene-d12

Concen: 20.000 ng

RT: 15.315 min Scan# 21

Delta R.T. 0.006 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:264 Resp: 319837

Ion Ratio Lower Upper

264 100

260 24.0 18.8 28.2

265 24.5 17.0 25.4



#87

Indeno(1,2,3-cd)pyrene

Concen: 21.253 ng

RT: 16.703 min Scan# 2485

Delta R.T. 0.012 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:276 Resp: 386657


Ion Ratio Lower Upper

276 100

138 26.8 25.5 38.3

277 23.5 20.4 30.6





#88

Benzo(b)fluoranthene

Concen: 88.769 ng

RT: 14.915 min Scan# 2181

Delta R.T. 0.012 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:252 Resp: 1806245

Ion Ratio Lower Upper

252 100

253 23.0 17.6 26.4

125 13.9 8.9 13.3#



#89

Benzo(k)fluoranthene

Concen: 89.549 ng

RT: 14.915 min Scan# 2181

Delta R.T. -0.018 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:252 Resp: 1806245


Ion Ratio Lower Upper

252 100

253 23.0 17.7 26.5

125 13.9 9.0 13.4#





#90

Benzo(a)pyrene

Concen: 36.375 ng

RT: 15.262 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54

Instrument :

BNA_F

ClientSampleId :

RBR-251671





Tgt Ion:252 Resp: 673682

Ion Ratio Lower Upper

252 100

253 23.1 17.7 26.5

125 13.3 10.5 15.7



#91

Dibenzo(a,h)anthracene

Concen: 6.276 ng

RT: 16.709 min Scan# 2486

Delta R.T. 0.000 min

Lab File: BF133198.D

Acq: 02 May 2023 19:54




Tgt Ion:278 Resp: 95226


Ion Ratio Lower Upper

278 100

139 26.2 16.6 24.8#

279 29.6 18.8 28.2#





#92
 Benzo(g,h,i)perylene
 Concen: 23.021 ng
 RT: 17.121 min Scan# 21
 Delta R.T. 0.012 min
 Lab File: BF133198.D
 Acq: 02 May 2023 19:54

Instrument :
 BNA_F
 ClientSampleId :
 RBR-251671

Tgt Ion:276 Resp: 344860
 Ion Ratio Lower Upper
 276 100
 277 24.0 19.4 29.0
 138 26.3 22.7 34.1

