

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
 Data File : BF133179.D
 Acq On : 02 May 2023 09:54
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 10:16:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 01 15:02:31 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.614	0.587	4.4	102	0.00
3	Pyridine	1.631	1.606	1.5	102	0.00
4	n-Nitrosodimethylamine	0.845	0.751	11.1	94	0.00
5 S	2-Fluorophenol	1.324	1.272	3.9	102	0.00
6	Aniline	2.117	2.020	4.6	98	0.00
7 S	Phenol-d6	1.643	1.555	5.4	101	0.00
8	2-Chlorophenol	1.344	1.295	3.6	102	0.00
9	Benzaldehyde	0.791	0.765	3.3	107	0.00
10 C	Phenol	1.813	1.666	8.1	97	0.00
11	bis(2-Chloroethyl)ether	1.360	1.266	6.9	98	0.00
12	1,3-Dichlorobenzene	1.429	1.348	5.7	100	0.00
13 C	1,4-Dichlorobenzene	1.432	1.355	5.4	99	0.00
14	1,2-Dichlorobenzene	1.321	1.261	4.5	101	0.00
15	Benzyl Alcohol	1.135	1.071	5.6	96	0.00
16	2,2'-oxybis(1-Chloropropane)	2.363	2.046	13.4	91	0.00
17	2-Methylphenol	1.231	1.158	5.9	99	0.00
18	Hexachloroethane	0.498	0.468	6.0	99	0.00
19 P	n-Nitroso-di-n-propylamine	1.023	0.856	16.3	91	0.00
20	3+4-Methylphenols	1.450	1.351	6.8	100	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
22	Acetophenone	0.463	0.435	6.0	100	0.00
23 S	Nitrobenzene-d5	0.371	0.350	5.7	97	0.00
24	Nitrobenzene	0.375	0.361	3.7	99	0.00
25	Isophorone	0.704	0.645	8.4	95	0.00
26 C	2-Nitrophenol	0.181	0.187	-3.3	102	0.00
27	2,4-Dimethylphenol	0.311	0.300	3.5	98	0.00
28	bis(2-Chloroethoxy)methane	0.416	0.382	8.2	95	0.00
29 C	2,4-Dichlorophenol	0.283	0.278	1.8	100	0.00
30	1,2,4-Trichlorobenzene	0.293	0.281	4.1	100	0.00
31	Naphthalene	1.000	0.957	4.3	99	0.00
32	Benzoic acid	0.192	0.209	-8.9	104	0.00
33	4-Chloroaniline	0.404	0.419	-3.7	109	0.00
34 C	Hexachlorobutadiene	0.162	0.155	4.3	99	0.00
35	Caprolactam	0.095	0.100	-5.3	105	0.00
36 C	4-Chloro-3-methylphenol	0.305	0.299	2.0	100	0.00
37	2-Methylnaphthalene	0.684	0.645	5.7	98	0.00
38	1-Methylnaphthalene	0.640	0.607	5.2	99	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
40	1,2,4,5-Tetrachlorobenzene	0.538	0.512	4.8	99	0.00
41 P	Hexachlorocyclopentadiene	0.314	0.286	8.9	89	0.00
42 S	2,4,6-Tribromophenol	0.201	0.198	1.5	100	0.00
43 C	2,4,6-Trichlorophenol	0.372	0.364	2.2	100	0.00
44	2,4,5-Trichlorophenol	0.430	0.419	2.6	99	0.00
45 S	2-Fluorobiphenyl	1.195	1.150	3.8	99	0.00
46	1,1'-Biphenyl	1.586	1.516	4.4	99	0.00
47	2-Chloronaphthalene	1.161	1.119	3.6	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
 Data File : BF133179.D
 Acq On : 02 May 2023 09:54
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 10:16:06 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon May 01 15:02:31 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.384	0.385	-0.3	100	0.00
49	Acenaphthylene	1.855	1.819	1.9	100	0.00
50	Dimethylphthalate	1.395	1.356	2.8	100	0.00
51	2,6-Dinitrotoluene	0.314	0.311	1.0	101	0.00
52 C	Acenaphthene	1.242	1.207	2.8	100	0.00
53	3-Nitroaniline	0.373	0.393	-5.4	105	0.00
54 P	2,4-Dinitrophenol	0.158	0.168	-6.3	100	0.00
55	Dibenzofuran	1.695	1.646	2.9	101	0.00
56 P	4-Nitrophenol	0.289	0.292	-1.0	99	0.00
57	2,4-Dinitrotoluene	0.400	0.424	-6.0	104	0.00
58	Fluorene	1.241	1.213	2.3	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.333	0.332	0.3	102	0.00
60	Diethylphthalate	1.317	1.299	1.4	102	0.00
61	4-Chlorophenyl-phenylether	0.605	0.577	4.6	101	0.00
62	4-Nitroaniline	0.343	0.395	-15.2	119	0.00
63	Azobenzene	1.395	1.315	5.7	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	0.121	0.133	-9.9	108	0.00
66 c	n-Nitrosodiphenylamine	0.649	0.611	5.9	103	0.00
67	4-Bromophenyl-phenylether	0.217	0.203	6.5	101	0.00
68	Hexachlorobenzene	0.222	0.212	4.5	102	0.00
69	Atrazine	0.187	0.190	-1.6	106	0.00
70 C	Pentachlorophenol	0.158	0.150	5.1	96	0.00
71	Phenanthrene	1.075	1.010	6.0	104	0.00
72	Anthracene	1.100	1.055	4.1	105	0.00
73	Carbazole	0.980	0.971	0.9	107	0.00
74	Di-n-butylphthalate	1.109	1.131	-2.0	107	0.00
75 C	Fluoranthene	1.079	1.081	-0.2	108	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	120	0.00
77	Benzidine	0.338	0.413	-22.2	146	0.00
78	Pyrene	1.714	1.554	9.3	106	0.00
79 S	Terphenyl-d14	1.160	1.016	12.4	104	0.00
80	Butylbenzylphthalate	0.607	0.682	-12.4	123	0.00
81	Benzo(a)anthracene	1.375	1.256	8.7	107	0.00
82	3,3'-Dichlorobenzidine	0.380	0.421	-10.8	123	0.00
83	Chrysene	1.375	1.273	7.4	109	0.00
84	Bis(2-ethylhexyl)phthalate	0.762	0.769	-0.9	113	0.00
85 c	Di-n-octyl phthalate	1.242	1.349	-8.6	120	0.00
86 I	Perylene-d12	1.000	1.000	0.0	121	0.00
87	Indeno(1,2,3-cd)pyrene	1.138	1.138	0.0	114	0.00
88	Benzo(b)fluoranthene	1.272	1.291	-1.5	117	0.00
89	Benzo(k)fluoranthene	1.261	1.137	9.8	107	0.00
90 C	Benzo(a)pyrene	1.158	1.110	4.1	112	0.00
91	Dibenzo(a,h)anthracene	0.949	0.939	1.1	112	0.00
92	Benzo(g,h,i)perylene	0.937	0.948	-1.2	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
Data File : BF133179.D
Acq On : 02 May 2023 09:54
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 10:16:06 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon May 01 15:02:31 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0