

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
 Data File : BF133183.D
 Acq On : 02 May 2023 12:11
 Operator : CG\JU
 Sample : 02521-05
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 GMW-10

Quant Time: May 03 06:17:59 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 03 04:52:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.728	152	222157	20.000	ng	0.00
21) Naphthalene-d8	8.016	136	884085	20.000	ng	0.00
39) Acenaphthene-d10	9.769	164	465386	20.000	ng	0.00
64) Phenanthrene-d10	11.251	188	807304	20.000	ng	0.00
76) Chrysene-d12	13.880	240	532323	20.000	ng	-0.01
86) Perylene-d12	15.304	264	408265	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	800652	54.442	ng	0.00
7) Phenol-d6	6.363	99	614294	33.653	ng	-0.01
23) Nitrobenzene-d5	7.298	82	1226039	74.854	ng	0.00
42) 2,4,6-Tribromophenol	10.557	330	529655	113.166	ng	0.00
45) 2-Fluorobiphenyl	9.092	172	2090898	75.165	ng	0.00
79) Terphenyl-d14	12.839	244	2345841	76.003	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050223\
Data File : BF133183.D
Acq On : 02 May 2023 12:11
Operator : CG\JU
Sample : 02521-05
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
GMW-10

Quant Time: May 03 06:17:59 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042123.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 03 04:52:44 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.728 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: BF133183.D
 Acq: 02 May 2023 12:11




Instrument :
 BNA_F
 ClientSampleId :
 GMW-10


Tgt Ion:152 Resp: 222157
 Ion Ratio Lower Upper
 152 100
 150 158.6 124.8 187.2
 115 62.5 49.9 74.9



#5
 2-Fluorophenol
 Concen: 54.442 ng
 RT: 5.340 min Scan# 553
 Delta R.T. 0.000 min
 Lab File: BF133183.D
 Acq: 02 May 2023 12:11

Tgt Ion:112 Resp: 800652
 Ion Ratio Lower Upper
 112 100
 64 57.6 46.8 70.2
 63 28.4 23.8 35.8









#7
Phenol-d6
Concen: 33.653 ng
RT: 6.363 min Scan# 71
Delta R.T. -0.012 min
Lab File: BF133183.D
Acq: 02 May 2023 12:11




Instrument :
BNA_F
ClientSampleId :
GMW-10


Tgt Ion: 99 Resp: 614294
Ion Ratio Lower Upper
99 100
42 19.4 16.3 24.5
71 31.3 25.6 38.4



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.016 min Scan# 1008
Delta R.T. 0.000 min
Lab File: BF133183.D
Acq: 02 May 2023 12:11

Tgt Ion: 136 Resp: 884085
Ion Ratio Lower Upper
136 100
137 10.7 8.6 13.0
54 8.2 7.0 10.4
68 6.8 5.4 8.2





#23

Nitrobenzene-d5

Concen: 74.854 ng

RT: 7.298 min Scan# 81

Delta R.T. 0.000 min

Lab File: BF133183.D

Acq: 02 May 2023 12:11

Instrument :





BNA_F

ClientSampleId :

GMW-10

Tgt Ion: 82 Resp: 1226039

Ion	Ratio	Lower	Upper
82	100		
128	49.3	37.7	56.5
54	51.7	43.0	64.4



#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.769 min Scan# 1306




Delta R.T. 0.000 min


Lab File: BF133183.D

Acq: 02 May 2023 12:11

Tgt Ion: 164 Resp: 465386

Ion	Ratio	Lower	Upper
164	100		
162	97.0	78.7	118.1
160	43.8	35.8	53.6









#42
2,4,6-Tribromophenol
Concen: 113.166 ng
RT: 10.557 min Scan# 1440
Delta R.T. -0.006 min
Lab File: BF133183.D
Acq: 02 May 2023 12:11




Instrument :
BNA_F
ClientSampleId :
GMW-10


Tgt Ion:330 Resp: 529655
Ion Ratio Lower Upper
330 100
332 96.4 77.1 115.7
141 29.8 25.8 38.8



#45
2-Fluorobiphenyl
Concen: 75.165 ng
RT: 9.092 min Scan# 1191
Delta R.T. -0.006 min
Lab File: BF133183.D
Acq: 02 May 2023 12:11

Tgt Ion:172 Resp: 2090898
Ion Ratio Lower Upper
172 100
171 36.6 28.8 43.2
170 24.1 19.4 29.2









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.251 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF133183.D
Acq: 02 May 2023 12:11




Instrument :
BNA_F
ClientSampleId :
GMW-10


Tgt Ion:188 Resp: 807304
Ion Ratio Lower Upper
188 100
94 9.9 9.7 14.5
80 10.6 10.4 15.6



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.880 min Scan# 2005
Delta R.T. -0.012 min
Lab File: BF133183.D
Acq: 02 May 2023 12:11

Tgt Ion:240 Resp: 532323
Ion Ratio Lower Upper
240 100
120 15.5 10.5 15.7
236 26.5 21.5 32.3









#79
 Terphenyl-d14
 Concen: 76.003 ng
 RT: 12.839 min Scan# 1828
 Delta R.T. 0.000 min
 Lab File: BF133183.D
 Acq: 02 May 2023 12:11

Instrument :
 BNA_F
 ClientSampleId :
 GMW-10

Tgt Ion:244 Resp: 2345841
 Ion Ratio Lower Upper
 244 100
 212 6.6 5.1 7.7
 122 12.2 8.6 13.0



#86
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.304 min Scan# 2247
 Delta R.T. -0.006 min
 Lab File: BF133183.D
 Acq: 02 May 2023 12:11

Tgt Ion:264 Resp: 408265
 Ion Ratio Lower Upper
 264 100
 260 23.5 18.8 28.2
 265 21.4 17.0 25.4

