

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050224\
 Data File : BF137805.D
 Acq On : 02 May 2024 09:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 11:46:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.538	0.525	2.4	79	0.00
3	Pyridine	1.302	1.240	4.8	77	0.00
4	n-Nitrosodimethylamine	0.586	0.543	7.3	75	0.00
5 S	2-Fluorophenol	1.195	1.183	1.0	81	0.00
6	Aniline	1.502	1.384	7.9	74	0.00
7 S	Phenol-d6	1.515	1.459	3.7	79	0.00
8	2-Chlorophenol	1.302	1.272	2.3	80	0.00
9	Benzaldehyde	0.816	0.845	-3.6	87	0.00
10 C	Phenol	1.553	1.493	3.9	79	0.00
11	bis(2-Chloroethyl)ether	1.216	1.149	5.5	77	0.00
12	1,3-Dichlorobenzene	1.468	1.443	1.7	81	0.00
13 C	1,4-Dichlorobenzene	1.473	1.466	0.5	81	0.00
14	1,2-Dichlorobenzene	1.387	1.361	1.9	80	0.00
15	Benzyl Alcohol	1.048	1.050	-0.2	79	0.00
16	2,2'-oxybis(1-Chloropropane	1.709	1.533	10.3	73	0.00
17	2-Methylphenol	1.050	1.039	1.0	80	0.00
18	Hexachloroethane	0.498	0.492	1.2	79	0.00
19 P	n-Nitroso-di-n-propylamine	0.905	0.840	7.2	76	0.00
20	3+4-Methylphenols	1.382	1.365	1.2	79	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
22	Acetophenone	0.450	0.441	2.0	78	0.00
23 S	Nitrobenzene-d5	0.344	0.340	1.2	78	0.00
24	Nitrobenzene	0.355	0.346	2.5	77	0.00
25	Isophorone	0.621	0.599	3.5	78	0.00
26 C	2-Nitrophenol	0.161	0.180	-11.8	83	0.00
27	2,4-Dimethylphenol	0.231	0.228	1.3	79	0.00
28	bis(2-Chloroethoxy)methane	0.376	0.356	5.3	76	0.00
29 C	2,4-Dichlorophenol	0.280	0.285	-1.8	80	0.00
30	1,2,4-Trichlorobenzene	0.305	0.307	-0.7	80	0.00
31	Naphthalene	0.994	0.971	2.3	78	0.00
32	Benzoic acid	0.200	0.220	-10.0	85	0.00
33	4-Chloroaniline	0.369	0.346	6.2	75	0.00
34 C	Hexachlorobutadiene	0.187	0.190	-1.6	81	0.00
35	Caprolactam	0.089	0.088	1.1	78	0.00
36 C	4-Chloro-3-methylphenol	0.292	0.292	0.0	79	0.00
37	2-Methylnaphthalene	0.643	0.627	2.5	78	0.00
38	1-Methylnaphthalene	0.633	0.620	2.1	79	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	78	0.00
40	1,2,4,5-Tetrachlorobenzene	0.530	0.529	0.2	79	0.00
41 P	Hexachlorocyclopentadiene	0.215	0.222	-3.3	78	0.00
42 S	2,4,6-Tribromophenol	0.204	0.210	-2.9	82	0.00
43 C	2,4,6-Trichlorophenol	0.361	0.362	-0.3	78	0.00
44	2,4,5-Trichlorophenol	0.398	0.404	-1.5	78	0.00
45 S	2-Fluorobiphenyl	1.233	1.218	1.2	80	0.00
46	1,1'-Biphenyl	1.373	1.357	1.2	79	0.00
47	2-Chloronaphthalene	1.108	1.094	1.3	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050224\
 Data File : BF137805.D
 Acq On : 02 May 2024 09:40
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 02 11:46:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 01 03:25:08 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.301	0.306	-1.7	78	0.00
49	Acenaphthylene	1.603	1.590	0.8	79	0.00
50	Dimethylphthalate	1.263	1.264	-0.1	80	0.00
51	2,6-Dinitrotoluene	0.289	0.298	-3.1	80	0.00
52 C	Acenaphthene	1.063	1.052	1.0	79	0.00
53	3-Nitroaniline	0.302	0.306	-1.3	78	0.00
54 P	2,4-Dinitrophenol	0.142	0.154	-8.5	82	0.00
55	Dibenzofuran	1.546	1.509	2.4	78	0.00
56 P	4-Nitrophenol	0.250	0.243	2.8	74	0.00
57	2,4-Dinitrotoluene	0.377	0.400	-6.1	81	0.00
58	Fluorene	1.231	1.201	2.4	78	0.00
59	2,3,4,6-Tetrachlorophenol	0.323	0.320	0.9	77	0.00
60	Diethylphthalate	1.199	1.202	-0.3	80	0.00
61	4-Chlorophenyl-phenylether	0.606	0.604	0.3	80	0.00
62	4-Nitroaniline	0.306	0.305	0.3	76	0.00
63	Azobenzene	1.159	1.097	5.3	75	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
65	4,6-Dinitro-2-methylphenol	0.115	0.126	-9.6	83	0.00
66 c	n-Nitrosodiphenylamine	0.599	0.590	1.5	78	0.00
67	4-Bromophenyl-phenylether	0.214	0.212	0.9	78	0.00
68	Hexachlorobenzene	0.234	0.235	-0.4	80	0.00
69	Atrazine	0.183	0.167	8.7	72	0.00
70 C	Pentachlorophenol	0.161	0.162	-0.6	77	0.00
71	Phenanthrene	0.975	0.952	2.4	78	0.00
72	Anthracene	0.970	0.951	2.0	78	0.00
73	Carbazole	0.935	0.908	2.9	78	0.00
74	Di-n-butylphthalate	1.020	1.029	-0.9	80	0.00
75 C	Fluoranthene	1.066	1.036	2.8	78	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	78	0.00
77	Benzydine	0.463	0.494	-6.7	72	0.00
78	Pyrene	1.586	1.531	3.5	76	0.00
79 S	Terphenyl-d14	1.193	1.189	0.3	80	0.00
80	Butylbenzylphthalate	0.503	0.563	-11.9	85	0.00
81	Benzo(a)anthracene	1.279	1.269	0.8	79	0.00
82	3,3'-Dichlorobenzidine	0.383	0.387	-1.0	82	0.00
83	Chrysene	1.198	1.169	2.4	79	0.00
84	Bis(2-ethylhexyl)phthalate	0.634	0.729	-15.0	91	0.00
85 c	Di-n-octyl phthalate	0.853	1.075	-26.0#	100	0.00
86 I	Perylene-d12	1.000	1.000	0.0	91	0.00
87	Indeno(1,2,3-cd)pyrene	1.137	1.128	0.8	91	0.00
88	Benzo(b)fluoranthene	1.244	1.196	3.9	89	0.00
89	Benzo(k)fluoranthene	1.203	1.173	2.5	88	0.00
90 C	Benzo(a)pyrene	1.015	1.011	0.4	92	0.00
91	Dibenzo(a,h)anthracene	0.926	0.923	0.3	91	0.00
92	Benzo(g,h,i)perylene	0.970	0.927	4.4	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050224\
Data File : BF137805.D
Acq On : 02 May 2024 09:40
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 02 11:46:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF043024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 01 03:25:08 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1