


Data Path : Z:\HPCHEM1\BNA_F\Data\BF050316\
 Data File : BF086635.D
 Acq On : 3 May 2016 9:35
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF042716.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Apr 27 14:20:52 2016



AutoFind: Scans 826, 827, 828; Background Corrected with Scan 823

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	61.9	51392	PASS
68	69	0.00	2	1.7	624	PASS
69	198	0.00	100	45.1	37405	PASS
70	69	0.00	2	0.5	194	PASS
127	198	10	80	54.4	45167	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	82986	PASS
199	198	5	9	7.1	5861	PASS
275	198	10	60	24.6	20445	PASS
365	198	1	100	3.3	2758	PASS
441	442	0.01	24	15.8	11164	PASS
442	198	50	100	85.0	70577	PASS
443	442	15	24	18.5	13081	PASS