

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050321\
 Data File : BF123784.D
 Acq On : 3 May 2021 17:33
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 03 18:01:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 16:30:39 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.680	0.670	1.5	94	0.00
3	Pyridine	1.662	1.643	1.1	93	0.00
4	n-Nitrosodimethylamine	0.916	0.932	-1.7	95	0.00
5 S	2-Fluorophenol	1.262	1.214	3.8	92	0.00
6	Aniline	2.011	1.926	4.2	90	0.00
7 S	Phenol-d6	1.739	1.663	4.4	93	0.00
8	2-Chlorophenol	1.358	1.288	5.2	90	0.00
9	Benzaldehyde	0.842	0.843	-0.1	100	0.00
10 C	Phenol	1.868	1.775	5.0	91	0.00
11	bis(2-Chloroethyl)ether	1.364	1.297	4.9	91	0.00
12	1,3-Dichlorobenzene	1.372	1.314	4.2	91	0.00
13 C	1,4-Dichlorobenzene	1.388	1.346	3.0	93	0.00
14	1,2-Dichlorobenzene	1.348	1.247	7.5	94	0.00
15	Benzyl Alcohol	1.363	1.358	0.4	94	0.00
16	2,2'-oxybis(1-Chloropropane	2.901	2.831	2.4	93	0.00
17	2-Methylphenol	1.157	1.106	4.4	91	0.00
18	Hexachloroethane	0.549	0.534	2.7	91	0.00
19 P	n-Nitroso-di-n-propylamine	1.074	1.009	6.1	90	0.00
20	3+4-Methylphenols	1.473	1.355	8.0	94	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.544	0.536	1.5	93	0.00
23 S	Nitrobenzene-d5	0.443	0.442	0.2	92	0.00
24	Nitrobenzene	0.461	0.465	-0.9	93	0.00
25	Isophorone	0.831	0.807	2.9	89	0.00
26 C	2-Nitrophenol	0.188	0.191	-1.6	91	0.00
27	2,4-Dimethylphenol	0.296	0.293	1.0	92	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.414	2.1	89	0.00
29 C	2,4-Dichlorophenol	0.292	0.290	0.7	92	0.00
30	1,2,4-Trichlorobenzene	0.306	0.299	2.3	91	0.00
31	Naphthalene	1.030	1.011	1.8	91	0.00
32	Benzoic acid	0.186	0.172	7.5	84	0.00
33	4-Chloroaniline	0.430	0.429	0.2	92	0.00
34 C	Hexachlorobutadiene	0.186	0.187	-0.5	91	0.00
35	Caprolactam	0.102	0.100	2.0	88	0.00
36 C	4-Chloro-3-methylphenol	0.364	0.359	1.4	89	0.00
37	2-Methylnaphthalene	0.671	0.661	1.5	91	0.00
38	1-Methylnaphthalene	0.631	0.617	2.2	92	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	0.589	0.581	1.4	91	0.00
41 P	Hexachlorocyclopentadiene	0.233	0.239	-2.6	88	0.00
42 S	2,4,6-Tribromophenol	0.249	0.245	1.6	90	0.00
43 C	2,4,6-Trichlorophenol	0.405	0.401	1.0	89	0.00
44	2,4,5-Trichlorophenol	0.435	0.430	1.1	89	0.00
45 S	2-Fluorobiphenyl	1.373	1.303	5.1	91	0.00
46	1,1'-Biphenyl	1.639	1.558	4.9	89	0.00
47	2-Chloronaphthalene	1.235	1.192	3.5	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050321\
 Data File : BF123784.D
 Acq On : 3 May 2021 17:33
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 03 18:01:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Apr 27 16:30:39 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.544	0.543	0.2	91	0.00
49	Acenaphthylene	1.994	1.935	3.0	90	0.00
50	Dimethylphthalate	1.409	1.329	5.7	89	0.00
51	2,6-Dinitrotoluene	0.326	0.317	2.8	89	0.00
52 C	Acenaphthene	1.215	1.189	2.1	92	0.00
53	3-Nitroaniline	0.389	0.369	5.1	87	0.00
54 P	2,4-Dinitrophenol	0.173	0.166	4.0	82	0.00
55	Dibenzofuran	1.838	1.764	4.0	89	0.00
56 P	4-Nitrophenol	0.283	0.282	0.4	87	0.00
57	2,4-Dinitrotoluene	0.444	0.434	2.3	89	0.00
58	Fluorene	1.412	1.341	5.0	89	0.00
59	2,3,4,6-Tetrachlorophenol	0.344	0.334	2.9	87	0.00
60	Diethylphthalate	1.499	1.421	5.2	88	0.00
61	4-Chlorophenyl-phenylether	0.658	0.614	6.7	88	0.00
62	4-Nitroaniline	0.385	0.375	2.6	88	0.00
63	Azobenzene	1.724	1.626	5.7	87	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	0.124	0.133	-7.3	90	0.00
66 c	n-Nitrosodiphenylamine	0.654	0.642	1.8	88	0.00
67	4-Bromophenyl-phenylether	0.211	0.207	1.9	87	0.00
68	Hexachlorobenzene	0.240	0.235	2.1	90	0.00
69	Atrazine	0.198	0.188	5.1	85	0.00
70 C	Pentachlorophenol	0.122	0.128	-4.9	92	0.00
71	Phenanthrene	1.049	1.022	2.6	89	0.00
72	Anthracene	1.043	1.021	2.1	89	0.00
73	Carbazole	1.054	1.033	2.0	90	0.00
74	Di-n-butylphthalate	1.184	1.131	4.5	85	0.00
75 C	Fluoranthene	1.144	1.109	3.1	88	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
77	Benzydine	0.518	0.498	3.9	72	0.00
78	Pyrene	1.552	1.634	-5.3	88	0.00
79 S	Terphenyl-d14	1.039	1.076	-3.6	87	0.00
80	Butylbenzylphthalate	0.663	0.662	0.2	81	0.00
81	Benzo(a)anthracene	1.212	1.198	1.2	83	0.00
82	3,3'-Dichlorobenzidine	0.372	0.403	-8.3	89	0.00
83	Chrysene	1.219	1.208	0.9	84	0.00
84	Bis(2-ethylhexyl)phthalate	0.824	0.780	5.3	78	0.00
85 c	Di-n-octyl phthalate	1.317	1.182	10.3	73	0.00
86 I	Perylene-d12	1.000	1.000	0.0	82	0.00
87	Indeno(1,2,3-cd)pyrene	1.165	1.258	-8.0	96	0.00
88	Benzo(b)fluoranthene	1.348	1.335	1.0	84	0.00
89	Benzo(k)fluoranthene	1.286	1.191	7.4	75	0.00
90 C	Benzo(a)pyrene	1.153	1.138	1.3	82	0.00
91	Dibenzo(a,h)anthracene	0.983	1.062	-8.0	96	0.00
92	Benzo(g,h,i)perylene	0.985	1.101	-11.8	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050321\
Data File : BF123784.D
Acq On : 3 May 2021 17:33
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 03 18:01:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF042721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Apr 27 16:30:39 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0