

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
 Data File : BF128049.D
 Acq On : 03 May 2022 16:49
 Operator : CG\JU
 Sample : N2664-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 S-6

Quant Time: May 04 04:27:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.922	152	73720	20.000	ng	0.00
21) Naphthalene-d8	8.222	136	209311	20.000	ng	# 0.02
39) Acenaphthene-d10	10.004	164	120174	20.000	ng	# 0.04
64) Phenanthrene-d10	11.492	188	159114	20.000	ng	# 0.04
76) Chrysene-d12	14.104	240	192778	20.000	ng	0.00
86) Perylene-d12	15.609	264	211426	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.539	112	547536	118.300	ng	0.00
7) Phenol-d6	6.545	99	712498	118.638	ng	0.00
23) Nitrobenzene-d5	7.486	82	480228	107.889	ng	0.00
42) 2,4,6-Tribromophenol	10.804	330	127893	105.718	ng	0.05
45) 2-Fluorobiphenyl	9.304	172	550039	77.431	ng	0.02
79) Terphenyl-d14	13.057	244	605515	57.435	ng	0.01
Target Compounds						
						Qvalue
37) 2-Methylnaphthalene	8.945	142	358602	50.898	ng	96
38) 1-Methylnaphthalene	9.051	142	1305434	190.627	ng	99
52) Acenaphthene	10.039	154	241709	34.683	ng	# 77
58) Fluorene	10.563	166	448113	59.726	ng	# 84
71) Phenanthrene	11.521	178	1017556	122.172	ng	# 96
75) Fluoranthene	12.698	202	55284	6.307	ng	71
78) Pyrene	12.927	202	235876	15.157	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
Data File : BF128049.D
Acq On : 03 May 2022 16:49
Operator : CG\JU
Sample : N2664-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
S-6

Quant Time: May 04 04:27:05 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 04:23:29 2022
Response via : Initial Calibration

