

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
 Data File : BF128040.D
 Acq On : 03 May 2022 12:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 04 04:25:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69	0.00
2	1,4-Dioxane	0.599	0.580	3.2	65	0.00
3	Pyridine	1.535	1.556	-1.4	67	0.00
4	n-Nitrosodimethylamine	0.743	0.788	-6.1	71	0.00
5 S	2-Fluorophenol	1.256	1.268	-1.0	70	0.00
6	Aniline	1.936	2.011	-3.9	69	0.00
7 S	Phenol-d6	1.629	1.690	-3.7	71	0.00
8	2-Chlorophenol	1.303	1.339	-2.8	69	0.00
9	Benzaldehyde	1.008	0.880	12.7	68	0.00
10 C	Phenol	1.643	1.755	-6.8	72	0.00
11	bis(2-Chloroethyl)ether	1.363	1.388	-1.8	69	0.00
12	1,3-Dichlorobenzene	1.500	1.498	0.1	69	0.00
13 C	1,4-Dichlorobenzene	1.497	1.499	-0.1	69	0.00
14	1,2-Dichlorobenzene	1.386	1.416	-2.2	71	0.00
15	Benzyl Alcohol	1.108	1.421	-28.2#	82	0.00
16	2,2'-oxybis(1-Chloropropane	2.083	2.127	-2.1	68	0.00
17	2-Methylphenol	1.124	1.143	-1.7	67	0.00
18	Hexachloroethane	0.550	0.578	-5.1	72	0.00
19 P	n-Nitroso-di-n-propylamine	0.997	1.063	-6.6	73	0.00
20	3+4-Methylphenols	1.358	1.410	-3.8	70	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	67	0.00
22	Acetophenone	0.487	0.517	-6.2	72	0.00
23 S	Nitrobenzene-d5	0.425	0.474	-11.5	74	0.00
24	Nitrobenzene	0.410	0.445	-8.5	72	0.00
25	Isophorone	0.716	0.771	-7.7	71	0.00
26 C	2-Nitrophenol	0.175	0.198	-13.1	72	0.00
27	2,4-Dimethylphenol	0.290	0.303	-4.5	67	0.00
28	bis(2-Chloroethoxy)methane	0.459	0.489	-6.5	71	0.00
29 C	2,4-Dichlorophenol	0.304	0.320	-5.3	70	0.00
30	1,2,4-Trichlorobenzene	0.345	0.361	-4.6	70	0.00
31	Naphthalene	1.019	1.051	-3.1	70	0.00
32	Benzoic acid	0.208	0.217	-4.3	63	0.00
33	4-Chloroaniline	0.403	0.419	-4.0	68	0.00
34 C	Hexachlorobutadiene	0.217	0.245	-12.9	76	0.00
35	Caprolactam	0.098	0.103	-5.1	67	0.00
36 C	4-Chloro-3-methylphenol	0.325	0.352	-8.3	71	0.00
37	2-Methylnaphthalene	0.673	0.711	-5.6	71	0.00
38	1-Methylnaphthalene	0.654	0.692	-5.8	71	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	69	0.00
40	1,2,4,5-Tetrachlorobenzene	0.590	0.635	-7.6	74	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.331	-3.8	67	0.00
42 S	2,4,6-Tribromophenol	0.201	0.239	-18.9	81	0.00
43 C	2,4,6-Trichlorophenol	0.388	0.413	-6.4	72	0.00
44	2,4,5-Trichlorophenol	0.422	0.435	-3.1	68	0.00
45 S	2-Fluorobiphenyl	1.182	1.343	-13.6	77	0.00
46	1,1'-Biphenyl	1.481	1.561	-5.4	73	0.00
47	2-Chloronaphthalene	1.128	1.171	-3.8	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
 Data File : BF128040.D
 Acq On : 03 May 2022 12:17
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: May 04 04:25:38 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed May 04 04:23:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.381	0.435	-14.2	75	0.00
49	Acenaphthylene	1.725	1.794	-4.0	71	0.00
50	Dimethylphthalate	1.341	1.418	-5.7	72	0.00
51	2,6-Dinitrotoluene	0.290	0.310	-6.9	70	0.00
52 C	Acenaphthene	1.160	1.222	-5.3	71	0.00
53	3-Nitroaniline	0.322	0.331	-2.8	67	0.00
54 P	2,4-Dinitrophenol	0.153	0.174	-13.7	71	0.00
55	Dibenzofuran	1.673	1.736	-3.8	71	0.00
56 P	4-Nitrophenol	0.246	0.235	4.5	61	0.00
57	2,4-Dinitrotoluene	0.385	0.426	-10.6	72	0.00
58	Fluorene	1.249	1.329	-6.4	74	0.00
59	2,3,4,6-Tetrachlorophenol	0.355	0.377	-6.2	71	0.00
60	Diethylphthalate	1.315	1.425	-8.4	73	0.00
61	4-Chlorophenyl-phenylether	0.620	0.672	-8.4	75	0.00
62	4-Nitroaniline	0.315	0.328	-4.1	68	0.00
63	Azobenzene	1.447	1.564	-8.1	73	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	69	0.00
65	4,6-Dinitro-2-methylphenol	0.119	0.144	-21.0	73	0.00
66 c	n-Nitrosodiphenylamine	0.622	0.641	-3.1	72	0.00
67	4-Bromophenyl-phenylether	0.222	0.237	-6.8	74	0.00
68	Hexachlorobenzene	0.234	0.255	-9.0	76	0.00
69	Atrazine	0.192	0.201	-4.7	73	0.00
70 C	Pentachlorophenol	0.137	0.131	4.4	61	0.00
71	Phenanthrene	1.047	1.068	-2.0	71	0.00
72	Anthracene	1.043	1.074	-3.0	72	0.00
73	Carbazole	1.005	1.000	0.5	69	0.00
74	Di-n-butylphthalate	1.130	1.199	-6.1	73	0.00
75 C	Fluoranthene	1.102	1.118	-1.5	71	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
77	Benzydine	0.378	0.506	-33.9#	83	0.00
78	Pyrene	1.614	1.678	-4.0	70	0.00
79 S	Terphenyl-d14	1.094	1.214	-11.0	77	0.00
80	Butylbenzylphthalate	0.633	0.706	-11.5	72	0.00
81	Benzo(a)anthracene	1.333	1.387	-4.1	69	0.00
82	3,3'-Dichlorobenzidine	0.423	0.402	5.0	64	0.00
83	Chrysene	1.273	1.281	-0.6	68	0.00
84	Bis(2-ethylhexyl)phthalate	0.839	0.921	-9.8	72	0.00
85 c	Di-n-octyl phthalate	1.300	1.389	-6.8	69	0.00
86 I	Perylene-d12	1.000	1.000	0.0	64	0.00
87	Indeno(1,2,3-cd)pyrene	1.335	1.397	-4.6	63	0.00
88	Benzo(b)fluoranthene	1.256	1.319	-5.0	66	0.00
89	Benzo(k)fluoranthene	1.266	1.291	-2.0	62	0.00
90 C	Benzo(a)pyrene	1.042	1.087	-4.3	64	0.00
91	Dibenzo(a,h)anthracene	1.071	1.130	-5.5	63	0.00
92	Benzo(g,h,i)perylene	1.128	1.179	-4.5	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF050322\
Data File : BF128040.D
Acq On : 03 May 2022 12:17
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: May 04 04:25:38 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF042222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed May 04 04:23:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0